



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5291

Client: PARSONS Main of New York, Inc.

Analytical Method: **8082A** **Datafile :** PO108563.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB165649BSD	AR1016	5	4.60	ug/L	92	0			61	112	20
	AR1260	5	4.80	ug/L	96	2			66	113	20