

Instrument ID: **BNA_F**

Daily Analysis Runlog For Sequence/QC Batch ID # BF121624

Review By	yogesh	Review On	12/18/2024 1:40:10 AM
Supervise By	mohammad	Supervise On	12/18/2024 2:36:39 AM
SubDirectory	BF121624	HP Acquire Method	BNA_F
		HP Processing Method	Bf121624
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6680,SP6679,SP6678,SP6677,SP6676,SP6675,SP6674,SP6673		
CCC	SP6676		
Internal Standard/PEM	S12331,10ul/1000ul sample		
ICV/I.BLK	SP6686		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF140854.D	16 Dec 2024 11:21	RC/JU	Ok
2	SSTDICC2.5	BF140855.D	16 Dec 2024 12:13	RC/JU	Ok
3	SSTDICC005	BF140856.D	16 Dec 2024 12:39	RC/JU	Ok
4	SSTDICC010	BF140857.D	16 Dec 2024 13:05	RC/JU	Ok,M
5	SSTDICC020	BF140858.D	16 Dec 2024 14:00	RC/JU	Ok,M
6	SSTDICCC040	BF140859.D	16 Dec 2024 14:26	RC/JU	Ok
7	SSTDICC050	BF140860.D	16 Dec 2024 15:23	RC/JU	Ok
8	SSTDICC060	BF140861.D	16 Dec 2024 15:49	RC/JU	Ok
9	SSTDICC080	BF140862.D	16 Dec 2024 16:15	RC/JU	Ok,M
10	SSTDICV040	BF140863.D	16 Dec 2024 16:41	RC/JU	Ok
11	PB165543BL	BF140864.D	16 Dec 2024 17:08	RC/JU	Ok
12	PB165543BS	BF140865.D	16 Dec 2024 17:34	RC/JU	Ok,M
13	PB165590BL	BF140866.D	16 Dec 2024 18:00	RC/JU	Ok
14	PB165479BS	BF140867.D	16 Dec 2024 18:26	RC/JU	Ok,M
15	PB165479BSD	BF140868.D	16 Dec 2024 18:53	RC/JU	Ok,M
16	PB165479BL	BF140869.D	16 Dec 2024 19:19	RC/JU	Ok
17	P5192-02	BF140870.D	16 Dec 2024 19:45	RC/JU	ReRun

M : Manual Integration

Instrument ID: **BNA_F**

Daily Analysis Runlog For Sequence/QC Batch ID # BF121824

Review By	yogesh	Review On	12/21/2024 1:11:35 AM
Supervise By	mohammad	Supervise On	12/21/2024 2:30:36 AM
SubDirectory	BF121824	HP Acquire Method	BNA_F
		HP Processing Method	bf121624
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6680,SP6679,SP6678,SP6677,SP6676,SP6675,SP6674,SP6673		
CCC	SP6676		
Internal Standard/PEM	S12646,10ul/1000ul sample		
ICV/I.BLK	SP6686		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF140895.D	18 Dec 2024 09:41	RC/JU	Ok
2	SSTDCCC040	BF140896.D	18 Dec 2024 10:43	RC/JU	Ok
3	PB165648BL	BF140897.D	18 Dec 2024 11:27	RC/JU	Ok
4	PB165648BS	BF140898.D	18 Dec 2024 11:53	RC/JU	Ok,M
5	PB165685BL	BF140899.D	18 Dec 2024 12:19	RC/JU	Ok
6	PB165685BS	BF140900.D	18 Dec 2024 12:45	RC/JU	Ok,M
7	PB165682BL	BF140901.D	18 Dec 2024 13:12	RC/JU	Ok
8	PB165682BS	BF140902.D	18 Dec 2024 13:37	RC/JU	Ok,M
9	P5270-07DL2	BF140903.D	18 Dec 2024 14:09	RC/JU	Ok,M
10	P5270-08DL2	BF140904.D	18 Dec 2024 14:35	RC/JU	Ok
11	P5277-02	BF140905.D	18 Dec 2024 15:02	RC/JU	Ok
12	P5277-03	BF140906.D	18 Dec 2024 15:28	RC/JU	Ok
13	P5291-11	BF140907.D	18 Dec 2024 15:54	RC/JU	Ok
14	P5277-01	BF140908.D	18 Dec 2024 16:20	RC/JU	Ok
15	P5277-01MS	BF140909.D	18 Dec 2024 16:46	RC/JU	Ok,M
16	P5277-01MSD	BF140910.D	18 Dec 2024 17:12	RC/JU	Ok,M
17	P5291-06	BF140911.D	18 Dec 2024 17:38	RC/JU	Dilution
18	P5291-08MS	BF140912.D	18 Dec 2024 18:04	RC/JU	Ok
19	P5291-09MSD	BF140913.D	18 Dec 2024 18:30	RC/JU	Ok,M
20	P5291-07	BF140914.D	18 Dec 2024 18:56	RC/JU	Ok
21	P5291-10	BF140915.D	18 Dec 2024 19:23	RC/JU	Ok

Instrument ID: **BNA_F**

Daily Analysis Runlog For Sequence/QC Batch ID # BF121824

Review By	yogesh	Review On	12/21/2024 1:11:35 AM
Supervise By	mohammad	Supervise On	12/21/2024 2:30:36 AM
SubDirectory	BF121824	HP Acquire Method	BNA_F
		HP Processing Method	bf121624
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6680,SP6679,SP6678,SP6677,SP6676,SP6675,SP6674,SP6673		
CCC	SP6676		
Internal Standard/PEM	S12646,10ul/1000ul sample		
ICV/I.BLK	SP6686		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	P5268-01DL	BF140916.D	18 Dec 2024 19:49	RC/JU	Ok,M
----	------------	------------	-------------------	-------	------

M : Manual Integration

Instrument ID: **BNA_F**

Daily Analysis Runlog For Sequence/QC Batch ID # BF121924

Review By	yogesh	Review On	12/21/2024 1:10:39 AM
Supervise By	mohammad	Supervise On	12/21/2024 2:31:44 AM
SubDirectory	BF121924	HP Acquire Method	BNA_F
		HP Processing Method	bf121624
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6680,SP6679,SP6678,SP6677,SP6676,SP6675,SP6674,SP6673		
CCC	SP6676		
Internal Standard/PEM	S12646,10ul/1000ul sample		
ICV/I.BLK	SP6686		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF140917.D	19 Dec 2024 10:30	RC/JU	Ok
2	SSTDCCC040	BF140918.D	19 Dec 2024 11:38	RC/JU	Ok
3	PB165719BL	BF140919.D	19 Dec 2024 12:04	RC/JU	Ok
4	PB165719BS	BF140920.D	19 Dec 2024 12:30	RC/JU	Ok,M
5	PB165680TB	BF140921.D	19 Dec 2024 12:56	RC/JU	Ok
6	P5312-03	BF140922.D	19 Dec 2024 13:29	RC/JU	Ok
7	P5291-13	BF140923.D	19 Dec 2024 13:55	RC/JU	Ok
8	P5291-06DL	BF140924.D	19 Dec 2024 14:21	RC/JU	Ok
9	P5337-01	BF140925.D	19 Dec 2024 14:47	RC/JU	Ok
10	P5291-13MS	BF140926.D	19 Dec 2024 15:13	RC/JU	Ok,M
11	P5291-13MSD	BF140927.D	19 Dec 2024 15:39	RC/JU	Ok,M
12	P5291-05	BF140928.D	19 Dec 2024 16:05	RC/JU	Dilution
13	P5343-01	BF140929.D	19 Dec 2024 16:31	RC/JU	Ok,M
14	P5343-01MS	BF140930.D	19 Dec 2024 16:57	RC/JU	Not Ok
15	P5343-01MSD	BF140931.D	19 Dec 2024 17:23	RC/JU	Not Ok
16	P5343-03	BF140932.D	19 Dec 2024 17:49	RC/JU	ReRun
17	P5330-01	BF140933.D	19 Dec 2024 18:15	RC/JU	ReRun
18	P5312-01	BF140934.D	19 Dec 2024 18:41	RC/JU	Ok
19	P5301-03MS	BF140935.D	19 Dec 2024 19:08	RC/JU	Ok,M
20	P5301-03MSD	BF140936.D	19 Dec 2024 19:34	RC/JU	Ok
21	P5339-01	BF140937.D	19 Dec 2024 20:00	RC/JU	Ok,M

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF121924

Review By	yogesh	Review On	12/21/2024 1:10:39 AM
Supervise By	mohammad	Supervise On	12/21/2024 2:31:44 AM
SubDirectory	BF121924	HP Acquire Method	BNA_F
		HP Processing Method	bf121624
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6680,SP6679,SP6678,SP6677,SP6676,SP6675,SP6674,SP6673		
CCC	SP6676		
Internal Standard/PEM	S12646,10ul/1000ul sample		
ICV/I.BLK	SP6686		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	P5342-01	BF140938.D	19 Dec 2024 20:26	RC/JU	Dilution
23	P5342-05	BF140939.D	19 Dec 2024 20:52	RC/JU	Ok
24	P5342-03	BF140940.D	19 Dec 2024 21:19	RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF122324

Review By	Yogesh	Review On	12/24/2024 8:42:59 AM
Supervise By	mohammad	Supervise On	12/24/2024 11:31:30 PM
SubDirectory	BF122324	HP Acquire Method	BNA_F
		HP Processing Method	bf121624
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6680,SP6679,SP6678,SP6677,SP6676,SP6675,SP6674,SP6673		
CCC	SP6676		
Internal Standard/PEM	S12646,10ul/1000ul sample		
ICV/I.BLK	SP6686		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF140960.D	23 Dec 2024 09:05	RC/JU	Ok
2	SSTDCCC040	BF140961.D	23 Dec 2024 10:18	RC/JU	Ok,M
3	PB165785BL	BF140962.D	23 Dec 2024 11:09	RC/JU	Not Ok
4	P5355-02	BF140963.D	23 Dec 2024 11:49	RC/JU	Ok
5	P5355-01	BF140964.D	23 Dec 2024 12:46	RC/JU	ReRun
6	P5346-01	BF140965.D	23 Dec 2024 14:36	RC/JU	Not Ok
7	P5291-05DL	BF140966.D	23 Dec 2024 15:01	RC/JU	Ok
8	P5342-01DL	BF140967.D	23 Dec 2024 15:27	RC/JU	Ok,M
9	P5346-01	BF140968.D	23 Dec 2024 16:14	RC/JU	Ok

M : Manual Integration