

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF121624

Review By	yogesh	Review On	12/18/2024 1:40:10 AM
Supervise By	mohammad	Supervise On	12/18/2024 2:36:39 AM
SubDirectory	BF121624	HP Acquire Method	BNA_F
		HP Processing Method	Bf121624

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6680,SP6679,SP6678,SP6677,SP6676,SP6675,SP6674,SP6673
CCC	SP6676
Internal Standard/PEM	S12331,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140854.D	16 Dec 2024 11:21		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF140855.D	16 Dec 2024 12:13		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF140856.D	16 Dec 2024 12:39	Compound#32,41,54,56,65,70,77 removed from 5 ppm	RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF140857.D	16 Dec 2024 13:05		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BF140858.D	16 Dec 2024 14:00	The Calibration is Good For 8270 DOD and good for 625.1 Method	RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BF140859.D	16 Dec 2024 14:26	Compound#41 Kept on QR & Compound#54,56 Kept on LR	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BF140860.D	16 Dec 2024 15:23		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BF140861.D	16 Dec 2024 15:49		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF140862.D	16 Dec 2024 16:15		RC/JU	Ok,M
10	SSTDICV040	ICVBF121624	BF140863.D	16 Dec 2024 16:41		RC/JU	Ok
11	PB165543BL	PB165543BL	BF140864.D	16 Dec 2024 17:08		RC/JU	Ok
12	PB165543BS	PB165543BS	BF140865.D	16 Dec 2024 17:34		RC/JU	Ok,M
13	PB165590BL	PB165590BL	BF140866.D	16 Dec 2024 18:00		RC/JU	Ok
14	PB165479BS	PB165479BS	BF140867.D	16 Dec 2024 18:26		RC/JU	Ok,M
15	PB165479BSD	PB165479BSD	BF140868.D	16 Dec 2024 18:53		RC/JU	Ok,M
16	PB165479BL	PB165479BL	BF140869.D	16 Dec 2024 19:19		RC/JU	Ok
17	P5192-02	EFF-WASTE WATER	BF140870.D	16 Dec 2024 19:45	Surrogate Fail	RC/JU	ReRun

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF121624

Review By	yogesh	Review On	12/18/2024 1:40:10 AM		
Supervise By	mohammad	Supervise On	12/18/2024 2:36:39 AM		
SubDirectory	BF121624	HP Acquire Method	BNA_F	HP Processing Method	Bf121624
STD. NAME	STD REF.#				
Tune/Reschk	SP6573				
Initial Calibration Stds	SP6680,SP6679,SP6678,SP6677,SP6676,SP6675,SP6674,SP6673				
CCC	SP6676				
Internal Standard/PEM	S12331,10ul/1000ul sample				
ICV/I.BLK	SP6686				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF121824

Review By	yogesh	Review On	12/21/2024 1:11:35 AM
Supervise By	mohammad	Supervise On	12/21/2024 2:30:36 AM
SubDirectory	BF121824	HP Acquire Method	BNA_F
		HP Processing Method	bf121624

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6680,SP6679,SP6678,SP6677,SP6676,SP6675,SP6674,SP6673
CCC	SP6676
Internal Standard/PEM	S12646,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140895.D	18 Dec 2024 09:41		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF140896.D	18 Dec 2024 10:43		RC/JU	Ok
3	PB165648BL	PB165648BL	BF140897.D	18 Dec 2024 11:27		RC/JU	Ok
4	PB165648BS	PB165648BS	BF140898.D	18 Dec 2024 11:53		RC/JU	Ok,M
5	PB165685BL	PB165685BL	BF140899.D	18 Dec 2024 12:19		RC/JU	Ok
6	PB165685BS	PB165685BS	BF140900.D	18 Dec 2024 12:45		RC/JU	Ok,M
7	PB165682BL	PB165682BL	BF140901.D	18 Dec 2024 13:12		RC/JU	Ok
8	PB165682BS	PB165682BS	BF140902.D	18 Dec 2024 13:37		RC/JU	Ok,M
9	P5270-07DL2	MW12-20241212DL2	BF140903.D	18 Dec 2024 14:09		RC/JU	Ok,M
10	P5270-08DL2	MW14-20241212DL2	BF140904.D	18 Dec 2024 14:35		RC/JU	Ok
11	P5277-02	COMP-2A	BF140905.D	18 Dec 2024 15:02		RC/JU	Ok
12	P5277-03	COMP-3A	BF140906.D	18 Dec 2024 15:28		RC/JU	Ok
13	P5291-11	EQUIP-BLANK-202412	BF140907.D	18 Dec 2024 15:54		RC/JU	Ok
14	P5277-01	COMP-1A	BF140908.D	18 Dec 2024 16:20		RC/JU	Ok
15	P5277-01MS	COMP-1AMS	BF140909.D	18 Dec 2024 16:46		RC/JU	Ok,M
16	P5277-01MSD	COMP-1AMSD	BF140910.D	18 Dec 2024 17:12		RC/JU	Ok,M
17	P5291-06	MW3-20241213	BF140911.D	18 Dec 2024 17:38	Need 5X Dilution	RC/JU	Dilution
18	P5291-08MS	MW6-20241213MS	BF140912.D	18 Dec 2024 18:04		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF121824

Review By	yogesh	Review On	12/21/2024 1:11:35 AM		
Supervise By	mohammad	Supervise On	12/21/2024 2:30:36 AM		
SubDirectory	BF121824	HP Acquire Method	BNA_F	HP Processing Method	bf121624
STD. NAME	STD REF.#				
Tune/Reschk	SP6573				
Initial Calibration Stds	SP6680,SP6679,SP6678,SP6677,SP6676,SP6675,SP6674,SP6673				
CCC	SP6676				
Internal Standard/PEM	S12646,10ul/1000ul sample				
ICV/I.BLK	SP6686				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	P5291-09MSD	MW6-20241213MSD	BF140913.D	18 Dec 2024 18:30		RC/JU	Ok,M
20	P5291-07	MW6-20241213	BF140914.D	18 Dec 2024 18:56		RC/JU	Ok
21	P5291-10	MW6-20241213-A	BF140915.D	18 Dec 2024 19:23		RC/JU	Ok
22	P5268-01DL	OR-02-12122024DL	BF140916.D	18 Dec 2024 19:49		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF121924

Review By	yogesh	Review On	12/21/2024 1:10:39 AM
Supervise By	mohammad	Supervise On	12/21/2024 2:31:44 AM
SubDirectory	BF121924	HP Acquire Method	BNA_F
		HP Processing Method	bf121624

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6680,SP6679,SP6678,SP6677,SP6676,SP6675,SP6674,SP6673
CCC	SP6676
Internal Standard/PEM	S12646,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140917.D	19 Dec 2024 10:30		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF140918.D	19 Dec 2024 11:38		RC/JU	Ok
3	PB165719BL	PB165719BL	BF140919.D	19 Dec 2024 12:04		RC/JU	Ok
4	PB165719BS	PB165719BS	BF140920.D	19 Dec 2024 12:30		RC/JU	Ok,M
5	PB165680TB	PB165680TB	BF140921.D	19 Dec 2024 12:56		RC/JU	Ok
6	P5312-03	CONCRETE-VNJ-222	BF140922.D	19 Dec 2024 13:29		RC/JU	Ok
7	P5291-13	WC-20241213	BF140923.D	19 Dec 2024 13:55		RC/JU	Ok
8	P5291-06DL	MW3-20241213DL	BF140924.D	19 Dec 2024 14:21		RC/JU	Ok
9	P5337-01	ETGI-357	BF140925.D	19 Dec 2024 14:47		RC/JU	Ok
10	P5291-13MS	WC-20241213MS	BF140926.D	19 Dec 2024 15:13		RC/JU	Ok,M
11	P5291-13MSD	WC-20241213MSD	BF140927.D	19 Dec 2024 15:39		RC/JU	Ok,M
12	P5291-05	MW2-20241213	BF140928.D	19 Dec 2024 16:05	Need 2X Dilution	RC/JU	Dilution
13	P5343-01	VNJ210	BF140929.D	19 Dec 2024 16:31		RC/JU	Ok,M
14	P5343-01MS	VNJ210MS	BF140930.D	19 Dec 2024 16:57	MSD is Not Ok	RC/JU	Not Ok
15	P5343-01MSD	VNJ210MSD	BF140931.D	19 Dec 2024 17:23	Internal Standard Fail	RC/JU	Not Ok
16	P5343-03	VNJ281	BF140932.D	19 Dec 2024 17:49	Internal Standard Fail	RC/JU	ReRun
17	P5330-01	TP-5	BF140933.D	19 Dec 2024 18:15	Internal Standard Fail	RC/JU	ReRun
18	P5312-01	SOIL-VNJ-222	BF140934.D	19 Dec 2024 18:41	Internal Standard Fail	RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF121924

Review By	yogesh	Review On	12/21/2024 1:10:39 AM		
Supervise By	mohammad	Supervise On	12/21/2024 2:31:44 AM		
SubDirectory	BF121924	HP Acquire Method	BNA_F	HP Processing Method	bf121624
STD. NAME		STD REF.#			
Tune/Reschk		SP6573			
Initial Calibration Stds		SP6680,SP6679,SP6678,SP6677,SP6676,SP6675,SP6674,SP6673			
CCC		SP6676			
Internal Standard/PEM		S12646,10ul/1000ul sample			
ICV/I.BLK		SP6686			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	P5301-03MS	HR-04-121624MS	BF140935.D	19 Dec 2024 19:08	Internal Standard Fail	RC/JU	Ok,M
20	P5301-03MSD	HR-04-121624MSD	BF140936.D	19 Dec 2024 19:34		RC/JU	Ok
21	P5339-01	TR-06-12182024	BF140937.D	19 Dec 2024 20:00	Internal Standard Fail	RC/JU	Ok,M
22	P5342-01	CHRT26634	BF140938.D	19 Dec 2024 20:26	Internal Standard Fail, Need 5X Dilution	RC/JU	Dilution
23	P5342-05	RB21198	BF140939.D	19 Dec 2024 20:52	Internal Standard Fail	RC/JU	Ok
24	P5342-03	HT2651	BF140940.D	19 Dec 2024 21:19	Internal Standard Fail	RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF122324

Review By	Yogesh	Review On	12/24/2024 8:42:59 AM
Supervise By	mohammad	Supervise On	12/24/2024 11:31:30 PM
SubDirectory	BF122324	HP Acquire Method	BNA_F
		HP Processing Method	bf121624

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6680,SP6679,SP6678,SP6677,SP6676,SP6675,SP6674,SP6673
CCC	SP6676
Internal Standard/PEM	S12646,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140960.D	23 Dec 2024 09:05		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF140961.D	23 Dec 2024 10:18	CCC fail high side for com. #87, 91, 92 and low for com. #41,56	RC/JU	Ok,M
3	PB165785BL	PB165785BL	BF140962.D	23 Dec 2024 11:09	CCC fail low side for compound #41,56	RC/JU	Not Ok
4	P5355-02	RBR251688	BF140963.D	23 Dec 2024 11:49		RC/JU	Ok
5	P5355-01	RBR251688	BF140964.D	23 Dec 2024 12:46	CCC fail low side for compound #41,56	RC/JU	ReRun
6	P5346-01	ETGI-328	BF140965.D	23 Dec 2024 14:36	Internal Standard Not added.	RC/JU	Not Ok
7	P5291-05DL	MW2-20241213DL	BF140966.D	23 Dec 2024 15:01		RC/JU	Ok
8	P5342-01DL	CHRT26634DL	BF140967.D	23 Dec 2024 15:27		RC/JU	Ok,M
9	P5346-01	ETGI-328	BF140968.D	23 Dec 2024 16:14		RC/JU	Ok

M : Manual Integration