

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF121624

Review By	yogesh	Review On	12/18/2024 1:40:10 AM
Supervise By	mohammad	Supervise On	12/18/2024 2:36:39 AM
SubDirectory	BF121624	HP Acquire Method	BNA_F
		HP Processing Method	Bf121624

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6680,SP6679,SP6678,SP6677,SP6676,SP6675,SP6674,SP6673
CCC	SP6676
Internal Standard/PEM	S12331,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140854.D	16 Dec 2024 11:21		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF140855.D	16 Dec 2024 12:13		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF140856.D	16 Dec 2024 12:39	Compound#32,41,54,56,65,70,77 removed from 5 ppm	RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF140857.D	16 Dec 2024 13:05		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BF140858.D	16 Dec 2024 14:00	The Calibration is Good For 8270 DOD and good for 625.1 Method	RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BF140859.D	16 Dec 2024 14:26	Compound#41 Kept on QR & Compound#54,56 Kept on LR	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BF140860.D	16 Dec 2024 15:23		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BF140861.D	16 Dec 2024 15:49		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF140862.D	16 Dec 2024 16:15		RC/JU	Ok,M
10	SSTDICV040	ICVBF121624	BF140863.D	16 Dec 2024 16:41		RC/JU	Ok
11	PB165543BL	PB165543BL	BF140864.D	16 Dec 2024 17:08		RC/JU	Ok
12	PB165543BS	PB165543BS	BF140865.D	16 Dec 2024 17:34		RC/JU	Ok,M
13	PB165590BL	PB165590BL	BF140866.D	16 Dec 2024 18:00		RC/JU	Ok
14	PB165479BS	PB165479BS	BF140867.D	16 Dec 2024 18:26		RC/JU	Ok,M
15	PB165479BSD	PB165479BSD	BF140868.D	16 Dec 2024 18:53		RC/JU	Ok,M
16	PB165479BL	PB165479BL	BF140869.D	16 Dec 2024 19:19		RC/JU	Ok
17	P5192-02	EFF-WASTE WATER	BF140870.D	16 Dec 2024 19:45	Surrogate Fail	RC/JU	ReRun

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF121624

Review By	yogesh	Review On	12/18/2024 1:40:10 AM		
Supervise By	mohammad	Supervise On	12/18/2024 2:36:39 AM		
SubDirectory	BF121624	HP Acquire Method	BNA_F	HP Processing Method	Bf121624
STD. NAME	STD REF.#				
Tune/Reschk	SP6573				
Initial Calibration Stds	SP6680,SP6679,SP6678,SP6677,SP6676,SP6675,SP6674,SP6673				
CCC	SP6676				
Internal Standard/PEM	S12331,10ul/1000ul sample				
ICV/I.BLK	SP6686				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF121924

Review By	yogesh	Review On	12/21/2024 1:10:39 AM
Supervise By	mohammad	Supervise On	12/21/2024 2:31:44 AM
SubDirectory	BF121924	HP Acquire Method	BNA_F
		HP Processing Method	bf121624

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6680,SP6679,SP6678,SP6677,SP6676,SP6675,SP6674,SP6673
CCC	SP6676
Internal Standard/PEM	S12646,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140917.D	19 Dec 2024 10:30		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF140918.D	19 Dec 2024 11:38		RC/JU	Ok
3	PB165719BL	PB165719BL	BF140919.D	19 Dec 2024 12:04		RC/JU	Ok
4	PB165719BS	PB165719BS	BF140920.D	19 Dec 2024 12:30		RC/JU	Ok,M
5	PB165680TB	PB165680TB	BF140921.D	19 Dec 2024 12:56		RC/JU	Ok
6	P5312-03	CONCRETE-VNJ-222	BF140922.D	19 Dec 2024 13:29		RC/JU	Ok
7	P5291-13	WC-20241213	BF140923.D	19 Dec 2024 13:55		RC/JU	Ok
8	P5291-06DL	MW3-20241213DL	BF140924.D	19 Dec 2024 14:21		RC/JU	Ok
9	P5337-01	ETGI-357	BF140925.D	19 Dec 2024 14:47		RC/JU	Ok
10	P5291-13MS	WC-20241213MS	BF140926.D	19 Dec 2024 15:13		RC/JU	Ok,M
11	P5291-13MSD	WC-20241213MSD	BF140927.D	19 Dec 2024 15:39		RC/JU	Ok,M
12	P5291-05	MW2-20241213	BF140928.D	19 Dec 2024 16:05	Need 2X Dilution	RC/JU	Dilution
13	P5343-01	VNJ210	BF140929.D	19 Dec 2024 16:31		RC/JU	Ok,M
14	P5343-01MS	VNJ210MS	BF140930.D	19 Dec 2024 16:57	MSD is Not Ok	RC/JU	Not Ok
15	P5343-01MSD	VNJ210MSD	BF140931.D	19 Dec 2024 17:23	Internal Standard Fail	RC/JU	Not Ok
16	P5343-03	VNJ281	BF140932.D	19 Dec 2024 17:49	Internal Standard Fail	RC/JU	ReRun
17	P5330-01	TP-5	BF140933.D	19 Dec 2024 18:15	Internal Standard Fail	RC/JU	ReRun
18	P5312-01	SOIL-VNJ-222	BF140934.D	19 Dec 2024 18:41	Internal Standard Fail	RC/JU	ReRun

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF121924

Review By	yogesh	Review On	12/21/2024 1:10:39 AM		
Supervise By	mohammad	Supervise On	12/21/2024 2:31:44 AM		
SubDirectory	BF121924	HP Acquire Method	BNA_F	HP Processing Method	bf121624
STD. NAME		STD REF.#			
Tune/Reschk		SP6573			
Initial Calibration Stds		SP6680,SP6679,SP6678,SP6677,SP6676,SP6675,SP6674,SP6673			
CCC		SP6676			
Internal Standard/PEM		S12646,10ul/1000ul sample			
ICV/I.BLK		SP6686			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	P5301-03MS	HR-04-121624MS	BF140935.D	19 Dec 2024 19:08	Internal Standard Fail, Com#41 Recovery is zero	RC/JU	Not Ok
20	P5301-03MSD	HR-04-121624MSD	BF140936.D	19 Dec 2024 19:34	Com#41 Recovery is zero	RC/JU	Not Ok
21	P5339-01	TR-06-12182024	BF140937.D	19 Dec 2024 20:00	Internal Standard Fail	RC/JU	Ok,M
22	P5342-01	CHRT26634	BF140938.D	19 Dec 2024 20:26	Internal Standard Fail, Need 5X Dilution	RC/JU	Dilution
23	P5342-05	RB21198	BF140939.D	19 Dec 2024 20:52	Internal Standard Fail	RC/JU	Ok
24	P5342-03	HT2651	BF140940.D	19 Dec 2024 21:19	Internal Standard Fail, Need Straight Run	RC/JU	Ok

M : Manual Integration