

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P5291

Client: PARSONS Main of New York, Inc.

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: P5291-13MS		Client Sample ID: WC-20241213MS		DataFile: BF140926.D							
Pyridine	500	0	300	ug/L	60				10	109	
1,4-Dichlorobenzene	500	0	510	ug/L	102				55	125	
2-Methylphenol	500	0	510	ug/L	102				37	126	
3+4-Methylphenols	500	0	520	ug/L	104				31	127	
Hexachloroethane	500	0	550	ug/L	110				49	110	
Nitrobenzene	500	0	500	ug/L	100				62	112	
Hexachlorobutadiene	500	0	500	ug/L	100				52	125	
2,4,6-Trichlorophenol	500	0	520	ug/L	104				78	112	
2,4,5-Trichlorophenol	500	0	490	ug/L	98				71	111	
2,4-Dinitrotoluene	500	0	480	ug/L	96				50	142	
Hexachlorobenzene	500	0	600	ug/L	120	*			72	115	
Pentachlorophenol	1000	0	890	ug/L	89				25	139	