

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P5291

Client: PARSONS Main of New York, Inc.

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	P5291-13MSD	Client Sample ID:	WC-20241213MSD					DataFile:	BF140927.D		
Pyridine	500	0	290	ug/L	58		3		10	109	20
1,4-Dichlorobenzene	500	0	540	ug/L	108		6		55	125	20
2-Methylphenol	500	0	530	ug/L	106		4		37	126	20
3+4-Methylphenols	500	0	530	ug/L	106		2		31	127	20
Hexachloroethane	500	0	540	ug/L	108		2		49	110	20
Nitrobenzene	500	0	520	ug/L	104		4		62	112	20
Hexachlorobutadiene	500	0	540	ug/L	108		8		52	125	20
2,4,6-Trichlorophenol	500	0	500	ug/L	100		4		78	112	20
2,4,5-Trichlorophenol	500	0	470	ug/L	94		4		71	111	20
2,4-Dinitrotoluene	500	0	510	ug/L	102		6		50	142	20
Hexachlorobenzene	500	0	600	ug/L	120	*	0		72	115	20
Pentachlorophenol	1000	0	940	ug/L	94		5		25	139	20