

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5291

Client: PARSONS Main of New York, Inc.

Analytical Method: 8270E DataFile: BF140920.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB165719BS	Pyridine	50	45.2	ug/L	90				29	97	
	1,4-Dichlorobenzene	50	49.2	ug/L	98				76	103	
	2-Methylphenol	50	45.7	ug/L	91				69	109	
	3+4-Methylphenols	50	47.0	ug/L	94				67	106	
	Hexachloroethane	50	43.4	ug/L	87				76	118	
	Nitrobenzene	50	46.0	ug/L	92				58	106	
	Hexachlorobutadiene	50	47.1	ug/L	94				69	101	
	2,4,6-Trichlorophenol	50	50.4	ug/L	101				61	110	
	2,4,5-Trichlorophenol	50	48.5	ug/L	97				70	106	
	2,4-Dinitrotoluene	50	49.4	ug/L	99				60	115	
	Hexachlorobenzene	50	53.6	ug/L	107		*		73	106	
	Pentachlorophenol	100	95.1	ug/L	95				47	114	