

Report of Analysis

Client:	PARSONS Main of New York, Inc.		Date Collected:	
Project:	Con Edison Non-MGP - Atlantic Avenue		Date Received:	
Client Sample ID:	VN1217WBS02		SDG No.:	P5291
Lab Sample ID:	VN1217WBS02		Matrix:	TCLP
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	TCLP VOA
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085222.D	1		12/17/24 14:09	VN121724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	21.8		0.34	1.00	ug/L
75-35-4	1,1-Dichloroethene	19.4		0.26	1.00	ug/L
78-93-3	2-Butanone	94.9		1.30	5.00	ug/L
56-23-5	Carbon Tetrachloride	21.9		0.25	1.00	ug/L
67-66-3	Chloroform	18.9		0.26	1.00	ug/L
71-43-2	Benzene	20.4		0.16	1.00	ug/L
107-06-2	1,2-Dichloroethane	20.2		0.24	1.00	ug/L
79-01-6	Trichloroethene	20.6		0.32	1.00	ug/L
127-18-4	Tetrachloroethene	20.8		0.25	1.00	ug/L
108-90-7	Chlorobenzene	19.1		0.13	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	49.8		74 - 125	100%	SPK: 50
1868-53-7	Dibromofluoromethane	53.0		75 - 124	106%	SPK: 50
2037-26-5	Toluene-d8	51.1		86 - 113	102%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.1		77 - 121	106%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	156000	8.224			
540-36-3	1,4-Difluorobenzene	251000	9.1			
3114-55-4	Chlorobenzene-d5	224000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	110000	13.788			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products