

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5291

Client: PARSONS Main of New York, Inc.

Analytical Method: SW8260-Low

Datafile : VN085222.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1217WBS02	Vinyl chloride	20	21.8	ug/L	109			65	117	
	1,1-Dichloroethene	20	19.4	ug/L	97			74	110	
	2-Butanone	100	94.9	ug/L	95			65	122	
	Carbon Tetrachloride	20	21.9	ug/L	110			77	113	
	Chloroform	20	18.9	ug/L	95			79	113	
	Benzene	20	20.4	ug/L	102			82	109	
	1,2-Dichloroethane	20	20.2	ug/L	101			80	115	
	Trichloroethene	20	20.6	ug/L	103			77	113	
	Tetrachloroethene	20	20.8	ug/L	104			67	123	
	Chlorobenzene	20	19.1	ug/L	96			82	109	