

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PARS02
 Lab Code: CHEM Case No.: P5291 SAS No.: P5291 SDG NO.: P5291
 Lab File ID: VN085218.D Date Analyzed: 12/17/2024
 Instrument ID: MSVOA_N Time Analyzed: 12:10
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	147251	8.22	228600	9.10	208199	11.87
UPPER LIMIT	294502	8.724	457200	9.6	416398	12.365
LOWER LIMIT	73625.5	7.724	114300	8.6	104100	11.365
EPA SAMPLE NO.						
MW6-20241213MS	150098	8.22	255819	9.10	227606	11.87
MW6-20241213MSD	153484	8.22	255174	9.10	228107	11.87
WC-20241213	147129	8.22	259253	9.10	224725	11.87
VN1217WBL01	142158	8.22	245850	9.10	212586	11.87
VN1217WBS02	156058	8.22	250626	9.10	224078	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.