

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PARS02
 Lab Code: CHEM Case No.: P5291 SAS No.: P5291 SDG NO.: P5291
 Lab File ID: VN085218.D Date Analyzed: 12/17/2024
 Instrument ID: MSVOA_N Time Analyzed: 12:10
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	110571	13.788				
UPPER LIMIT	221142	14.288				
LOWER LIMIT	55285.5	13.288				
EPA SAMPLE NO.						
MW6-20241213MS	111390	13.79				
MW6-20241213MSD	107049	13.79				
WC-20241213	99799	13.79				
VN1217WBL01	80399	13.79				
VN1217WBS02	109988	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.