

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PARS02
Lab Code: CHEM Case No.: P5291 SAS No.: P5291 SDG NO.: P5291
Lab File ID: VN085246.D Date Analyzed: 12/18/2024
Instrument ID: MSVOA_N Time Analyzed: 12:20
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	130776	8.22	226121	9.10	202175	11.87
UPPER LIMIT	261552	8.718	452242	9.6	404350	12.365
LOWER LIMIT	65388	7.718	113061	8.6	101088	11.365
EPA SAMPLE NO.						
MW2-20241213DL	99052	8.23	179932	9.10	153337	11.87
MW3-20241213DL	93760	8.22	172552	9.10	156849	11.87
MW6-20241213DL	100325	8.22	186563	9.10	158147	11.87
MW6-20241213-ADL	96192	8.22	181407	9.10	156548	11.86
EQUIP-BLANK-20241213	102235	8.22	187663	9.10	161222	11.87
VN1218WBL01	94664	8.22	169703	9.10	150910	11.87
VN1218WBS01	104586	8.22	185338	9.10	167159	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.