

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PARS02
Lab Code: CHEM Case No.: P5291 SAS No.: P5291 SDG NO.: P5291
Lab File ID: VN085218.D Date Analyzed: 12/17/2024
Instrument ID: MSVOA_N Time Analyzed: 12:10
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	110571	13.788				
UPPER LIMIT	221142	14.288				
LOWER LIMIT	55285.5	13.288				
EPA SAMPLE NO.						
MW1A-20241213	95605	13.79				
MW10D-20241213	104054	13.79				
MW4-20241213	91999	13.79				
MW5-20241213	86832	13.79				
MW2-20241213	108758	13.79				
MW3-20241213	108555	13.79				
MW6-20241213	111275	13.79				
MW6-20241213MS	111390	13.79				
MW6-20241213MSD	107049	13.79				
MW6-20241213-A	107107	13.79				
TRIP BLANK	101465	13.79				
WC-20241213	99799	13.79				
VN1217WBL01	80399	13.79				
VN1217WBS02	109988	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.