

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PARS02
 Lab Code: CHEM Case No.: P5291 SAS No.: P5291 SDG NO.: P5291
 Lab File ID: VN085246.D Date Analyzed: 12/18/2024
 Instrument ID: MSVOA_N Time Analyzed: 12:20
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	98582	13.788				
UPPER LIMIT	197164	14.288				
LOWER LIMIT	49291	13.288				
EPA SAMPLE NO.						
MW2-20241213DL	70062	13.79				
MW3-20241213DL	73363	13.79				
MW6-20241213DL	68008	13.79				
MW6-20241213-ADL	70463	13.79				
EQUIP-BLANK-20241213	62465	13.79				
VN1218WBL01	59088	13.79				
VN1218WBS01	82237	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.