

Process Report

Order ID : P5293

Date	Bottle ID	Lab Sample Number	InComing Vacuum	Canister ID	Can Size	Reg ID	Reg Size	Batch Code	Certification ID
12/17/2024	B2409035	P5293-01	-3.1	10604	6 L	10713		C2408003	VL041543.D Seq :vl081924 TUNE : VL041539.D ICAL : VL041404.D CCAL : VL041540.D MB : VL041541.D
12/17/2024	B2409035	P5293-03	-7.5	10258	6 L	10206		C2409002	VL041664.D Seq :vl101424 TUNE : VL041659.D ICAL : VL041597.D CCAL : VL041660.D MB : VL041661.D
12/17/2024	B2409035	P5293-02	-7.1	10153	6 L	10203		C2409003	VL041615.D Seq :vl091824 TUNE : VL041600.D ICAL : VL041597.D CCAL : VL041601.D MB : VL041602.D



Canister Cleaning

284 Sheffield Street, Mountainside, New Jersey 07092 Phone : 908 789 8900 Fax : 908 789 8922

Canister Batch : C2409002

Type of Canister : 6 L

Cleaning Date : 09/04/2024

Analyst : SAM

Oven Temperature : 125.0 c

Canister	Vaccum After Cleaning ("Hg)	Vaccum After Cleaning Date	Analysis Date	VAC 24Hr ("Hg)	Vaccum After 24hr Date	New Sample Number	QC Canister	File
10334	-30	09/05/2024	09/18/2024	-30	09/06/2024	P5298-02	10334	VL041666.D
10281	-30	09/05/2024	09/18/2024	-30	09/06/2024		10334	VL041614.D
10258	-30	09/05/2024	09/18/2024	-30	09/06/2024	P5293-03	10258	VL041664.D
10332	-30	09/05/2024	09/18/2024	-30	09/06/2024		10334	VL041614.D
10590	-30	09/05/2024	09/18/2024	-30	09/06/2024	P5298-03	10590	VL041665.D
10401	-30	09/05/2024	09/18/2024	-30	09/06/2024	P5294-02	10401	VL041614.D
10607	-30	09/05/2024	09/18/2024	-30	09/06/2024		10334	VL041614.D
10323	-30	09/05/2024	09/18/2024	-30	09/06/2024		10334	VL041614.D



Canister Cleaning

284 Sheffield Street, Mountainside, New Jersey 07092 Phone : 908 789 8900 Fax : 908 789 8922

Canister Batch : C2408003

Type of Canister : 6 L

Cleaning Date : 08/13/2024

Analyst : SAM

Oven Temperature : 125.0 c

Canister	Vaccum After Cleaning ("Hg)	Vaccum After Cleaning Date	Analysis Date	VAC 24Hr ("Hg)	Vaccum After 24hr Date	New Sample Number	QC Canister	File
10604	-30	08/14/2024	08/19/2024	-30	08/15/2024	P5293-01	10604	VL041543.D
10445	-30	08/14/2024	08/19/2024	-30	08/15/2024	P5150-04	10604	VL041543.D
10609	-30	08/14/2024	08/19/2024	-30	08/15/2024	P5150-05	10604	VL041543.D
10283	-30	08/14/2024	08/19/2024	-30	08/15/2024	P4696-02	10604	VL041543.D
10597	-30	08/14/2024	08/19/2024	-30	08/15/2024	P5150-06	10604	VL041543.D
10266	-30	08/14/2024	08/19/2024	-30	08/15/2024		10604	VL041543.D
10158	-30	08/14/2024	08/19/2024	-30	08/15/2024	P5151-01	10604	VL041543.D
10295	-30	08/14/2024	08/19/2024	-30	08/15/2024	P4696-01	10604	VL041543.D



Canister Cleaning

284 Sheffield Street, Mountainside, New Jersey 07092 Phone : 908 789 8900 Fax : 908 789 8922

Canister Batch : C2409003

Type of Canister : 6 L

Cleaning Date : 09/05/2024

Analyst : SAM

Oven Temperature : 125.0 c



Canister	Vaccum After Cleaning ("Hg)	Vaccum After Cleaning Date	Analysis Date	VAC 24Hr ("Hg)	Vaccum After 24hr Date	New Sample Number	QC Canister	File
10285	-30	09/06/2024	09/18/2024	-30	09/09/2024		10153	VL041615.D
10269	-30	09/06/2024	09/18/2024	-30	09/09/2024		10153	VL041615.D
10153	-30	09/06/2024	09/18/2024	-30	09/09/2024	P5293-02	10153	VL041615.D
10198	-30	09/06/2024	09/18/2024	-30	09/09/2024		10153	VL041615.D
10271	-30	09/06/2024	09/18/2024	-30	09/09/2024		10153	VL041615.D
10331	-30	09/06/2024	09/18/2024	-30	09/09/2024		10153	VL041615.D
10592	-30	09/06/2024	09/18/2024	-30	09/09/2024		10153	VL041615.D
10275	-30	09/06/2024	09/18/2024	-30	09/09/2024		10153	VL041615.D

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072524\
Data File : VL041397.D
Acq On : 25 Jul 2024 12:29
Operator : SY/MD
Sample : BFB
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
Last Update : Fri Jul 26 07:31:14 2024



AutoFind: Scans 3472, 3473, 3474; Background Corrected with Scan 3447

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	8	40	25.1	226560	PASS
75	95	30	66	57.2	517099	PASS
95	95	100	100	100.0	904149	PASS
96	95	5	9	6.8	61827	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	69.5	628715	PASS
175	174	4	9	7.9	49563	PASS
176	174	93	101	97.4	612373	PASS
177	176	5	9	6.5	39536	PASS

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\

Method File : VL072524AIR.M

Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022

Last Update : Fri Jul 26 07:31:14 2024

Response Via : Initial Calibration

Calibration Files

0.03=VL041404.D 0.1 =VL041403.D 0.5 =VL041398.D 1 =VL041401.D 2 =VL041400.D 10 =VL041399.D 15 =VL041405.D

Compound	0.03	0.1	0.5	1	2	10	15	Avg	%RSD

1) I Bromochloromethane	-----ISTD-----								
2) T Dichlorodifluo...			1.883	2.019	2.126	1.425	1.148	1.720	24.23
3) Chlorodifluoro...			1.302	1.262	1.264	1.070	1.054	1.190	9.97
4) Chloromethane			0.682	0.685	0.687	0.564	0.548	0.633	11.12
5) T Vinyl Chloride	0.719	0.563	0.628	0.663	0.653	0.570	0.571	0.624	9.45
6) T Bromomethane			0.331	0.344	0.333	0.295	0.301	0.321	6.69
7) Chloroethane			0.240	0.210	0.226	0.202	0.198	0.215	8.14
8) T Dichlorotetra...			1.593	1.646	1.681	1.381	1.321	1.524	10.69
9) T Propene			0.576	0.586	0.588	0.489	0.495	0.547	9.16
10) T Heptane			1.139	1.240	1.371	1.200	1.194	1.229	7.09
11) T Trichlorofluor...			1.621	1.638	1.636	1.383	1.362	1.528	9.32
12) T 1,1,2-Trichlor...			1.154	1.154	1.201	0.999	0.992	1.100	8.85
13) Ethanol			0.044	0.066	0.065	0.035	0.035	0.049#	31.91
14) T Bromoethene			0.417	0.430	0.468	0.412	0.417	0.429	5.27
15) T Acetone			1.292	1.192	1.231	0.996	0.969	1.136	12.76
16) T 1,3-Butadiene			0.605	0.652	0.664	0.556	0.558	0.607	8.38
17) tert-Butyl alc...			1.032	1.100	0.986	1.053	1.011	1.036	4.17
18) T 1,1-Dichloroet...			0.460	0.479	0.515	0.455	0.444	0.471	5.93
19) T Isopropyl Alcohol			0.711	0.674	0.661	0.646	0.630	0.664	4.64
20) T Methylene Chlo...			0.502	0.430	0.444	0.371	0.372	0.424	12.91
21) T Allyl Chloride			0.714	0.783	0.796	0.726	0.707	0.745	5.53
22) T trans-1,2-Dich...			0.462	0.481	0.516	0.448	0.446	0.471	6.16
23) T Vinyl Acetate			0.777	0.874	0.939	0.884	0.865	0.868	6.71
24) T 1,1-Dichloroet...			1.083	1.045	1.116	0.941	0.923	1.022	8.40
25) T Ethyl Acetate			2.177	2.381	2.517	2.116	2.094	2.257	8.16
26) T Hexane			0.897	0.951	1.056	0.889	0.889	0.936	7.64
27) T Carbon Disulfide			1.055	1.020	1.160	1.120	1.122	1.095	5.18
28) T Methyl tert-Bu...			0.565	0.637	0.652	0.555	0.543	0.591	8.53
29) T Chloroform			1.554	1.612	1.659	1.414	1.398	1.527	7.65
30) T Cyclohexane			0.602	0.795	0.875	0.764	0.777	0.763	13.04
31) T cis-1,2-Dichlo...			0.879	0.937	1.022	0.890	0.896	0.925	6.35
32) T 1,1,1-Trichlor...	1.713	1.310	1.498	1.517	1.581	1.398	1.388	1.487	9.11

33) I 1,4-Difluorobenzene	-----ISTD-----								
34) T 2-Butanone			0.465	0.491	0.528	0.467	0.467	0.484	5.58
35) T Carbon Tetrach...	0.584	0.500	0.542	0.558	0.589	0.548	0.540	0.552	5.42
36) T Benzene			0.691	0.778	0.792	0.705	0.689	0.731	6.82
37) T 1,2-Dichloroet...			0.413	0.416	0.428	0.393	0.396	0.409	3.55
38) T Trichloroethene	0.394	0.288	0.326	0.349	0.349	0.320	0.317	0.335	9.99
39) T 1,2-Dichloropr...			0.267	0.294	0.303	0.266	0.269	0.280	6.21

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\ Method File : VL072524AIR.M										
40) T	1,4-Dioxane			0.130	0.110	0.106	0.115	0.101	0.112	9.70
41) T	Tetrahydrofuran			0.226	0.291	0.308	0.278	0.280	0.277	11.09
42) T	Bromodichlorom...			0.573	0.613	0.622	0.576	0.568	0.591	4.26
43) T	Methyl Methacr...			0.205	0.252	0.287	0.283	0.284	0.262	13.30
44) T	2,2,4-Trimethy...			1.168	1.294	1.342	1.165	1.132	1.220	7.54
45) T	t-1,3-Dichloro...			0.166	0.202	0.248	0.287	0.302	0.241	23.76
46) T	cis-1,3-Dichlo...			0.270	0.325	0.372	0.382	0.392	0.348	14.53
47) T	1,1,2-Trichlor...			0.310	0.313	0.334	0.294	0.291	0.308	5.62
48) T	Dibromochlorom...			0.447	0.503	0.526	0.505	0.503	0.497	5.97
49) T	Bromoform			0.354	0.401	0.455	0.446	0.443	0.420	10.10
50) T	4-Methyl-2-Pen...			0.644	0.773	0.860	0.776	0.763	0.763	10.08
51) T	2-Hexanone			0.398	0.569	0.640	0.639	0.640	0.577	18.10
52) T	Tetrachloroethene	0.308	0.260	0.275	0.289	0.304	0.271	0.270	0.282	6.49
53) T	Toluene			0.678	0.851	0.889	0.829	0.822	0.814	9.86
54) T	1,2-Dibromoethane	0.340		0.377	0.424	0.463	0.429	0.429	0.410	10.74
-----ISTD-----										
55) I	Chlorobenzene-d5									
56) T	1,1,1,2-Tetrac...			0.401	0.401	0.412	0.372	0.353	0.388	6.36
57) T	Chlorobenzene			0.747	0.762	0.758	0.667	0.626	0.712	8.71
58) T	Ethyl Benzene			1.028	1.229	1.262	1.151	1.092	1.153	8.35
59) T	m/p-Xylene			0.934	1.065	1.083	0.979	0.914	0.995	7.68
60) T	o-Xylene			0.842	1.015	1.038	0.931	0.880	0.941	8.96
61) T	Styrene			0.313	0.424	0.468	0.464	0.455	0.425	15.32
62) T	Isopropylbenzene			1.404	1.552	1.560	1.368	1.270	1.431	8.68
63) T	1,1,2,2-Tetrac...	0.575	0.568	0.626	0.641	0.650	0.569	0.533	0.594	7.45
64) T	n-propylbenzene			0.353	0.393	0.408	0.370	0.347	0.374	6.89
65) T	tert-Butylbenzene			1.229	1.412	1.448	1.235	1.138	1.292	10.21
66) T	Benzyl Chloride			0.069	0.076	0.080	0.092	0.097	0.083	13.50
67) T	sec-Butylbenzene			1.641	1.940	2.017	1.719	1.586	1.781	10.59
68) S	1-Bromo-4-Fluo...	0.772	0.783	0.774	0.764	0.764	0.782	0.747	0.770	1.63
69) T	p-Isopropyltol...			1.299	1.617	1.674	1.453	1.346	1.478	11.11
70) T	n-Butylbenzene			1.347	1.622	1.731	1.488	1.386	1.515	10.64
71) T	2-Chlorotoluene			0.990	1.133	1.154	1.033	0.966	1.055	8.00
72) T	4-Ethyltoluene			0.976	1.143	1.229	1.138	1.054	1.108	8.70
73) T	1,3,5-Trimethy...			0.831	0.987	1.072	0.974	0.909	0.955	9.46
74) T	1,2,4-Trimethy...			1.034	1.199	1.238	1.074	1.003	1.110	9.32
75) T	1,3-Dichlorobe...			0.663	0.751	0.768	0.672	0.634	0.698	8.33
76) T	1,4-Dichlorobe...			0.673	0.710	0.716	0.656	0.623	0.676	5.74
77) T	1,2-Dichlorobe...			0.692	0.731	0.767	0.654	0.614	0.692	8.73
78) T	Hexachloro-1,3...			0.621	0.680	0.665	0.499	0.484	0.590	15.69
79) T	Naphthalene	0.490		0.808	1.106	1.252	1.125	1.096	0.980	28.67
80) T	Naphthalene,2-...			0.499	0.667	0.911	0.992	1.036	0.821	27.97
81) T	1,2,4-Trichlor...			0.504	0.606	0.646	0.587	0.573	0.583	8.95

(#) = Out of Range

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072524\
 Data File : VL041398.D
 Acq On : 25 Jul 2024 13:19
 Operator : SY/MD
 Sample : VSTDICC0.5
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.5

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 07/26/2024
 Supervised By :Mahesh Dadoda 07/29/2024

Quant Time: Jul 26 05:07:21 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Fri Jul 26 05:07:02 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.176	49	1163290	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.639	114	3245422	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.458	117	3043708	10.000	ppbv	# 0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 13.770 95 2357044 10.071 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 100.700%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.709	85	109541	0.539	ppbv	98
3) Chlorodifluoromethane	2.655	51	75741	0.549	ppbv	100
4) Chloromethane	2.793	50	39649	0.546	ppbv	96
5) Vinyl Chloride	2.906	62	36538	0.508	ppbv #	87
6) Bromomethane	3.102	94	19255	0.523	ppbv	93
7) Chloroethane	3.179	64	13984	0.572	ppbv	92
8) Dichlorotetrafluoroethane	2.838	85	92652	0.521	ppbv	93
9) Propene	2.671	41	33491	0.523	ppbv	98
10) Heptane	7.568	43	66249	0.465	ppbv	97
11) Trichlorofluoromethane	3.545	101	94267	0.536	ppbv	100
12) 1,1,2-Trichlorotrifluo...	4.050	101	67103	0.527	ppbv	99
13) Ethanol	3.237	45	2541m	0.454	ppbv	
14) Bromoethene	3.346	108	24275	0.495	ppbv #	90
15) Acetone	3.462	43	75165	0.574	ppbv	100
16) 1,3-Butadiene	2.963	39	35196m	0.505	ppbv	
17) tert-Butyl alcohol	3.870	59	59997	0.504	ppbv	99
18) 1,1-Dichloroethene	3.864	96	26744	0.484	ppbv	90
19) Isopropyl Alcohol	3.568	45	41335m	0.559	ppbv	
20) Methylene Chloride	3.915	84	29173	0.603	ppbv #	93
21) Allyl Chloride	3.979	41	41531	0.483	ppbv	94
22) trans-1,2-Dichloroethene	4.433	96	26864	0.500	ppbv	93
23) Vinyl Acetate	4.616	43	45209	0.444	ppbv #	97
24) 1,1-Dichloroethane	4.545	63	62999	0.542	ppbv #	97
25) Ethyl Acetate	5.195	43	126641	0.485	ppbv	99
26) Hexane	5.192	57	52190	0.475	ppbv	91
27) Carbon Disulfide	4.111	76	61342	0.489	ppbv #	93
28) Methyl tert-Butyl Ether	4.581	73	32882	0.475	ppbv	95
29) Chloroform	5.256	83	90359	0.503	ppbv	99
30) Cyclohexane	6.594	84	35035	0.389	ppbv #	88
31) cis-1,2-Dichloroethene	5.063	61	51136	0.472	ppbv #	87
32) 1,1,1-Trichloroethane	5.989	97	87155	0.508	ppbv	97
34) 2-Butanone	4.787	43	75522m	0.481	ppbv	
35) Carbon Tetrachloride	6.478	117	88027	0.495	ppbv	86
36) Benzene	6.356	78	112067	0.470	ppbv	95
37) 1,2-Dichloroethane	5.793	62	66949	0.507	ppbv	98
38) Trichloroethene	7.282	130	52825	0.488	ppbv	98
39) 1,2-Dichloropropane	7.060	63	43368	0.473	ppbv	95
40) 1,4-Dioxane	7.304	88	21057m	0.633	ppbv	
41) Tetrahydrofuran	5.568	42	36699	0.403	ppbv	98
42) Bromodichloromethane	7.237	83	92906	0.486	ppbv	98
43) Methyl Methacrylate	7.465	69	33258	0.386	ppbv	91
44) 2,2,4-Trimethylpentane	7.314	57	189569	0.479	ppbv	94
45) t-1,3-Dichloropropene	8.700	75	26864m	0.341	ppbv	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072524\
 Data File : VL041398.D
 Acq On : 25 Jul 2024 13:19
 Operator : SY/MD
 Sample : VSTDICC0.5
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.5

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 07/26/2024
 Supervised By :Mahesh Dadoda 07/29/2024

Quant Time: Jul 26 05:07:21 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Jul 26 05:07:02 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	8.134	75	43823m	0.384	ppbv	
47) 1,1,2-Trichloroethane	8.886	97	50232	0.505	ppbv	95
48) Dibromochloromethane	9.700	129	72480	0.452	ppbv	100
49) Bromoform	12.400	173	57383	0.423	ppbv	95
50) 4-Methyl-2-Pentanone	8.185	43	104559m	0.417	ppbv	
51) 2-Hexanone	9.558	43	64659	0.338	ppbv #	97
52) Tetrachloroethene	10.600	164	44583	0.486	ppbv	92
53) Toluene	9.211	91	110057	0.411	ppbv	98
54) 1,2-Dibromoethane	10.002	107	61242	0.455	ppbv	89
56) 1,1,1,2-Tetrachloroethane	11.497	131	61039	0.519	ppbv #	1
57) Chlorobenzene	11.516	112	113739	0.524	ppbv	94
58) Ethyl Benzene	12.079	91	156513	0.441	ppbv	96
59) m/p-Xylene	12.365	91	284326m	0.936	ppbv	
60) o-Xylene	13.050	91	128195	0.440	ppbv	99
61) Styrene	12.889	104	47561	0.362	ppbv	94
62) Isopropylbenzene	14.018	105	213603m	0.492	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.040	83	95214	0.526	ppbv	97
64) n-propylbenzene	14.915	120	53748m	0.473	ppbv	
65) tert-Butylbenzene	16.082	119	187031m	0.476	ppbv	
66) Benzyl Chloride	16.323	91	10564m	0.422	ppbv	
67) sec-Butylbenzene	16.632	105	249741	0.458	ppbv	99
69) p-Isopropyltoluene	16.992	119	197625	0.434	ppbv	96
70) n-Butylbenzene	17.886	91	204999	0.443	ppbv	98
71) 2-Chlorotoluene	14.796	91	150717	0.465	ppbv	100
72) 4-Ethyltoluene	15.188	105	148551	0.441	ppbv	97
73) 1,3,5-Trimethylbenzene	15.346	105	126480	0.429	ppbv	92
74) 1,2,4-Trimethylbenzene	16.098	105	157396m	0.463	ppbv	
75) 1,3-Dichlorobenzene	16.326	146	100938	0.476	ppbv	94
76) 1,4-Dichlorobenzene	16.468	146	102368m	0.492	ppbv	
77) 1,2-Dichlorobenzene	17.143	146	105386m	0.503	ppbv	
78) Hexachloro-1,3-Butadiene	22.631	225	94555m	0.520	ppbv	
79) Naphthalene	21.387	128	123028m	0.412	ppbv	
80) Naphthalene,2-methyl-	24.374	142	75989	0.313	ppbv	93
81) 1,2,4-Trichlorobenzene	21.143	180	76641m	0.428	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072524\
 Data File : VL041399.D
 Acq On : 25 Jul 2024 13:59
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Quant Time: Jul 26 05:05:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri Jul 26 04:51:44 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 07/26/2024
 Supervised By : Mahesh Dadoda 07/29/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.176	49	1227171	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.645	114	3353351	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.465	117	3260833m	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.770	95	2550034	10.170	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.700%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.706	85	1748662	8.150	ppbv	100
3) Chlorodifluoromethane	2.652	51	1312868	9.025	ppbv	100
4) Chloromethane	2.790	50	692256	9.024	ppbv	99
5) Vinyl Chloride	2.899	62	699389	9.214	ppbv	100
6) Bromomethane	3.099	94	361665	9.227	ppbv	100
7) Chloroethane	3.176	64	248340	9.630	ppbv	100
8) Dichlorotetrafluoroethane	2.835	85	1694634	9.039	ppbv	100
9) Propene	2.671	41	600396	8.889	ppbv	100
10) Heptane	7.568	43	1472225	9.801	ppbv	99
11) Trichlorofluoromethane	3.542	101	1696907	9.145	ppbv	100
12) 1,1,2-Trichlorotrifluo...	4.047	101	1226328	9.136	ppbv	100
13) Ethanol	3.224	45	43077m	7.073	ppbv	
14) Bromoethene	3.346	108	505960	9.786	ppbv	100
15) Acetone	3.446	43	1222864	8.859	ppbv	100
16) 1,3-Butadiene	2.964	39	681843	9.269	ppbv	100
17) tert-Butyl alcohol	3.861	59	1291681	10.276	ppbv	100
18) 1,1-Dichloroethene	3.861	96	558470	9.575	ppbv	100
19) Isopropyl Alcohol	3.555	45	792740	10.155	ppbv	100
20) Methylene Chloride	3.915	84	455611	8.929	ppbv	100
21) Allyl Chloride	3.980	41	890718	9.827	ppbv	100
22) trans-1,2-Dichloroethene	4.430	96	549902	9.711	ppbv	100
23) Vinyl Acetate	4.610	43	1085315	10.104	ppbv	100
24) 1,1-Dichloroethane	4.549	63	1154644	9.415	ppbv	100
25) Ethyl Acetate	5.189	43	2597295	9.430	ppbv	100
26) Hexane	5.198	57	1090849	9.420	ppbv	100
27) Carbon Disulfide	4.105	76	1375021	10.384	ppbv	99
28) Methyl tert-Butyl Ether	4.568	73	681033	9.329	ppbv	100
29) Chloroform	5.259	83	1735318	9.163	ppbv	100
30) Cyclohexane	6.590	84	937363	9.878	ppbv	100
31) cis-1,2-Dichloroethene	5.063	61	1091838	9.548	ppbv	100
32) 1,1,1-Trichloroethane	5.989	97	1715896	9.483	ppbv	100
34) 2-Butanone	4.771	43	1567685	9.654	ppbv	100
35) Carbon Tetrachloride	6.484	117	1836059	9.983	ppbv	100
36) Benzene	6.359	78	2363293	9.591	ppbv	100
37) 1,2-Dichloroethane	5.796	62	1317693	9.664	ppbv	99
38) Trichloroethene	7.282	130	1074518	9.605	ppbv	100
39) 1,2-Dichloropropane	7.063	63	891517	9.403	ppbv	100
40) 1,4-Dioxane	7.275	88	384473	11.177	ppbv	100
41) Tetrahydrofuran	5.542	42	931064	9.892	ppbv	100
42) Bromodichloromethane	7.240	83	1932798	9.793	ppbv	100
43) Methyl Methacrylate	7.465	69	948521	10.646	ppbv	100
44) 2,2,4-Trimethylpentane	7.317	57	3907968	9.554	ppbv	100
45) t-1,3-Dichloropropene	8.703	75	963383	11.835	ppbv	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072524\
 Data File : VL041399.D
 Acq On : 25 Jul 2024 13:59
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 07/26/2024
 Supervised By : Mahesh Dadoda 07/29/2024

Quant Time: Jul 26 05:05:14 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Jul 26 04:51:44 2024


Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	8.134	75	1281640	10.855	ppbv	100
47) 1,1,2-Trichloroethane	8.893	97	987016	9.602	ppbv	100
48) Dibromochloromethane	9.706	129	1692234	10.224	ppbv	100
49) Bromoform	12.404	173	1496236	10.670	ppbv	99
50) 4-Methyl-2-Pentanone	8.172	43	2603717	10.051	ppbv	100
51) 2-Hexanone	9.542	43	2141345	10.819	ppbv	100
52) Tetrachloroethene	10.613	164	907950	9.585	ppbv	100
53) Toluene	9.214	91	2779504	10.036	ppbv	100
54) 1,2-Dibromoethane	10.002	107	1438535	10.342	ppbv	100
56) 1,1,1,2-Tetrachloroethane	11.503	131	1213659	9.640	ppbv #	58
57) Chlorobenzene	11.523	112	2174265	9.343	ppbv	100
58) Ethyl Benzene	12.085	91	3752771	9.870	ppbv	100
59) m/p-Xylene	12.368	91	6382118m	19.609	ppbv	
60) o-Xylene	13.053	91	3036791	9.732	ppbv	100
61) Styrene	12.896	104	1513410	10.748	ppbv	100
62) Isopropylbenzene	14.024	105	4462038	9.595	ppbv	100
63) 1,1,2,2-Tetrachloroethane	13.040	83	1853886	9.567	ppbv	100
64) n-propylbenzene	14.912	120	1205437	9.895	ppbv	100
65) tert-Butylbenzene	16.085	119	4027165	9.564	ppbv	100
66) Benzyl Chloride	16.323	91	299511	11.162	ppbv	100
67) sec-Butylbenzene	16.635	105	5605419	9.602	ppbv	100
69) p-Isopropyltoluene	16.989	119	4737222	9.713	ppbv	100
70) n-Butylbenzene	17.883	91	4851380	9.796	ppbv	100
71) 2-Chlorotoluene	14.799	91	3368997	9.707	ppbv	100
72) 4-Ethyltoluene	15.188	105	3711228	10.293	ppbv	99
73) 1,3,5-Trimethylbenzene	15.346	105	3175291	10.049	ppbv	100
74) 1,2,4-Trimethylbenzene	16.105	105	3502689	9.621	ppbv	100
75) 1,3-Dichlorobenzene	16.330	146	2192883	9.642	ppbv	100
76) 1,4-Dichlorobenzene	16.471	146	2140459	9.596	ppbv	100
77) 1,2-Dichlorobenzene	17.140	146	2131506	9.501	ppbv	100
78) Hexachloro-1,3-Butadiene	22.628	225	1626038	8.347	ppbv	100
79) Naphthalene	21.381	128	3669460	11.481	ppbv	100
80) Naphthalene,2-methyl-	24.371	142	3235986	12.454	ppbv	100
81) 1,2,4-Trichlorobenzene	21.133	180	1912951	9.962	ppbv	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Manual Integrations APPROVED

Quant Time: Jul 26 05:05:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Fri Jul 26 04:51:44 2024
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072524\
 Data File : VL041400.D
 Acq On : 25 Jul 2024 14:40
 Operator : SY/MD
 Sample : VSTDICC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC002

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 07/26/2024
 Supervised By :Mahesh Dadoda 07/29/2024

Quant Time: Jul 25 15:37:54 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Thu Jul 25 14:31:18 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.176	49	1187869	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.642	114	3361544	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.462	117	3300753	10.000	ppbv	# 0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 13.773 95 2521024 10.190 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 101.900%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.710	85	505198	2.492	ppbv	99
3) Chlorodifluoromethane	2.652	51	300375	1.467	ppbv	99
4) Chloromethane	2.793	50	163139	2.084	ppbv	96
5) Vinyl Chloride	2.899	62	155185	1.835	ppbv	99
6) Bromomethane	3.102	94	79103	2.956	ppbv	98
7) Chloroethane	3.176	64	53717	1.922	ppbv	94
8) Dichlorotetrafluoroethane	2.835	85	399415	2.027	ppbv	100
9) Propene	2.671	41	139773	1.972	ppbv	99
10) Heptane	7.568	43	325614	1.956	ppbv	100
11) Trichlorofluoromethane	3.546	101	388765	1.970	ppbv	99
12) 1,1,2-Trichlorotrifluo...	4.050	101	285420	2.064	ppbv	100
13) Ethanol	3.237	45	15410m	0.980	ppbv	
14) Bromoethene	3.346	108	111081	2.003	ppbv	93
15) Acetone	3.456	43	292550	1.865	ppbv	99
16) 1,3-Butadiene	2.967	39	157764	2.113	ppbv	99
17) tert-Butyl alcohol	3.861	59	234318	1.936	ppbv	99
18) 1,1-Dichloroethene	3.861	96	122348	2.013	ppbv	97
19) Isopropyl Alcohol	3.562	45	156947m	1.822	ppbv	
20) Methylene Chloride	3.915	84	105428	1.852	ppbv	98
21) Allyl Chloride	3.980	41	189033	1.998	ppbv	90
22) trans-1,2-Dichloroethene	4.427	96	122560	2.030	ppbv	90
23) Vinyl Acetate	4.613	43	223002	1.442	ppbv	# 98
24) 1,1-Dichloroethane	4.546	63	265124	1.977	ppbv	98
25) Ethyl Acetate	5.192	43	597999	2.070	ppbv	99
26) Hexane	5.198	57	250786	1.961	ppbv	98
27) Carbon Disulfide	4.108	76	275637	1.771	ppbv	94
28) Methyl tert-Butyl Ether	4.578	73	154942	1.633	ppbv	97
29) Chloroform	5.259	83	394045	1.997	ppbv	99
30) Cyclohexane	6.587	84	207896	2.040	ppbv	93
31) cis-1,2-Dichloroethene	5.063	61	242910	1.981	ppbv	97
32) 1,1,1-Trichloroethane	5.992	97	375703	1.814	ppbv	97
34) 2-Butanone	4.774	43	355170	2.002	ppbv	98
35) Carbon Tetrachloride	6.481	117	395969	1.784	ppbv	98
36) Benzene	6.359	78	532213	1.913	ppbv	97
37) 1,2-Dichloroethane	5.793	62	287587	1.802	ppbv	98
38) Trichloroethene	7.282	130	234513	1.935	ppbv	93
39) 1,2-Dichloropropane	7.060	63	203420	1.892	ppbv	96
40) 1,4-Dioxane	7.288	88	71274	2.068	ppbv	# 1
41) Tetrahydrofuran	5.552	42	207220	1.886	ppbv	99
42) Bromodichloromethane	7.237	83	418292	1.873	ppbv	98
43) Methyl Methacrylate	7.468	69	192639	1.843	ppbv	94
44) 2,2,4-Trimethylpentane	7.317	57	902570	1.909	ppbv	99
45) t-1,3-Dichloropropene	8.700	75	166952	2.524	ppbv	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072524\
 Data File : VL041400.D
 Acq On : 25 Jul 2024 14:40
 Operator : SY/MD
 Sample : VSTDIC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC002

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 07/26/2024
 Supervised By :Mahesh Dadoda 07/29/2024

Quant Time: Jul 25 15:37:54 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Thu Jul 25 14:31:18 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	8.134	75	249911	2.810	ppbv	97
47) 1,1,2-Trichloroethane	8.889	97	224865	1.950	ppbv	94
48) Dibromochloromethane	9.703	129	353718	1.910	ppbv	100
49) Bromoform	12.401	173	305646	2.073	ppbv	99
50) 4-Methyl-2-Pentanone	8.176	43	577958	2.029	ppbv	99
51) 2-Hexanone	9.555	43	430093	2.124	ppbv #	96
52) Tetrachloroethene	10.606	164	204374	1.829	ppbv	98
53) Toluene	9.214	91	597804	1.905	ppbv	100
54) 1,2-Dibromoethane	10.005	107	311374	2.182	ppbv	98
56) 1,1,1,2-Tetrachloroethane	11.503	131	272178	1.832	ppbv #	1
57) Chlorobenzene	11.520	112	500573	1.933	ppbv	96
58) Ethyl Benzene	12.082	91	833428	1.887	ppbv	96
59) m/p-Xylene	12.365	91	1430508m	3.743	ppbv	
60) o-Xylene	13.053	91	685320	1.900	ppbv	99
61) Styrene	12.896	104	308833	1.763	ppbv	99
62) Isopropylbenzene	14.027	105	1029899	1.850	ppbv	99
63) 1,1,2,2-Tetrachloroethane	13.044	83	428827	1.509	ppbv	98
64) n-propylbenzene	14.912	120	269236	1.852	ppbv	91
65) tert-Butylbenzene	16.085	119	956093	1.928	ppbv	100
66) Benzyl Chloride	16.326	91	52950m	1.625	ppbv	
67) sec-Butylbenzene	16.635	105	1331315	1.898	ppbv	99
69) p-Isopropyltoluene	16.992	119	1105164	1.908	ppbv	100
70) n-Butylbenzene	17.886	91	1142994	1.931	ppbv	99
71) 2-Chlorotoluene	14.799	91	762002	1.842	ppbv	99
72) 4-Ethyltoluene	15.188	105	811504	1.889	ppbv	99
73) 1,3,5-Trimethylbenzene	15.349	105	707827	1.925	ppbv	95
74) 1,2,4-Trimethylbenzene	16.105	105	817315	1.972	ppbv	95
75) 1,3-Dichlorobenzene	16.326	146	506675	2.061	ppbv	99
76) 1,4-Dichlorobenzene	16.468	146	472957	1.931	ppbv	97
77) 1,2-Dichlorobenzene	17.140	146	506084	2.019	ppbv	99
78) Hexachloro-1,3-Butadiene	22.628	225	438766	2.407	ppbv	98
79) Naphthalene	21.384	128	826782	3.710	ppbv	98
80) Naphthalene,2-methyl-	24.371	142	601537	1025.625	ppbv	99
81) 1,2,4-Trichlorobenzene	21.140	180	426225	3.095	ppbv	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072524\
Data File : VL041400.D
Acq On : 25 Jul 2024 14:40
Operator : SY/MD
Sample : VSTDIC002
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC002

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 07/26/2024
Supervised By : Mahesh Dadoda 07/29/2024


Quant Time: Jul 25 15:37:54 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug

QLast Update : Thu Jul 25 14:31:18 2024

Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072524\
 Data File : VL041401.D
 Acq On : 25 Jul 2024 15:19
 Operator : SY/MD
 Sample : VSTDICC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 07/26/2024
 Supervised By : Mahesh Dadoda 07/29/2024

Quant Time: Jul 26 02:43:04 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Thu Jul 25 14:31:18 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.176	49	1198624	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.639	114	3342988	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.459	117	3161618	10.000	ppbv	# 0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 13.774 95 2414921 10.191 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 101.900%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.706	85	241964	1.183	ppbv	95
3) Chlorodifluoromethane	2.652	51	151299	0.732	ppbv	98
4) Chloromethane	2.793	50	82099	1.039	ppbv	96
5) Vinyl Chloride	2.899	62	79527	0.932	ppbv	98
6) Bromomethane	3.099	94	41204	1.526	ppbv	92
7) Chloroethane	3.179	64	25119	0.891	ppbv	92
8) Dichlorotetrafluoroethane	2.835	85	197281	0.992	ppbv	99
9) Propene	2.674	41	70203	0.982	ppbv	97
10) Heptane	7.568	43	148675	0.885	ppbv	96
11) Trichlorofluoromethane	3.542	101	196294	0.986	ppbv	96
12) 1,1,2-Trichlorotrifluo...	4.050	101	138368	0.992	ppbv	99
13) Ethanol	3.240	45	7925m	0.500	ppbv	
14) Bromoethene	3.350	108	51557	0.921	ppbv #	85
15) Acetone	3.456	43	142863	0.902	ppbv	97
16) 1,3-Butadiene	2.967	39	78187	1.038	ppbv	96
17) tert-Butyl alcohol	3.867	59	131833	1.080	ppbv	100
18) 1,1-Dichloroethene	3.861	96	57445	0.937	ppbv	99
19) Isopropyl Alcohol	3.568	45	80830	0.930	ppbv #	98
20) Methylene Chloride	3.912	84	51566	0.898	ppbv	96
21) Allyl Chloride	3.980	41	93864	0.983	ppbv	95
22) trans-1,2-Dichloroethene	4.433	96	57666	0.947	ppbv	91
23) Vinyl Acetate	4.613	43	104758	0.671	ppbv #	94
24) 1,1-Dichloroethane	4.546	63	125314	0.926	ppbv	100
25) Ethyl Acetate	5.195	43	285442	0.979	ppbv	98
26) Hexane	5.198	57	113990	0.883	ppbv	88
27) Carbon Disulfide	4.112	76	122266	0.779	ppbv	97
28) Methyl tert-Butyl Ether	4.575	73	76370	0.798	ppbv	99
29) Chloroform	5.259	83	193219	0.970	ppbv	91
30) Cyclohexane	6.594	84	95243m	0.926	ppbv	
31) cis-1,2-Dichloroethene	5.066	61	112360	0.908	ppbv	96
32) 1,1,1-Trichloroethane	5.986	97	181889	0.870	ppbv	99
34) 2-Butanone	4.777	43	163975	0.930	ppbv	97
35) Carbon Tetrachloride	6.481	117	186609	0.845	ppbv	99
36) Benzene	6.356	78	259942	0.940	ppbv	95
37) 1,2-Dichloroethane	5.793	62	138981	0.875	ppbv	100
38) Trichloroethene	7.279	130	116586m	0.967	ppbv	
39) 1,2-Dichloropropane	7.063	63	98425m	0.921	ppbv	
40) 1,4-Dioxane	7.298	88	36813m	1.074	ppbv	
41) Tetrahydrofuran	5.558	42	97249	0.890	ppbv	98
42) Bromodichloromethane	7.240	83	204988	0.923	ppbv	99
43) Methyl Methacrylate	7.472	69	84268	0.811	ppbv	88
44) 2,2,4-Trimethylpentane	7.317	57	432528	0.920	ppbv	99
45) t-1,3-Dichloropropene	8.706	75	67675	1.029	ppbv	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072524\
 Data File : VL041401.D
 Acq On : 25 Jul 2024 15:19
 Operator : SY/MD
 Sample : VSTDICC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 07/26/2024
 Supervised By :Mahesh Dadoda 07/29/2024

Quant Time: Jul 26 02:43:04 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Thu Jul 25 14:31:18 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	8.134	75	108779	1.230	ppbv	99
47) 1,1,2-Trichloroethane	8.889	97	104479	0.911	ppbv	97
48) Dibromochloromethane	9.703	129	168168	0.913	ppbv	97
49) Bromoform	12.404	173	133988	0.914	ppbv	99
50) 4-Methyl-2-Pentanone	8.182	43	258275	0.912	ppbv	96
51) 2-Hexanone	9.558	43	190138	0.944	ppbv #	100
52) Tetrachloroethene	10.613	164	96753m	0.871	ppbv	
53) Toluene	9.211	91	284562	0.912	ppbv	98
54) 1,2-Dibromoethane	10.005	107	141838	0.999	ppbv	91
56) 1,1,1,2-Tetrachloroethane	11.504	131	126753	0.891	ppbv #	1
57) Chlorobenzene	11.520	112	240871	0.971	ppbv	95
58) Ethyl Benzene	12.086	91	388676	0.919	ppbv	98
59) m/p-Xylene	12.368	91	673526m	1.840	ppbv	
60) o-Xylene	13.053	91	320813	0.929	ppbv	99
61) Styrene	12.896	104	133924	0.798	ppbv	99
62) Isopropylbenzene	14.024	105	490658	0.920	ppbv	97
63) 1,1,2,2-Tetrachloroethane	13.047	83	202749	0.745	ppbv	98
64) n-propylbenzene	14.915	120	124102	0.891	ppbv	83
65) tert-Butylbenzene	16.085	119	446299	0.940	ppbv	99
66) Benzyl Chloride	16.326	91	24142m	0.774	ppbv	
67) sec-Butylbenzene	16.632	105	613358	0.913	ppbv	99
69) p-Isopropyltoluene	16.992	119	511080	0.921	ppbv	100
70) n-Butylbenzene	17.886	91	512783	0.904	ppbv	98
71) 2-Chlorotoluene	14.796	91	358079	0.904	ppbv	100
72) 4-Ethyltoluene	15.188	105	361470	0.878	ppbv	97
73) 1,3,5-Trimethylbenzene	15.349	105	312184	0.886	ppbv	98
74) 1,2,4-Trimethylbenzene	16.101	105	379183	0.955	ppbv #	91
75) 1,3-Dichlorobenzene	16.330	146	237337	1.008	ppbv	99
76) 1,4-Dichlorobenzene	16.468	146	224373	0.956	ppbv	98
77) 1,2-Dichlorobenzene	17.140	146	231111	0.963	ppbv	98
78) Hexachloro-1,3-Butadiene	22.632	225	215079	1.232	ppbv	99
79) Naphthalene	21.387	128	349592m	1.638	ppbv	
80) Naphthalene,2-methyl-	24.374	142	210840	375.304	ppbv	98
81) 1,2,4-Trichlorobenzene	21.143	180	191623	1.452	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072524\
 Data File : VL041403.D
 Acq On : 25 Jul 2024 16:42
 Operator : SY/MD
 Sample : VSTDIC0.1
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.1

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 07/26/2024
 Supervised By :Mahesh Dadoda 07/29/2024

Quant Time: Jul 26 02:44:42 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Thu Jul 25 14:31:18 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.176	49	1207844	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.639	114	3343894	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.458	117	3087423	10.000	ppbv	# 0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.767	95	2418801	10.453	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.500%
Target Compounds						
						Qvalue
5) Vinyl Chloride	2.899	62	6804	0.079	ppbv	94
32) 1,1,1-Trichloroethane	5.989	97	15825m	0.075	ppbv	
35) Carbon Tetrachloride	6.478	117	16719m	0.076	ppbv	
38) Trichloroethene	7.278	130	9639m	0.080	ppbv	
52) Tetrachloroethene	10.593	164	8679m	0.078	ppbv	
54) 1,2-Dibromoethane	9.995	107	11367m	0.080	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.040	83	17530m	0.066	ppbv	
79) Naphthalene	21.403	128	15122m	0.073	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072524\
Data File : VL041403.D
Acq On : 25 Jul 2024 16:42
Operator : SY/MD
Sample : VSTDICC0.1
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 07/26/2024
Supervised By :Mahesh Dadoda 07/29/2024

Quant Time: Jul 26 02:44:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Thu Jul 25 14:31:18 2024
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072524\
 Data File : VL041404.D
 Acq On : 25 Jul 2024 17:24
 Operator : SY/MD
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 07/26/2024
 Supervised By : Mahesh Dadoda 07/29/2024

Quant Time: Jul 26 02:45:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Thu Jul 25 14:31:18 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.176	49	1173154	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.639	114	3272729	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.458	117	3056197	10.000	ppbv	# 0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.767	95	2360153	10.303	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	= 103.000%	
Target Compounds						
					Qvalue	
5) Vinyl Chloride	2.899	62	2531m	0.030	ppbv	
32) 1,1,1-Trichloroethane	5.983	97	6029m	0.029	ppbv	
35) Carbon Tetrachloride	6.478	117	5730m	0.027	ppbv	
38) Trichloroethene	7.275	130	3872m	0.033	ppbv	
52) Tetrachloroethene	10.610	164	3020m	0.028	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.037	83	5269m	0.020	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072524\
Data File : VL041404.D
Acq On : 25 Jul 2024 17:24
Operator : SY/MD
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Quant Time: Jul 26 02:45:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Thu Jul 25 14:31:18 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 07/26/2024
Supervised By : Mahesh Dadoda 07/29/2024



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072524\
 Data File : VL041405.D
 Acq On : 25 Jul 2024 18:06
 Operator : SY/MD
 Sample : VSTDICC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC015

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 07/26/2024
 Supervised By :Mahesh Dadoda 07/29/2024

Quant Time: Jul 26 02:45:39 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Thu Jul 25 14:31:18 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.179	49	1211244	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.642	114	3327019	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.458	117	3365868	10.000	ppbv	# 0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 13.770 95 2515088 9.970 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 99.700%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.706	85	2086477	10.095	ppbv	99
3) Chlorodifluoromethane	2.652	51	1914240	9.165	ppbv	100
4) Chloromethane	2.790	50	996482	12.485	ppbv	99
5) Vinyl Chloride	2.899	62	1038128	12.036	ppbv	98
6) Bromomethane	3.099	94	547488	20.067	ppbv	96
7) Chloroethane	3.176	64	360614	12.655	ppbv	95
8) Dichlorotetrafluoroethane	2.835	85	2399417	11.943	ppbv	99
9) Propene	2.671	41	899840	12.451	ppbv	98
10) Heptane	7.571	43	2169367	12.780	ppbv	99
11) Trichlorofluoromethane	3.542	101	2473823	12.296	ppbv	100
12) 1,1,2-Trichlorotrifluo...	4.047	101	1802452	12.784	ppbv	99
13) Ethanol	3.221	45	62897m	3.923	ppbv	
14) Bromoethene	3.346	108	757933	13.400	ppbv	98
15) Acetone	3.446	43	1760479	11.005	ppbv	99
16) 1,3-Butadiene	2.964	39	1014052	13.319	ppbv	99
17) tert-Butyl alcohol	3.857	59	1836705	14.883	ppbv	99
18) 1,1-Dichloroethene	3.861	96	806522	13.011	ppbv	100
19) Isopropyl Alcohol	3.555	45	1143994	13.027	ppbv	100
20) Methylene Chloride	3.915	84	675282	11.634	ppbv	99
21) Allyl Chloride	3.980	41	1284615	13.313	ppbv	97
22) trans-1,2-Dichloroethene	4.430	96	809963	13.156	ppbv	99
23) Vinyl Acetate	4.613	43	1570687	9.960	ppbv	98
24) 1,1-Dichloroethane	4.545	63	1677304	12.269	ppbv	99
25) Ethyl Acetate	5.189	43	3804371	12.915	ppbv	100
26) Hexane	5.198	57	1615258	12.384	ppbv	98
27) Carbon Disulfide	4.105	76	2038049	12.842	ppbv	99
28) Methyl tert-Butyl Ether	4.568	73	986442	10.196	ppbv	100
29) Chloroform	5.263	83	2540539	12.624	ppbv	99
30) Cyclohexane	6.590	84	1411237	13.581	ppbv	96
31) cis-1,2-Dichloroethene	5.063	61	1628471	13.023	ppbv	100
32) 1,1,1-Trichloroethane	5.989	97	2522136	11.940	ppbv	100
34) 2-Butanone	4.771	43	2330483	13.274	ppbv	99
35) Carbon Tetrachloride	6.481	117	2695636	12.269	ppbv	98
36) Benzene	6.359	78	3438269	12.487	ppbv	100
37) 1,2-Dichloroethane	5.793	62	1975307	12.503	ppbv	100
38) Trichloroethene	7.282	130	1582683	13.195	ppbv	98
39) 1,2-Dichloropropane	7.063	63	1340584	12.599	ppbv	96
40) 1,4-Dioxane	7.269	88	505735	14.825	ppbv	80
41) Tetrahydrofuran	5.542	42	1399340	12.868	ppbv	100
42) Bromodichloromethane	7.240	83	2836611	12.832	ppbv	99
43) Methyl Methacrylate	7.465	69	1415809	13.686	ppbv	100
44) 2,2,4-Trimethylpentane	7.314	57	5650271	12.075	ppbv	99
45) t-1,3-Dichloropropene	8.703	75	1508473	23.041	ppbv	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072524\
 Data File : VL041405.D
 Acq On : 25 Jul 2024 18:06
 Operator : SY/MD
 Sample : VSTDIC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC015

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 07/26/2024
 Supervised By :Mahesh Dadoda 07/29/2024

Quant Time: Jul 26 02:45:39 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Thu Jul 25 14:31:18 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	8.134	75	1956016	22.219	ppbv	96
47) 1,1,2-Trichloroethane	8.893	97	1451859	12.721	ppbv	96
48) Dibromochloromethane	9.706	129	2512083	13.707	ppbv	100
49) Bromoform	12.404	173	2210280	15.143	ppbv	99
50) 4-Methyl-2-Pentanone	8.169	43	3807510	13.505	ppbv	99
51) 2-Hexanone	9.542	43	3193628	15.936	ppbv	100
52) Tetrachloroethene	10.610	164	1349134	12.202	ppbv	98
53) Toluene	9.214	91	4102245	13.210	ppbv	99
54) 1,2-Dibromoethane	10.002	107	2140365	15.154	ppbv	96
56) 1,1,1,2-Tetrachloroethane	11.503	131	1780102	11.753	ppbv #	58
57) Chlorobenzene	11.523	112	3158639	11.963	ppbv	98
58) Ethyl Benzene	12.082	91	5512808	12.242	ppbv	99
59) m/p-Xylene	12.365	91	9224976m	23.671	ppbv	
60) o-Xylene	13.053	91	4441698	12.076	ppbv	100
61) Styrene	12.892	104	2297760	12.864	ppbv	99
62) Isopropylbenzene	14.027	105	6413793	11.295	ppbv	99
63) 1,1,2,2-Tetrachloroethane	13.044	83	2691069	9.287	ppbv	99
64) n-propylbenzene	14.915	120	1752978	11.826	ppbv	100
65) tert-Butylbenzene	16.088	119	5746630	11.366	ppbv	99
66) Benzyl Chloride	16.323	91	487926	14.686	ppbv	99
67) sec-Butylbenzene	16.632	105	8009840	11.201	ppbv	98
69) p-Isopropyltoluene	16.992	119	6796381	11.508	ppbv	99
70) n-Butylbenzene	17.883	91	6996243	11.589	ppbv	98
71) 2-Chlorotoluene	14.796	91	4875167	11.557	ppbv	98
72) 4-Ethyltoluene	15.188	105	5320289	12.144	ppbv	97
73) 1,3,5-Trimethylbenzene	15.349	105	4587750	12.234	ppbv	99
74) 1,2,4-Trimethylbenzene	16.105	105	5066348	11.990	ppbv	97
75) 1,3-Dichlorobenzene	16.330	146	3202871	12.778	ppbv	100
76) 1,4-Dichlorobenzene	16.471	146	3142958	12.585	ppbv	100
77) 1,2-Dichlorobenzene	17.140	146	3101661	12.137	ppbv	99
78) Hexachloro-1,3-Butadiene	22.628	225	2444092	13.150	ppbv	99
79) Naphthalene	21.381	128	5532989	24.347	ppbv	99
80) Naphthalene,2-methyl-	24.368	142	5230189	8744.993	ppbv	100
81) 1,2,4-Trichlorobenzene	21.136	180	2891669	20.588	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072524\
 Data File : VL041406.D
 Acq On : 25 Jul 2024 19:31
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL072524

Quant Time: Jul 29 01:38:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Jul 26 07:31:14 2024
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	10.000	10.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	10.000	9.847	1.5	119	0.00
3	Chlorodifluoromethane	10.000	8.981	10.2	100	0.00
4	Chloromethane	10.000	8.963	10.4	101	0.00
5 T	Vinyl Chloride	10.000	9.264	7.4	102	0.00
6 T	Bromomethane	10.000	9.214	7.9	101	0.00
7	Chloroethane	10.000	9.496	5.0	102	0.00
8 T	Dichlorotetrafluoroethane	10.000	9.031	9.7	100	0.00
9 T	Propene	10.000	9.482	5.2	107	0.00
10 T	Heptane	10.000	9.841	1.6	101	0.00
11 T	Trichlorofluoromethane	10.000	8.800	12.0	98	0.00
12 T	1,1,2-Trichlorotrifluoroeth	10.000	8.941	10.6	99	0.00
13	Ethanol	10.000	6.269	37.3#	88	0.00
14 T	Bromoethene	10.000	9.593	4.1	100	0.00
15 T	Acetone	10.000	8.780	12.2	101	0.00
16 T	1,3-Butadiene	10.000	8.923	10.8	98	0.00
17	tert-Butyl alcohol	10.000	10.362	-3.6	103	0.00
18 T	1,1-Dichloroethene	10.000	9.602	4.0	100	0.00
19 T	Isopropyl Alcohol	10.000	8.805	12.0	91	0.00
20 T	Methylene Chloride	10.000	8.588	14.1	99	0.00
21 T	Allyl Chloride	10.000	9.744	2.6	101	0.00
22 T	trans-1,2-Dichloroethene	10.000	9.627	3.7	102	0.00
23 T	Vinyl Acetate	10.000	10.238	-2.4	101	0.00
24 T	1,1-Dichloroethane	10.000	8.974	10.3	98	0.00
25 T	Ethyl Acetate	10.000	9.581	4.2	103	0.00
26 T	Hexane	10.000	9.737	2.6	103	0.00
27 T	Carbon Disulfide	10.000	9.875	1.3	97	0.00
28 T	Methyl tert-Butyl Ether	10.000	9.572	4.3	102	0.00
29 T	Chloroform	10.000	9.179	8.2	100	0.00
30 T	Cyclohexane	10.000	10.217	-2.2	103	0.00
31 T	cis-1,2-Dichloroethene	10.000	9.817	1.8	103	0.00
32 T	1,1,1-Trichloroethane	10.000	9.280	7.2	99	0.00
33 I	1,4-Difluorobenzene	10.000	10.000	0.0	101	0.00
34 T	2-Butanone	10.000	9.886	1.1	103	0.00
35 T	Carbon Tetrachloride	10.000	9.425	5.7	96	0.00
36 T	Benzene	10.000	9.678	3.2	101	0.00
37 T	1,2-Dichloroethane	10.000	9.592	4.1	101	0.00
38 T	Trichloroethene	10.000	9.485	5.2	100	0.00
39 T	1,2-Dichloropropane	10.000	9.722	2.8	103	0.00
40 T	1,4-Dioxane	10.000	7.654	23.5	76	0.00
41 T	Tetrahydrofuran	10.000	10.640	-6.4	107	0.00
42 T	Bromodichloromethane	10.000	9.550	4.5	99	0.00
43	Methyl Methacrylate	10.000	10.898	-9.0	102	0.00
44 T	2,2,4-Trimethylpentane	10.000	9.470	5.3	100	0.00
45 T	t-1,3-Dichloropropene	10.000	12.021	-20.2	102	0.00
46 T	cis-1,3-Dichloropropene	10.000	11.015	-10.2	102	0.00
47 T	1,1,2-Trichloroethane	10.000	9.452	5.5	100	0.00
48 T	Dibromochloromethane	10.000	9.797	2.0	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072524\
 Data File : VL041406.D
 Acq On : 25 Jul 2024 19:31
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL072524

Quant Time: Jul 29 01:38:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Jul 26 07:31:14 2024
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

49 T	Bromoform	10.000	10.340	-3.4	98	0.00
50 T	4-Methyl-2-Pentanone	10.000	10.222	-2.2	102	0.00
51 T	2-Hexanone	10.000	10.994	-9.9	101	0.00
52 T	Tetrachloroethene	10.000	9.604	4.0	101	0.00
53 T	Toluene	10.000	10.183	-1.8	101	0.00
54 T	1,2-Dibromoethane	10.000	10.313	-3.1	100	0.00
55 I	Chlorobenzene-d5	10.000	10.000	0.0	104	0.00
56	1,1,1,2-Tetrachloroethane	10.000	8.940	10.6	97	0.00
57 T	Chlorobenzene	10.000	8.892	11.1	99	0.00
58 T	Ethyl Benzene	10.000	9.542	4.6	100	0.00
59 T	m/p-Xylene	20.000	18.591	7.0	99	0.00
60 T	o-Xylene	10.000	9.434	5.7	100	0.00
61 T	Styrene	10.000	10.485	-4.8	100	0.00
62	Isopropylbenzene	10.000	9.026	9.7	99	0.00
63 T	1,1,2,2-Tetrachloroethane	10.000	8.895	11.1	97	0.00
64	n-propylbenzene	10.000	9.254	7.5	98	0.00
65	tert-Butylbenzene	10.000	8.930	10.7	98	0.00
66 T	Benzyl Chloride	10.000	11.045	-10.4	104	0.00
67	sec-Butylbenzene	10.000	9.010	9.9	97	0.00
68 S	1-Bromo-4-Fluorobenzene	10.000	9.830	1.7	101	0.00
69	p-Isopropyltoluene	10.000	9.165	8.4	97	0.00
70	n-Butylbenzene	10.000	9.165	8.4	97	0.00
71	2-Chlorotoluene	10.000	9.245	7.6	99	0.00
72 T	4-Ethyltoluene	10.000	9.535	4.6	97	0.00
73 T	1,3,5-Trimethylbenzene	10.000	9.563	4.4	98	0.00
74 T	1,2,4-Trimethylbenzene	10.000	9.057	9.4	98	0.00
75 T	1,3-Dichlorobenzene	10.000	9.027	9.7	98	0.00
76 T	1,4-Dichlorobenzene	10.000	9.089	9.1	98	0.00
77 T	1,2-Dichlorobenzene	10.000	8.874	11.3	98	0.00
78 T	Hexachloro-1,3-Butadiene	10.000	8.012	19.9	99	0.00
79 T	Naphthalene	10.000	10.810	-8.1	98	0.00
80 T	Naphthalene,2-methyl-	10.000	11.790	-17.9	102	0.00
81 T	1,2,4-Trichlorobenzene	10.000	9.517	4.8	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072524\
 Data File : VL041406.D
 Acq On : 25 Jul 2024 19:31
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL072524

Quant Time: Jul 29 01:38:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Jul 26 07:31:14 2024
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	1.720	1.694	1.5	119	0.00
3	Chlorodifluoromethane	1.190	1.069	10.2	100	0.00
4	Chloromethane	0.633	0.568	10.3	101	0.00
5 T	Vinyl Chloride	0.624	0.578	7.4	102	0.00
6 T	Bromomethane	0.321	0.296	7.8	101	0.00
7	Chloroethane	0.215	0.205	4.7	102	0.00
8 T	Dichlorotetrafluoroethane	1.524	1.377	9.6	100	0.00
9 T	Propene	0.547	0.519	5.1	107	0.00
10 T	Heptane	1.229	1.209	1.6	101	0.00
11 T	Trichlorofluoromethane	1.528	1.345	12.0	98	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.100	0.984	10.5	99	0.00
13	Ethanol	0.049	0.031#	36.7#	88	0.00
14 T	Bromoethene	0.429	0.411	4.2	100	0.00
15 T	Acetone	1.136	0.998	12.1	101	0.00
16 T	1,3-Butadiene	0.607	0.542	10.7	98	0.00
17	tert-Butyl alcohol	1.036	1.074	-3.7	103	0.00
18 T	1,1-Dichloroethene	0.471	0.452	4.0	100	0.00
19 T	Isopropyl Alcohol	0.664	0.585	11.9	91	0.00
20 T	Methylene Chloride	0.424	0.364	14.2	99	0.00
21 T	Allyl Chloride	0.745	0.726	2.6	101	0.00
22 T	trans-1,2-Dichloroethene	0.471	0.453	3.8	102	0.00
23 T	Vinyl Acetate	0.868	0.888	-2.3	101	0.00
24 T	1,1-Dichloroethane	1.022	0.917	10.3	98	0.00
25 T	Ethyl Acetate	2.257	2.163	4.2	103	0.00
26 T	Hexane	0.936	0.912	2.6	103	0.00
27 T	Carbon Disulfide	1.095	1.082	1.2	97	0.00
28 T	Methyl tert-Butyl Ether	0.591	0.565	4.4	102	0.00
29 T	Chloroform	1.527	1.402	8.2	100	0.00
30 T	Cyclohexane	0.763	0.779	-2.1	103	0.00
31 T	cis-1,2-Dichloroethene	0.925	0.908	1.8	103	0.00
32 T	1,1,1-Trichloroethane	1.487	1.380	7.2	99	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	101	0.00
34 T	2-Butanone	0.484	0.478	1.2	103	0.00
35 T	Carbon Tetrachloride	0.552	0.520	5.8	96	0.00
36 T	Benzene	0.731	0.707	3.3	101	0.00
37 T	1,2-Dichloroethane	0.409	0.392	4.2	101	0.00
38 T	Trichloroethene	0.335	0.318	5.1	100	0.00
39 T	1,2-Dichloropropane	0.280	0.272	2.9	103	0.00
40 T	1,4-Dioxane	0.112	0.086	23.2	76	0.00
41 T	Tetrahydrofuran	0.277	0.294	-6.1	107	0.00
42 T	Bromodichloromethane	0.591	0.564	4.6	99	0.00
43	Methyl Methacrylate	0.262	0.286	-9.2	102	0.00
44 T	2,2,4-Trimethylpentane	1.220	1.156	5.2	100	0.00
45 T	t-1,3-Dichloropropene	0.241	0.290	-20.3	102	0.00
46 T	cis-1,3-Dichloropropene	0.348	0.384	-10.3	102	0.00
47 T	1,1,2-Trichloroethane	0.308	0.291	5.5	100	0.00
48 T	Dibromochloromethane	0.497	0.487	2.0	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072524\
 Data File : VL041406.D
 Acq On : 25 Jul 2024 19:31
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL072524

Quant Time: Jul 29 01:38:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Jul 26 07:31:14 2024
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.420	0.434	-3.3	98	0.00
50 T	4-Methyl-2-Pentanone	0.763	0.780	-2.2	102	0.00
51 T	2-Hexanone	0.577	0.634	-9.9	101	0.00
52 T	Tetrachloroethene	0.282	0.271	3.9	101	0.00
53 T	Toluene	0.814	0.829	-1.8	101	0.00
54 T	1,2-Dibromoethane	0.410	0.423	-3.2	100	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	104	0.00
56	1,1,1,2-Tetrachloroethane	0.388	0.347	10.6	97	0.00
57 T	Chlorobenzene	0.712	0.633	11.1	99	0.00
58 T	Ethyl Benzene	1.153	1.100	4.6	100	0.00
59 T	m/p-Xylene	0.995	0.925	7.0	99	0.00
60 T	o-Xylene	0.941	0.888	5.6	100	0.00
61 T	Styrene	0.425	0.445	-4.7	100	0.00
62	Isopropylbenzene	1.431	1.292	9.7	99	0.00
63 T	1,1,2,2-Tetrachloroethane	0.594	0.529	10.9	97	0.00
64	n-propylbenzene	0.374	0.346	7.5	98	0.00
65	tert-Butylbenzene	1.292	1.154	10.7	98	0.00
66 T	Benzyl Chloride	0.083	0.092	-10.8	104	0.00
67	sec-Butylbenzene	1.781	1.604	9.9	97	0.00
68 S	1-Bromo-4-Fluorobenzene	0.770	0.756	1.8	101	0.00
69	p-Isopropyltoluene	1.478	1.354	8.4	97	0.00
70	n-Butylbenzene	1.515	1.388	8.4	97	0.00
71	2-Chlorotoluene	1.055	0.976	7.5	99	0.00
72 T	4-Ethyltoluene	1.108	1.057	4.6	97	0.00
73 T	1,3,5-Trimethylbenzene	0.955	0.913	4.4	98	0.00
74 T	1,2,4-Trimethylbenzene	1.110	1.005	9.5	98	0.00
75 T	1,3-Dichlorobenzene	0.698	0.630	9.7	98	0.00
76 T	1,4-Dichlorobenzene	0.676	0.614	9.2	98	0.00
77 T	1,2-Dichlorobenzene	0.692	0.614	11.3	98	0.00
78 T	Hexachloro-1,3-Butadiene	0.590	0.473	19.8	99	0.00
79 T	Naphthalene	0.980	1.059	-8.1	98	0.00
80 T	Naphthalene,2-methyl-	0.821	0.968	-17.9	102	0.00
81 T	1,2,4-Trichlorobenzene	0.583	0.555	4.8	99	0.00

(#) = Out of Range



SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081924\
Data File : VL041539.D
Acq On : 19 Aug 2024 8:10
Operator : SY/MD
Sample : BFB
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
Last Update : Tue Aug 20 02:10:04 2024



AutoFind: Scans 3478, 3479, 3480; Background Corrected with Scan 3456

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	24.9	188907	PASS
75	95	30	66	56.1	425728	PASS
95	95	100	100	100.0	758997	PASS
96	95	5	9	6.9	52083	PASS
173	174	0.00	2	1.0	5439	PASS
174	95	50	120	68.5	519701	PASS
175	174	4	9	7.2	37573	PASS
176	174	93	101	96.6	502080	PASS
177	176	5	9	6.6	33307	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081924\
 Data File : VL041540.D
 Acq On : 19 Aug 2024 8:59
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 08/20/2024
 Supervised By : Mahesh Dadoda 08/20/2024

Quant Time: Aug 20 02:13:28 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 20 02:13:28 2024

QLast Update : Tue Aug 20 02:10:04 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.211	49	949994	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.671	114	2559374	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.491	117	2515938m	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 13.799 95 1947516 10.059 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 100.600%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.748	85	884681	5.413	ppbv	99
3) Chlorodifluoromethane	2.693	51	1159695	10.254	ppbv	100
4) Chloromethane	2.832	50	629369	10.463	ppbv	99
5) Vinyl Chloride	2.938	62	619805	10.454	ppbv	99
6) Bromomethane	3.137	94	327691	10.754	ppbv	92
7) Chloroethane	3.214	64	229324	11.207	ppbv	97
8) Dichlorotetrafluoroethane	2.874	85	1383474	9.554	ppbv	100
9) Propene	2.713	41	521016	10.029	ppbv	98
10) Heptane	7.597	43	1321671	11.323	ppbv	99
11) Trichlorofluoromethane	3.581	101	1519219	10.467	ppbv	98
12) 1,1,2-Trichlorotrifluo...	4.082	101	1094070	10.468	ppbv	98
13) Ethanol	3.256	45	36044m	7.762	ppbv	
14) Bromoethene	3.385	108	440924	10.821	ppbv	99
15) Acetone	3.484	43	1130098	10.470	ppbv	100
16) 1,3-Butadiene	3.002	39	624120	10.822	ppbv	97
17) tert-Butyl alcohol	3.893	59	1096646	11.140	ppbv	100
18) 1,1-Dichloroethene	3.899	96	487068	10.895	ppbv	92
19) Isopropyl Alcohol	3.591	45	651343	10.322	ppbv #	96
20) Methylene Chloride	3.951	84	412911	10.258	ppbv	98
21) Allyl Chloride	4.012	41	770702	10.888	ppbv	100
22) trans-1,2-Dichloroethene	4.462	96	491446	10.994	ppbv	98
23) Vinyl Acetate	4.642	43	966046	11.719	ppbv #	98
24) 1,1-Dichloroethane	4.581	63	1045215	10.768	ppbv	99
25) Ethyl Acetate	5.221	43	2381732	11.107	ppbv	99
26) Hexane	5.230	57	979260	11.009	ppbv	99
27) Carbon Disulfide	4.144	76	1163109	11.177	ppbv	99
28) Methyl tert-Butyl Ether	4.603	73	587594	10.474	ppbv	100
29) Chloroform	5.295	83	1547234	10.664	ppbv	99
30) Cyclohexane	6.623	84	817710	11.288	ppbv	99
31) cis-1,2-Dichloroethene	5.095	61	948884	10.798	ppbv	98
32) 1,1,1-Trichloroethane	6.021	97	1527033	10.812	ppbv	99
34) 2-Butanone	4.803	43	1405398	11.352	ppbv	99
35) Carbon Tetrachloride	6.513	117	1609727	11.403	ppbv	99
36) Benzene	6.391	78	2059410	11.012	ppbv	99
37) 1,2-Dichloroethane	5.825	62	1153588	11.021	ppbv	99
38) Trichloroethene	7.311	130	978782	11.424	ppbv	97
39) 1,2-Dichloropropane	7.095	63	811438	11.333	ppbv	97
40) 1,4-Dioxane	7.307	88	311632	10.835	ppbv	78
41) Tetrahydrofuran	5.574	42	831216	11.739	ppbv	98
42) Bromodichloromethane	7.269	83	1693152	11.203	ppbv	99
43) Methyl Methacrylate	7.494	69	821493	12.250	ppbv	97
44) 2,2,4-Trimethylpentane	7.343	57	3513799	11.249	ppbv	99
45) t-1,3-Dichloropropene	8.729	75	843236	13.661	ppbv	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081924\
 Data File : VL041540.D
 Acq On : 19 Aug 2024 8:59
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 08/20/2024
 Supervised By : Mahesh Dadoda 08/20/2024

Quant Time: Aug 20 02:13:28 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 20 02:13:28 2024

QLast Update : Tue Aug 20 02:10:04 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	8.163	75	1113175	12.489	ppbv	97
47) 1,1,2-Trichloroethane	8.918	97	874878	11.085	ppbv	96
48) Dibromochloromethane	9.732	129	1471389	11.573	ppbv	100
49) Bromoform	12.429	173	1290363	12.015	ppbv	99
50) 4-Methyl-2-Pentanone	8.198	43	2348083	12.021	ppbv	100
51) 2-Hexanone	9.571	43	1889904	12.796	ppbv #	98
52) Tetrachloroethene	10.635	164	785134	10.865	ppbv	98
53) Toluene	9.240	91	2434932	11.689	ppbv	100
54) 1,2-Dibromoethane	10.031	107	1279830	12.183	ppbv	99
56) 1,1,1,2-Tetrachloroethane	11.529	131	1082376	11.093	ppbv #	57
57) Chlorobenzene	11.545	112	1917343	10.704	ppbv	99
58) Ethyl Benzene	12.111	91	3295215	11.363	ppbv	100
59) m/p-Xylene	12.394	91	5677231m	22.679	ppbv	
60) o-Xylene	13.079	91	2717555	11.476	ppbv	100
61) Styrene	12.918	104	1314555	12.305	ppbv	99
62) Isopropylbenzene	14.050	105	3944805	10.958	ppbv	99
63) 1,1,2,2-Tetrachloroethane	13.069	83	1675653	11.206	ppbv	99
64) n-propylbenzene	14.937	120	1056999	11.231	ppbv	95
65) tert-Butylbenzene	16.111	119	3568987	10.976	ppbv	100
66) Benzyl Chloride	16.346	91	294537	14.122	ppbv	98
67) sec-Butylbenzene	16.658	105	5012445	11.189	ppbv	99
69) p-Isopropyltoluene	17.018	119	4225410	11.366	ppbv	100
70) n-Butylbenzene	17.908	91	4335585	11.376	ppbv	100
71) 2-Chlorotoluene	14.822	91	2983953	11.240	ppbv	99
72) 4-Ethyltoluene	15.214	105	3230798	11.588	ppbv	98
73) 1,3,5-Trimethylbenzene	15.371	105	2794784	11.636	ppbv	98
74) 1,2,4-Trimethylbenzene	16.127	105	3136247	11.232	ppbv	97
75) 1,3-Dichlorobenzene	16.355	146	1953807	11.131	ppbv	100
76) 1,4-Dichlorobenzene	16.494	146	1933808	11.378	ppbv	98
77) 1,2-Dichlorobenzene	17.162	146	1900542	10.922	ppbv	100
78) Hexachloro-1,3-Butadiene	22.657	225	1416463	9.546	ppbv	99
79) Naphthalene	21.410	128	3145958	12.765	ppbv	99
80) Naphthalene,2-methyl-	24.390	142	2696816	13.054	ppbv	100
81) 1,2,4-Trichlorobenzene	21.162	180	1663658	11.343	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081924\
 Data File : VL041540.D
 Acq On : 19 Aug 2024 8:59
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Aug 20 02:13:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Aug 20 02:10:04 2024
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	77	0.00
2 T	Dichlorodifluoromethane	1.720	0.931	45.9#	51	0.00
3	Chlorodifluoromethane	1.190	1.221	-2.6	88	0.00
4	Chloromethane	0.633	0.662	-4.6	91	0.00
5 T	Vinyl Chloride	0.624	0.652	-4.5	89	0.00
6 T	Bromomethane	0.321	0.345	-7.5	91	0.00
7	Chloroethane	0.215	0.241	-12.1	92	0.00
8 T	Dichlorotetrafluoroethane	1.524	1.456	4.5	82	0.00
9 T	Propene	0.547	0.548	-0.2	87	0.00
10 T	Heptane	1.229	1.391	-13.2	90	0.00
11 T	Trichlorofluoromethane	1.528	1.599	-4.6	90	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.100	1.152	-4.7	89	0.00
13	Ethanol	0.049	0.038#	22.4	84	0.00
14 T	Bromoethene	0.429	0.464	-8.2	87	0.00
15 T	Acetone	1.136	1.190	-4.8	92	0.00
16 T	1,3-Butadiene	0.607	0.657	-8.2	92	0.00
17	tert-Butyl alcohol	1.036	1.154	-11.4	85	0.00
18 T	1,1-Dichloroethene	0.471	0.513	-8.9	87	0.00
19 T	Isopropyl Alcohol	0.664	0.686	-3.3	82	0.00
20 T	Methylene Chloride	0.424	0.435	-2.6	91	0.00
21 T	Allyl Chloride	0.745	0.811	-8.9	87	0.00
22 T	trans-1,2-Dichloroethene	0.471	0.517	-9.8	89	0.00
23 T	Vinyl Acetate	0.868	1.017	-17.2	89	0.00
24 T	1,1-Dichloroethane	1.022	1.100	-7.6	91	0.00
25 T	Ethyl Acetate	2.257	2.507	-11.1	92	0.00
26 T	Hexane	0.936	1.031	-10.1	90	0.00
27 T	Carbon Disulfide	1.095	1.224	-11.8	85	0.00
28 T	Methyl tert-Butyl Ether	0.591	0.619	-4.7	86	0.00
29 T	Chloroform	1.527	1.629	-6.7	89	0.00
30 T	Cyclohexane	0.763	0.861	-12.8	87	0.00
31 T	cis-1,2-Dichloroethene	0.925	0.999	-8.0	87	0.00
32 T	1,1,1-Trichloroethane	1.487	1.607	-8.1	89	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	76	0.00
34 T	2-Butanone	0.484	0.549	-13.4	90	0.00
35 T	Carbon Tetrachloride	0.552	0.629	-13.9	88	0.00
36 T	Benzene	0.731	0.805	-10.1	87	0.00
37 T	1,2-Dichloroethane	0.409	0.451	-10.3	88	0.00
38 T	Trichloroethene	0.335	0.382	-14.0	91	0.00
39 T	1,2-Dichloropropane	0.280	0.317	-13.2	91	0.00
40 T	1,4-Dioxane	0.112	0.122	-8.9	81	0.00
41 T	Tetrahydrofuran	0.277	0.325	-17.3	89	0.00
42 T	Bromodichloromethane	0.591	0.662	-12.0	88	0.00
43	Methyl Methacrylate	0.262	0.321	-22.5	87	0.00
44 T	2,2,4-Trimethylpentane	1.220	1.373	-12.5	90	0.00
45 T	t-1,3-Dichloropropene	0.241	0.329	-36.5#	88	0.00
46 T	cis-1,3-Dichloropropene	0.348	0.435	-25.0	87	0.00
47 T	1,1,2-Trichloroethane	0.308	0.342	-11.0	89	0.00
48 T	Dibromochloromethane	0.497	0.575	-15.7	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081924\
 Data File : VL041540.D
 Acq On : 19 Aug 2024 8:59
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Aug 20 02:13:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Aug 20 02:10:04 2024
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.420	0.504	-20.0	86	0.00
50 T	4-Methyl-2-Pentanone	0.763	0.917	-20.2	90	0.00
51 T	2-Hexanone	0.577	0.738	-27.9	88	0.00
52 T	Tetrachloroethene	0.282	0.307	-8.9	86	0.00
53 T	Toluene	0.814	0.951	-16.8	88	0.00
54 T	1,2-Dibromoethane	0.410	0.500	-22.0	89	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	77	0.00
56	1,1,1,2-Tetrachloroethane	0.388	0.430	-10.8	89	0.00
57 T	Chlorobenzene	0.712	0.762	-7.0	88	0.00
58 T	Ethyl Benzene	1.153	1.310	-13.6	88	0.00
59 T	m/p-Xylene	0.995	1.128	-13.4	89	0.00
60 T	o-Xylene	0.941	1.080	-14.8	89	0.00
61 T	Styrene	0.425	0.522	-22.8	87	0.00
62	Isopropylbenzene	1.431	1.568	-9.6	88	0.00
63 T	1,1,2,2-Tetrachloroethane	0.594	0.666	-12.1	90	0.00
64	n-propylbenzene	0.374	0.420	-12.3	88	0.00
65	tert-Butylbenzene	1.292	1.419	-9.8	89	0.00
66 T	Benzyl Chloride	0.083	0.117	-41.0#	98	0.00
67	sec-Butylbenzene	1.781	1.992	-11.8	89	0.00
68 S	1-Bromo-4-Fluorobenzene	0.770	0.774	-0.5	76	0.00
69	p-Isopropyltoluene	1.478	1.679	-13.6	89	0.00
70	n-Butylbenzene	1.515	1.723	-13.7	89	0.00
71	2-Chlorotoluene	1.055	1.186	-12.4	89	0.00
72 T	4-Ethyltoluene	1.108	1.284	-15.9	87	0.00
73 T	1,3,5-Trimethylbenzene	0.955	1.111	-16.3	88	0.00
74 T	1,2,4-Trimethylbenzene	1.110	1.247	-12.3	90	0.00
75 T	1,3-Dichlorobenzene	0.698	0.777	-11.3	89	0.00
76 T	1,4-Dichlorobenzene	0.676	0.769	-13.8	90	0.00
77 T	1,2-Dichlorobenzene	0.692	0.755	-9.1	89	0.00
78 T	Hexachloro-1,3-Butadiene	0.590	0.563	4.6	87	0.00
79 T	Naphthalene	0.980	1.250	-27.6	86	0.00
80 T	Naphthalene,2-methyl-	0.821	1.072	-30.6#	83	0.00
81 T	1,2,4-Trichlorobenzene	0.583	0.661	-13.4	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081924\
 Data File : VL041541.D
 Acq On : 19 Aug 2024 10:07
 Operator : SY/MD
 Sample : VL0819ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0819ABL01

Quant Time: Aug 20 02:14:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Aug 20 02:10:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.211	49	895764	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.671	114	2374230	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.484	117	2131055	10.000	ppbv	# 0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.799	95	1588219	9.684	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.800%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081924\
Data File : VL041541.D
Acq On : 19 Aug 2024 10:07
Operator : SY/MD
Sample : VL0819ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0819ABL01

Quant Time: Aug 20 02:14:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Tue Aug 20 02:10:04 2024
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081924\
 Data File : VL041542.D
 Acq On : 19 Aug 2024 10:48
 Operator : SY/MD
 Sample : VL0819ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0819ABS01

Quant Time: Aug 20 02:14:35 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 20 2024

QLast Update : Tue Aug 20 02:10:04 2024

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 08/20/2024
 Supervised By : Mahesh Dadoda 08/20/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.211	49	953834	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.674	114	2568604	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.487	117	2570572	10.000	ppbv	# 0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 13.796 95 1934486 9.779 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 97.800%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.748	85	866011	5.278	ppbv	100
3) Chlorodifluoromethane	2.694	51	1148497	10.115	ppbv	100
4) Chloromethane	2.832	50	626957	10.381	ppbv	99
5) Vinyl Chloride	2.941	62	632244	10.621	ppbv	96
6) Bromomethane	3.141	94	337860	11.043	ppbv	98
7) Chloroethane	3.214	64	227400	11.069	ppbv	96
8) Dichlorotetrafluoroethane	2.877	85	1381485	9.502	ppbv	99
9) Propene	2.713	41	529749	10.156	ppbv	98
10) Heptane	7.600	43	1331349	11.360	ppbv	100
11) Trichlorofluoromethane	3.581	101	1495695	10.264	ppbv	97
12) 1,1,2-Trichlorotrifluo...	4.086	101	1072747	10.223	ppbv	100
13) Ethanol	3.259	45	36448m	7.818	ppbv	
14) Bromoethene	3.385	108	429951	10.510	ppbv	99
15) Acetone	3.488	43	1136536	10.487	ppbv	99
16) 1,3-Butadiene	3.005	39	616058	10.640	ppbv	99
17) tert-Butyl alcohol	3.896	59	1138725	11.521	ppbv	100
18) 1,1-Dichloroethene	3.899	96	490575	10.929	ppbv	96
19) Isopropyl Alcohol	3.594	45	662593	10.458	ppbv #	98
20) Methylene Chloride	3.951	84	468002	11.580	ppbv	95
21) Allyl Chloride	4.015	41	776304	10.923	ppbv	100
22) trans-1,2-Dichloroethene	4.465	96	469979	10.471	ppbv	97
23) Vinyl Acetate	4.645	43	984169	11.890	ppbv #	98
24) 1,1-Dichloroethane	4.581	63	1033773	10.608	ppbv	99
25) Ethyl Acetate	5.224	43	2417577	11.229	ppbv	99
26) Hexane	5.234	57	1027698	11.507	ppbv	97
27) Carbon Disulfide	4.144	76	1155043	11.055	ppbv	99
28) Methyl tert-Butyl Ether	4.603	73	589223	10.461	ppbv	100
29) Chloroform	5.295	83	1534946	10.536	ppbv	99
30) Cyclohexane	6.623	84	826550	11.364	ppbv	99
31) cis-1,2-Dichloroethene	5.099	61	959317	10.873	ppbv	99
32) 1,1,1-Trichloroethane	6.025	97	1517453	10.701	ppbv	100
34) 2-Butanone	4.806	43	1409794	11.346	ppbv	100
35) Carbon Tetrachloride	6.513	117	1592749	11.242	ppbv	98
36) Benzene	6.391	78	2079587	11.080	ppbv	98
37) 1,2-Dichloroethane	5.829	62	1168791	11.126	ppbv	100
38) Trichloroethene	7.311	130	970933	11.292	ppbv	97
39) 1,2-Dichloropropane	7.095	63	802701	11.171	ppbv	96
40) 1,4-Dioxane	7.308	88	276693	9.586	ppbv #	52
41) Tetrahydrofuran	5.574	42	830620	11.688	ppbv	98
42) Bromodichloromethane	7.269	83	1674518	11.039	ppbv	98
43) Methyl Methacrylate	7.494	69	830856	12.345	ppbv	99
44) 2,2,4-Trimethylpentane	7.346	57	3498360	11.160	ppbv	99
45) t-1,3-Dichloropropene	8.732	75	832465	13.438	ppbv	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081924\
 Data File : VL041542.D
 Acq On : 19 Aug 2024 10:48
 Operator : SY/MD
 Sample : VL0819ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0819ABS01

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024
 Supervised By :Mahesh Dadoda 08/20/2024

Quant Time: Aug 20 02:14:35 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Tue Aug 20 02:10:04 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	8.166	75	1134089	12.678	ppbv	98
47) 1,1,2-Trichloroethane	8.922	97	871843	11.007	ppbv	98
48) Dibromochloromethane	9.732	129	1455263	11.405	ppbv	100
49) Bromoform	12.430	173	1269504	11.778	ppbv	99
50) 4-Methyl-2-Pentanone	8.201	43	2350707	11.991	ppbv	99
51) 2-Hexanone	9.571	43	1863582	12.572	ppbv #	99
52) Tetrachloroethene	10.639	164	782284	10.787	ppbv	98
53) Toluene	9.243	91	2430940	11.628	ppbv	100
54) 1,2-Dibromoethane	10.031	107	1250868	11.865	ppbv	97
56) 1,1,1,2-Tetrachloroethane	11.529	131	1071372	10.747	ppbv #	59
57) Chlorobenzene	11.549	112	1920481	10.493	ppbv	98
58) Ethyl Benzene	12.111	91	3271939	11.043	ppbv	98
59) m/p-Xylene	12.391	91	5625619m	21.995	ppbv	
60) o-Xylene	13.079	91	2684986	11.097	ppbv	100
61) Styrene	12.918	104	1314788	12.045	ppbv	98
62) Isopropylbenzene	14.050	105	3916892	10.649	ppbv	100
63) 1,1,2,2-Tetrachloroethane	13.069	83	1660380	10.867	ppbv	100
64) n-propylbenzene	14.937	120	1051699	10.937	ppbv	95
65) tert-Butylbenzene	16.111	119	3561329	10.720	ppbv	100
66) Benzyl Chloride	16.349	91	293870	13.791	ppbv	99
67) sec-Butylbenzene	16.658	105	4963512	10.844	ppbv	100
69) p-Isopropyltoluene	17.015	119	4185466	11.019	ppbv	99
70) n-Butylbenzene	17.908	91	4322218	11.100	ppbv	100
71) 2-Chlorotoluene	14.822	91	2956439	10.899	ppbv	99
72) 4-Ethyltoluene	15.214	105	3303117	11.596	ppbv	99
73) 1,3,5-Trimethylbenzene	15.372	105	2807444	11.440	ppbv	98
74) 1,2,4-Trimethylbenzene	16.130	105	3111891	10.908	ppbv	99
75) 1,3-Dichlorobenzene	16.352	146	1936719	10.799	ppbv	100
76) 1,4-Dichlorobenzene	16.494	146	1874477	10.794	ppbv	99
77) 1,2-Dichlorobenzene	17.166	146	1866349	10.498	ppbv	100
78) Hexachloro-1,3-Butadiene	22.654	225	1419954	9.366	ppbv	100
79) Naphthalene	21.407	128	3143301	12.483	ppbv	99
80) Naphthalene,2-methyl-	24.394	142	2571622	12.183	ppbv	99
81) 1,2,4-Trichlorobenzene	21.166	180	1655281	11.046	ppbv	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081924\
 Data File : VL041543.D
 Acq On : 19 Aug 2024 11:29
 Operator : SY/MD
 Sample : QC081324 CAN 10604
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC081324 CAN 10604

Quant Time: Aug 20 02:15:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Aug 20 02:10:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.211	49	990128	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.671	114	2684055	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.487	117	2394412	10.000	ppbv	# 0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.799	95	1834029	9.953	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.500%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081924\
Data File : VL041543.D
Acq On : 19 Aug 2024 11:29
Operator : SY/MD
Sample : QC081324 CAN 10604
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC081324 CAN 10604

Quant Time: Aug 20 02:15:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Tue Aug 20 02:10:04 2024
Response via : Initial Calibration




Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091624\
Data File : VL041590.D
Acq On : 16 Sep 2024 11:50
Operator : SY/MD
Sample : BFB
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
Last Update : Tue Sep 17 09:12:21 2024



AutoFind: Scans 3466, 3467, 3468; Background Corrected with Scan 3441

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	8	40	24.8	206293	PASS
75	95	30	66	56.9	472747	PASS
95	95	100	100	100.0	831488	PASS
96	95	5	9	6.6	54749	PASS
173	174	0.00	2	0.7	4144	PASS
174	95	50	120	69.3	576085	PASS
175	174	4	9	7.6	43507	PASS
176	174	93	101	96.6	556480	PASS
177	176	5	9	6.7	37259	PASS

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\

Method File : VL091624AIR.M

Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022

Last Update : Tue Sep 17 09:12:21 2024

Response Via : Initial Calibration

Calibration Files

0.03=VL041597.D 0.1 =VL041596.D 0.5 =VL041591.D 1 =VL041594.D 2 =VL041593.D 10 =VL041592.D 15 =VL041598.D

Compound	0.03	0.1	0.5	1	2	10	15	Avg	%RSD

1) I Bromochloromethane	-----ISTD-----								
2) T Dichlorodifluo...			2.008	1.517	1.307	1.311	1.078	1.444	24.33
3) Chlorodifluoro...			1.427	1.234	1.245	1.104	1.055	1.213	11.94
4) Chloromethane			0.775	0.677	0.681	0.611	0.592	0.667	10.74
5) T Vinyl Chloride	0.459	0.525	0.634	0.633	0.625	0.587	0.580	0.578	11.28
6) T Bromomethane			0.325	0.277	0.298	0.274	0.262	0.287	8.62
7) Chloroethane			0.264	0.222	0.219	0.213	0.205	0.225	10.30
8) T Dichlorotetra...			1.783	1.546	1.526	1.422	1.298	1.515	11.84
9) T Propene			0.501	0.517	0.538	0.513	0.489	0.512	3.61
10) T Heptane			1.051	1.173	1.259	1.270	1.237	1.198	7.55
11) T Trichlorofluor...			1.732	1.636	1.535	1.429	1.353	1.537	9.95
12) T 1,1,2-Trichlor...			1.226	1.162	1.152	1.040	0.986	1.113	8.77
13) Ethanol			0.047	0.057	0.046	0.024	0.018	0.038#	42.94
14) T Bromoethene			0.429	0.421	0.427	0.423	0.410	0.422	1.76
15) T Acetone			1.775	1.244	1.240	1.027	0.990	1.255	24.97
16) T 1,3-Butadiene			0.674	0.612	0.622	0.579	0.566	0.611	6.93
17) tert-Butyl alc...			1.085	1.138	1.172	1.090	1.036	1.104	4.75
18) T 1,1-Dichloroet...			0.463	0.490	0.482	0.476	0.445	0.471	3.79
19) T Isopropyl Alcohol			0.690	0.659	0.674	0.625	0.566	0.643	7.61
20) T Methylene Chlo...			0.675	0.478	0.453	0.388	0.365	0.472	26.01
21) T Allyl Chloride			0.729	0.717	0.763	0.734	0.695	0.728	3.40
22) T trans-1,2-Dich...			0.497	0.462	0.488	0.460	0.443	0.470	4.65
23) T Vinyl Acetate			0.710	0.652	0.713	0.690	0.688	0.691	3.54
24) T 1,1-Dichloroet...			1.141	1.056	1.053	0.992	0.939	1.036	7.31
25) T Ethyl Acetate			2.106	2.286	2.408	2.286	2.184	2.254	5.08
26) T Hexane			0.979	0.924	0.974	0.927	0.895	0.940	3.81
27) T Carbon Disulfide			0.865	0.940	1.005	1.108	1.077	0.999	9.93
28) T Methyl tert-Bu...			0.628	0.578	0.604	0.558	0.530	0.579	6.60
29) T Chloroform			1.715	1.573	1.543	1.457	1.402	1.538	7.81
30) T Cyclohexane			0.603	0.670	0.780	0.785	0.785	0.725	11.57
31) T cis-1,2-Dichlo...			0.837	0.885	0.948	0.900	0.892	0.892	4.44
32) T 1,1,1-Trichlor...	1.165	1.138	1.529	1.452	1.479	1.438	1.389	1.370	11.32

33) I 1,4-Difluorobenzene	-----ISTD-----								
34) T 2-Butanone			0.521	0.437	0.500	0.470	0.456	0.477	7.07
35) T Carbon Tetrach...	0.363	0.395	0.536	0.518	0.555	0.543	0.525	0.491	15.89
36) T Benzene			0.640	0.677	0.746	0.714	0.696	0.694	5.71
37) T 1,2-Dichloroet...			0.409	0.406	0.413	0.399	0.388	0.403	2.48
38) T Trichloroethene	0.347	0.305	0.358	0.357	0.394	0.378	0.364	0.358	7.80
39) T 1,2-Dichloropr...			0.293	0.278	0.297	0.275	0.270	0.283	4.16

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\ Method File : VL091624AIR.M										
40) T	1,4-Dioxane			0.111	0.100	0.115	0.114	0.093	0.107	9.00
41) T	Tetrahydrofuran			0.209	0.257	0.272	0.289	0.285	0.262	12.31
42) T	Bromodichlorom...			0.546	0.557	0.572	0.573	0.554	0.561	2.07
43) T	Methyl Methacr...			0.195	0.225	0.261	0.281	0.280	0.249	15.02
44) T	2,2,4-Trimethy...			1.133	1.210	1.304	1.210	1.150	1.202	5.59
45) T	t-1,3-Dichloro...			0.146	0.157	0.194	0.247	0.259	0.200	25.54
46) T	cis-1,3-Dichlo...			0.197	0.256	0.306	0.353	0.355	0.294	22.93
47) T	1,1,2-Trichlor...			0.297	0.303	0.312	0.299	0.290	0.300	2.69
48) T	Dibromochlorom...			0.402	0.428	0.461	0.493	0.487	0.455	8.52
49) T	Bromoform			0.324	0.363	0.401	0.432	0.431	0.390	11.89
50) T	4-Methyl-2-Pen...			0.616	0.697	0.804	0.793	0.771	0.736	10.74
51) T	2-Hexanone			0.419	0.510	0.603	0.631	0.622	0.557	16.33
52) T	Tetrachloroethene	0.218	0.216	0.248	0.268	0.277	0.267	0.266	0.251	9.96
53) T	Toluene			0.526	0.703	0.806	0.834	0.816	0.737	17.42
54) T	1,2-Dibromoethane	0.259		0.370	0.379	0.402	0.407	0.399	0.369	15.11
-----ISTD-----										
55) I	Chlorobenzene-d5									
56) T	1,1,1,2-Tetrac...			0.421	0.401	0.406	0.378	0.345	0.390	7.55
57) T	Chlorobenzene			0.793	0.767	0.739	0.685	0.621	0.721	9.51
58) T	Ethyl Benzene			0.838	1.096	1.179	1.173	1.067	1.070	12.97
59) T	m/p-Xylene			0.830	0.974	1.061	0.996	0.904	0.953	9.31
60) T	o-Xylene			0.782	0.931	1.016	0.960	0.878	0.913	9.75
61) T	Styrene			0.247	0.338	0.413	0.468	0.442	0.382	23.54
62) T	Isopropylbenzene			1.282	1.461	1.522	1.394	1.260	1.384	8.14
63) T	1,1,2,2-Tetrac...	0.418	0.394	0.547	0.532	0.510	0.456	0.413	0.467	13.31
64) T	n-propylbenzene			0.308	0.356	0.401	0.375	0.347	0.357	9.56
65) T	tert-Butylbenzene			1.181	1.338	1.433	1.256	1.134	1.269	9.47
66) T	Benzyl Chloride			0.064	0.064	0.065	0.067	0.069	0.066	3.08
67) T	sec-Butylbenzene			1.575	1.882	2.001	1.770	1.572	1.760	10.74
68) S	1-Bromo-4-Fluo...	0.737	0.751	0.749	0.749	0.723	0.744	0.701	0.736	2.47
69) T	p-Isopropyltol...			1.164	1.476	1.677	1.481	1.325	1.425	13.46
70) T	n-Butylbenzene			1.271	1.572	1.732	1.520	1.368	1.492	12.04
71) T	2-Chlorotoluene			0.917	1.047	1.121	1.050	0.957	1.018	8.00
72) T	4-Ethyltoluene			0.850	1.086	1.197	1.154	1.052	1.068	12.57
73) T	1,3,5-Trimethy...			0.790	0.975	1.054	0.988	0.907	0.943	10.63
74) T	1,2,4-Trimethy...			0.986	1.181	1.267	1.095	0.989	1.104	11.09
75) T	1,3-Dichlorobe...			0.676	0.712	0.766	0.680	0.624	0.692	7.56
76) T	1,4-Dichlorobe...			0.598	0.708	0.733	0.661	0.614	0.663	8.77
77) T	1,2-Dichlorobe...			0.687	0.711	0.754	0.658	0.604	0.683	8.25
78) T	Hexachloro-1,3...			0.607	0.621	0.635	0.491	0.461	0.563	14.33
79) T	Naphthalene	0.311		0.570	0.869	1.107	1.031	0.985	0.812	38.05
80) T	Naphthalene,2-...			0.267	0.435	0.613	0.749	0.781	0.569	38.19
81) T	1,2,4-Trichlor...			0.405	0.503	0.607	0.543	0.524	0.516	14.24

(#) = Out of Range

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091624\
 Data File : VL041591.D
 Acq On : 16 Sep 2024 12:36
 Operator : SY/MD
 Sample : VSTDICC0.5
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.5

Quant Time: Sep 17 07:23:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Tue Sep 17 05:12:12 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 09/17/2024
 Supervised By : Mahesh Dadoda 09/17/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.166	49	1007876	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.626	114	2828824	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.439	117	2557349	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	13.751	95	1914464	10.111	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.100%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.710	85	101189	0.723	ppbv	96
3) Chlorodifluoromethane	2.655	51	71898	0.601	ppbv	99
4) Chloromethane	2.793	50	39035	0.689	ppbv	90
5) Vinyl Chloride	2.896	62	31972	0.554	ppbv	91
6) Bromomethane	3.099	94	16392	0.584	ppbv	99
7) Chloroethane	3.179	64	13318	0.613	ppbv	98
8) Dichlorotetrafluoroethane	2.835	85	89844	0.600	ppbv	95
9) Propene	2.677	41	25238	0.485	ppbv	88
10) Heptane	7.549	43	52953m	0.447	ppbv	
11) Trichlorofluoromethane	3.539	101	87284	0.576	ppbv	98
12) 1,1,2-Trichlorotrifluo...	4.041	101	61766	0.562	ppbv	93
13) Ethanol	3.227	45	2374m	0.709	ppbv	
14) Bromoethene	3.343	108	21604	0.511	ppbv #	91
15) Acetone	3.455	43	89461	0.777	ppbv	92
16) 1,3-Butadiene	2.967	39	33974	0.578	ppbv	93
17) tert-Butyl alcohol	3.870	59	54674	0.500	ppbv	95
18) 1,1-Dichloroethene	3.861	96	23326	0.483	ppbv	96
19) Isopropyl Alcohol	3.565	45	34766m	0.560	ppbv	
20) Methylene Chloride	3.909	84	34011	0.768	ppbv	93
21) Allyl Chloride	3.976	41	36721	0.501	ppbv	88
22) trans-1,2-Dichloroethene	4.420	96	25050	0.536	ppbv #	81
23) Vinyl Acetate	4.610	43	35784	0.523	ppbv #	87
24) 1,1-Dichloroethane	4.539	63	57495	0.562	ppbv #	95
25) Ethyl Acetate	5.185	43	106139	0.465	ppbv #	95
26) Hexane	5.189	57	49338	0.535	ppbv	90
27) Carbon Disulfide	4.105	76	43606	0.435	ppbv #	87
28) Methyl tert-Butyl Ether	4.571	73	31629m	0.552	ppbv	
29) Chloroform	5.246	83	86434	0.566	ppbv	97
30) Cyclohexane	6.578	84	30381m	0.419	ppbv	
31) cis-1,2-Dichloroethene	5.053	61	42156	0.473	ppbv	93
32) 1,1,1-Trichloroethane	5.979	97	77056	0.568	ppbv	96
34) 2-Butanone	4.777	43	73717m	0.569	ppbv	
35) Carbon Tetrachloride	6.471	117	75844	0.554	ppbv #	94
36) Benzene	6.346	78	90490	0.465	ppbv #	95
37) 1,2-Dichloroethane	5.783	62	57866	0.511	ppbv #	96
38) Trichloroethene	7.269	130	50656m	0.502	ppbv	
39) 1,2-Dichloropropane	7.044	63	41453m	0.523	ppbv	
40) 1,4-Dioxane	7.304	88	15716m	0.554	ppbv	
41) Tetrahydrofuran	5.558	42	29577	0.397	ppbv	96
42) Bromodichloromethane	7.224	83	77287	0.499	ppbv #	96
43) Methyl Methacrylate	7.452	69	27616	0.393	ppbv	95
44) 2,2,4-Trimethylpentane	7.304	57	160219	0.479	ppbv #	90
45) t-1,3-Dichloropropene	8.690	75	20590m	0.367	ppbv	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091624\
 Data File : VL041591.D
 Acq On : 16 Sep 2024 12:36
 Operator : SY/MD
 Sample : VSTDIC0.5
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.5

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/17/2024
 Supervised By :Mahesh Dadoda 09/17/2024

Quant Time: Sep 17 07:23:30 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Tue Sep 17 05:12:12 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	8.121	75	27915	0.331	ppbv	90
47) 1,1,2-Trichloroethane	8.876	97	41983	0.499	ppbv	92
48) Dibromochloromethane	9.687	129	56924m	0.447	ppbv	
49) Bromoform	12.375	173	45890m	0.424	ppbv	
50) 4-Methyl-2-Pentanone	8.176	43	87157m	0.424	ppbv	
51) 2-Hexanone	9.555	43	59285m	0.378	ppbv	
52) Tetrachloroethene	10.587	164	35100m	0.496	ppbv	
53) Toluene	9.192	91	74445	0.355	ppbv	95
54) 1,2-Dibromoethane	9.983	107	52272m	0.509	ppbv	
56) 1,1,1,2-Tetrachloroethane	11.487	131	53872	0.548	ppbv #	1
57) Chlorobenzene	11.503	112	101381m	0.554	ppbv	
58) Ethyl Benzene	12.063	91	107104	0.390	ppbv	94
59) m/p-Xylene	12.349	91	212193m	0.867	ppbv	
60) o-Xylene	13.037	91	99984m	0.428	ppbv	
61) Styrene	12.876	104	31545	0.319	ppbv	94
62) Isopropylbenzene	14.005	105	163927	0.462	ppbv	99
63) 1,1,2,2-Tetrachloroethane	13.021	83	69917m	0.592	ppbv	
64) n-propylbenzene	14.899	120	39445m	0.429	ppbv	
65) tert-Butylbenzene	16.066	119	151018m	0.474	ppbv	
66) Benzyl Chloride	16.304	91	8202m	0.484	ppbv	
67) sec-Butylbenzene	16.616	105	201367	0.442	ppbv	99
69) p-Isopropyltoluene	16.973	119	148897	0.407	ppbv	94
70) n-Butylbenzene	17.870	91	162469	0.426	ppbv	98
71) 2-Chlorotoluene	14.780	91	117206m	0.454	ppbv	
72) 4-Ethyltoluene	15.172	105	108676	0.395	ppbv #	95
73) 1,3,5-Trimethylbenzene	15.333	105	100995m	0.423	ppbv	
74) 1,2,4-Trimethylbenzene	16.079	105	126032	0.442	ppbv	95
75) 1,3-Dichlorobenzene	16.307	146	86427	0.491	ppbv	93
76) 1,4-Dichlorobenzene	16.452	146	76460	0.441	ppbv	87
77) 1,2-Dichlorobenzene	17.117	146	87843m	0.507	ppbv	
78) Hexachloro-1,3-Butadiene	22.609	225	77655m	0.537	ppbv	
79) Naphthalene	21.358	128	72907m	0.360	ppbv	
80) Naphthalene,2-methyl-	24.355	142	34080	0.240	ppbv	92
81) 1,2,4-Trichlorobenzene	21.124	180	51734m	0.392	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091624\
 Data File : VL041592.D
 Acq On : 16 Sep 2024 13:15
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 09/17/2024
 Supervised By : Mahesh Dadoda 09/17/2024

Quant Time: Sep 17 02:35:11 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Sep 17 02:32:57 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.169	49	1153662	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.629	114	3191034	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.445	117	3069576m	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 13.751 95 2284732 9.739 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 97.400%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.709	85	1512734	7.724	ppbv	100
3) Chlorodifluoromethane	2.652	51	1273490	9.220	ppbv	100
4) Chloromethane	2.790	50	704969	9.499	ppbv	100
5) Vinyl Chloride	2.896	62	676970	9.366	ppbv	100
6) Bromomethane	3.095	94	316373	8.660	ppbv	100
7) Chloroethane	3.172	64	246025	9.802	ppbv	100
8) Dichlorotetrafluoroethane	2.835	85	1640863	9.280	ppbv	100
9) Propene	2.671	41	591499	9.296	ppbv	100
10) Heptane	7.555	43	1465058	10.218	ppbv	100
11) Trichlorofluoromethane	3.539	101	1648291	9.296	ppbv	100
12) 1,1,2-Trichlorotrifluo...	4.041	101	1199947	9.385	ppbv	100
13) Ethanol	3.224	45	27404m	5.148	ppbv	
14) Bromoethene	3.343	108	488418	9.820	ppbv	100
15) Acetone	3.443	43	1184767	8.990	ppbv	100
16) 1,3-Butadiene	2.963	39	668292	9.469	ppbv	100
17) tert-Butyl alcohol	3.851	59	1256999	10.440	ppbv	100
18) 1,1-Dichloroethene	3.857	96	549303	10.028	ppbv	100
19) Isopropyl Alcohol	3.552	45	721155	9.470	ppbv	100
20) Methylene Chloride	3.909	84	447422	9.082	ppbv	100
21) Allyl Chloride	3.973	41	846436	9.826	ppbv	100
22) trans-1,2-Dichloroethene	4.420	96	531058	9.732	ppbv	100
23) Vinyl Acetate	4.603	43	796205	8.326	ppbv	100
24) 1,1-Dichloroethane	4.539	63	1144566	9.614	ppbv	100
25) Ethyl Acetate	5.179	43	2636978	9.977	ppbv	100
26) Hexane	5.188	57	1069971	9.824	ppbv	100
27) Carbon Disulfide	4.098	76	1278721	10.141	ppbv	100
28) Methyl tert-Butyl Ether	4.561	73	643194	9.433	ppbv	100
29) Chloroform	5.250	83	1680555	9.485	ppbv	100
30) Cyclohexane	6.577	84	905301	10.235	ppbv	100
31) cis-1,2-Dichloroethene	5.053	61	1038715	9.711	ppbv	100
32) 1,1,1-Trichloroethane	5.979	97	1658808	9.635	ppbv	100
34) 2-Butanone	4.761	43	1501240	9.714	ppbv	100
35) Carbon Tetrachloride	6.471	117	1733798	9.862	ppbv	100
36) Benzene	6.346	78	2276805	9.741	ppbv	100
37) 1,2-Dichloroethane	5.783	62	1272445	9.723	ppbv	100
38) Trichloroethene	7.266	130	1207402	11.030	ppbv	100
39) 1,2-Dichloropropane	7.050	63	877299	9.764	ppbv	100
40) 1,4-Dioxane	7.262	88	362562	10.129	ppbv #	100
41) Tetrahydrofuran	5.532	42	921404	10.354	ppbv	100
42) Bromodichloromethane	7.227	83	1829048	9.717	ppbv	100
43) Methyl Methacrylate	7.449	69	895478	10.728	ppbv	100
44) 2,2,4-Trimethylpentane	7.304	57	3860881	9.842	ppbv	100
45) t-1,3-Dichloropropene	8.687	75	786684	10.580	ppbv	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091624\
 Data File : VL041592.D
 Acq On : 16 Sep 2024 13:15
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 09/17/2024
 Supervised By : Mahesh Dadoda 09/17/2024

Quant Time: Sep 17 02:35:11 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Sep 17 02:32:57 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	8.121	75	1127288	10.315	ppbv	100
47) 1,1,2-Trichloroethane	8.873	97	952785	9.656	ppbv	100
48) Dibromochloromethane	9.687	129	1573314	9.972	ppbv	100
49) Bromoform	12.384	173	1380117	10.374	ppbv	100
50) 4-Methyl-2-Pentanone	8.156	43	2529791	10.343	ppbv	100
51) 2-Hexanone	9.526	43	2011980	10.956	ppbv	100
52) Tetrachloroethene	10.590	164	852321	9.478	ppbv	100
53) Toluene	9.198	91	2661666	10.235	ppbv	100
54) 1,2-Dibromoethane	9.986	107	1299501	10.010	ppbv	100
56) 1,1,1,2-Tetrachloroethane	11.487	131	1161709	9.726	ppbv	100
57) Chlorobenzene	11.503	112	2103559	9.574	ppbv	100
58) Ethyl Benzene	12.066	91	3600157	10.136	ppbv	100
59) m/p-Xylene	12.346	91	6116054m	19.957	ppbv	
60) o-Xylene	13.034	91	2948315	10.140	ppbv	100
61) Styrene	12.873	104	1437066	11.003	ppbv	100
62) Isopropylbenzene	14.005	105	4279685	9.708	ppbv	100
63) 1,1,2,2-Tetrachloroethane	13.024	83	1398913	7.881	ppbv	100
64) n-propylbenzene	14.895	120	1149808	9.986	ppbv	100
65) tert-Butylbenzene	16.066	119	3855478	9.686	ppbv	100
66) Benzyl Chloride	16.304	91	205903	8.606	ppbv	100
67) sec-Butylbenzene	16.612	105	5434135	9.884	ppbv	100
69) p-Isopropyltoluene	16.973	119	4544895	9.982	ppbv	100
70) n-Butylbenzene	17.863	91	4664460	9.989	ppbv	100
71) 2-Chlorotoluene	14.776	91	3221836	9.915	ppbv	100
72) 4-Ethyltoluene	15.169	105	3541706	10.382	ppbv	100
73) 1,3,5-Trimethylbenzene	15.326	105	3032119	10.316	ppbv	100
74) 1,2,4-Trimethylbenzene	16.085	105	3360833	9.827	ppbv	100
75) 1,3-Dichlorobenzene	16.307	146	2088291	9.728	ppbv	100
76) 1,4-Dichlorobenzene	16.452	146	2030349	9.776	ppbv	100
77) 1,2-Dichlorobenzene	17.120	146	2019099	9.498	ppbv	100
78) Hexachloro-1,3-Butadiene	22.609	225	1508307	8.351	ppbv	100
79) Naphthalene	21.358	128	3163696	10.692	ppbv	100
80) Naphthalene,2-methyl-	24.352	142	2300402	9.700	ppbv	100
81) 1,2,4-Trichlorobenzene	21.114	180	1667878	9.460	ppbv	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091624\
 Data File : VL041593.D
 Acq On : 16 Sep 2024 13:56
 Operator : SY/MD
 Sample : VSTDICC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC002

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 09/17/2024
 Supervised By : Mahesh Dadoda 09/17/2024

Quant Time: Sep 17 02:38:44 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Sep 17 02:32:57 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.166	49	1182080	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.629	114	3244465	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.439	117	3052584	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 13.754 95 2207152 9.461 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 94.600%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.706	85	309050	1.540	ppbv	97
3) Chlorodifluoromethane	2.652	51	294373	2.080	ppbv	98
4) Chloromethane	2.793	50	160883	2.116	ppbv	95
5) Vinyl Chloride	2.899	62	147802	1.996	ppbv	95
6) Bromomethane	3.095	94	70434	1.882	ppbv	99
7) Chloroethane	3.173	64	51705	2.010	ppbv	98
8) Dichlorotetrafluoroethane	2.835	85	360778	1.991	ppbv	97
9) Propene	2.671	41	127237	1.952	ppbv	92
10) Heptane	7.555	43	297624	2.026	ppbv	98
11) Trichlorofluoromethane	3.539	101	362790	1.997	ppbv	100
12) 1,1,2-Trichlorotrifluo...	4.044	101	272276	2.078	ppbv	99
13) Ethanol	3.234	45	10961m	2.010	ppbv	
14) Bromoethene	3.340	108	101050	1.983	ppbv	92
15) Acetone	3.449	43	293207	2.171	ppbv	99
16) 1,3-Butadiene	2.964	39	147142	2.035	ppbv	97
17) tert-Butyl alcohol	3.857	59	277171	2.247	ppbv	99
18) 1,1-Dichloroethene	3.857	96	113969	2.031	ppbv	90
19) Isopropyl Alcohol	3.558	45	159227	2.041	ppbv #	97
20) Methylene Chloride	3.909	84	107110	2.122	ppbv	92
21) Allyl Chloride	3.973	41	180432	2.044	ppbv	99
22) trans-1,2-Dichloroethene	4.420	96	115268	2.062	ppbv	92
23) Vinyl Acetate	4.607	43	168539	1.720	ppbv #	97
24) 1,1-Dichloroethane	4.539	63	249059	2.042	ppbv	99
25) Ethyl Acetate	5.185	43	569204	2.102	ppbv	99
26) Hexane	5.185	57	230270	2.063	ppbv	95
27) Carbon Disulfide	4.102	76	237668	1.840	ppbv	98
28) Methyl tert-Butyl Ether	4.565	73	142865	2.045	ppbv	99
29) Chloroform	5.250	83	364849	2.010	ppbv	91
30) Cyclohexane	6.574	84	184390	2.035	ppbv	95
31) cis-1,2-Dichloroethene	5.054	61	224037	2.044	ppbv	93
32) 1,1,1-Trichloroethane	5.980	97	349584	1.982	ppbv	98
34) 2-Butanone	4.764	43	324173	2.063	ppbv	98
35) Carbon Tetrachloride	6.468	117	360310	2.016	ppbv	99
36) Benzene	6.346	78	483918	2.036	ppbv	98
37) 1,2-Dichloroethane	5.780	62	268163	2.015	ppbv	98
38) Trichloroethene	7.266	130	255796	2.298	ppbv	94
39) 1,2-Dichloropropane	7.050	63	192586	2.108	ppbv	97
40) 1,4-Dioxane	7.272	88	74476	2.046	ppbv #	97
41) Tetrahydrofuran	5.542	42	176742	1.953	ppbv	95
42) Bromodichloromethane	7.224	83	371085	1.939	ppbv	95
43) Methyl Methacrylate	7.452	69	169666	1.999	ppbv	99
44) 2,2,4-Trimethylpentane	7.301	57	846282	2.122	ppbv	99
45) t-1,3-Dichloropropene	8.687	75	125807m	1.664	ppbv	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091624\
 Data File : VL041593.D
 Acq On : 16 Sep 2024 13:56
 Operator : SY/MD
 Sample : VSTDICC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC002

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 09/17/2024
 Supervised By : Mahesh Dadoda 09/17/2024

Quant Time: Sep 17 02:38:44 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Sep 17 02:32:57 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	8.118	75	198614	1.787	ppbv	98
47) 1,1,2-Trichloroethane	8.870	97	202286	2.016	ppbv	95
48) Dibromochloromethane	9.684	129	299312	1.866	ppbv	98
49) Bromoform	12.384	173	259933	1.922	ppbv	98
50) 4-Methyl-2-Pentanone	8.163	43	521882	2.099	ppbv	100
51) 2-Hexanone	9.536	43	391600	2.097	ppbv #	99
52) Tetrachloroethene	10.594	164	179606	1.964	ppbv	97
53) Toluene	9.198	91	523040	1.978	ppbv	99
54) 1,2-Dibromoethane	9.986	107	260693	1.975	ppbv	99
56) 1,1,1,2-Tetrachloroethane	11.484	131	247944	2.087	ppbv #	1
57) Chlorobenzene	11.503	112	451222	2.065	ppbv	96
58) Ethyl Benzene	12.066	91	719793	2.038	ppbv	96
59) m/p-Xylene	12.346	91	1294945m	4.249	ppbv	
60) o-Xylene	13.034	91	620587	2.146	ppbv	100
61) Styrene	12.880	104	252259	1.942	ppbv	98
62) Isopropylbenzene	14.005	105	929045	2.119	ppbv	99
63) 1,1,2,2-Tetrachloroethane	13.024	83	311063	1.762	ppbv	98
64) n-propylbenzene	14.892	120	244643	2.137	ppbv	93
65) tert-Butylbenzene	16.069	119	874824	2.210	ppbv	99
66) Benzyl Chloride	16.310	91	39929m	1.678	ppbv	
67) sec-Butylbenzene	16.613	105	1221574	2.234	ppbv	99
69) p-Isopropyltoluene	16.973	119	1023815	2.261	ppbv	99
70) n-Butylbenzene	17.863	91	1057176	2.276	ppbv	99
71) 2-Chlorotoluene	14.777	91	684496	2.118	ppbv	99
72) 4-Ethyltoluene	15.172	105	730570	2.153	ppbv	99
73) 1,3,5-Trimethylbenzene	15.330	105	643556	2.202	ppbv	96
74) 1,2,4-Trimethylbenzene	16.082	105	773729	2.275	ppbv	97
75) 1,3-Dichlorobenzene	16.307	146	467558	2.190	ppbv	99
76) 1,4-Dichlorobenzene	16.449	146	447493	2.167	ppbv	98
77) 1,2-Dichlorobenzene	17.121	146	460299	2.177	ppbv	99
78) Hexachloro-1,3-Butadiene	22.606	225	387729	2.159	ppbv	100
79) Naphthalene	21.362	128	675872	2.297	ppbv	97
80) Naphthalene,2-methyl-	24.358	142	374380	1.587	ppbv	99
81) 1,2,4-Trichlorobenzene	21.114	180	370420	2.113	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091624\
Data File : VL041593.D
Acq On : 16 Sep 2024 13:56
Operator : SY/MD
Sample : VSTDIC002
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC002

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 09/17/2024
Supervised By : Mahesh Dadoda 09/17/2024


Quant Time: Sep 17 02:38:44 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug

QLast Update : Tue Sep 17 02:32:57 2024

Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091624\
 Data File : VL041594.D
 Acq On : 16 Sep 2024 14:36
 Operator : SY/MD
 Sample : VSTDICC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/17/2024
 Supervised By :Mahesh Dadoda 09/17/2024

Quant Time: Sep 16 20:15:06 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Mon Sep 16 20:14:05 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.172	49	1163991	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.632	114	3218358	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.445	117	2901355	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 13.754 95 2171693 9.726 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 97.300%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.716	85	176543	0.882	ppbv	100
3) Chlorodifluoromethane	2.658	51	143593	1.036	ppbv	99
4) Chloromethane	2.796	50	78805m	1.069	ppbv	
5) Vinyl Chloride	2.906	62	73731	1.015	ppbv	99
6) Bromomethane	3.105	94	32268	0.864	ppbv	86
7) Chloroethane	3.182	64	25889	1.033	ppbv	91
8) Dichlorotetrafluoroethane	2.841	85	179916	1.014	ppbv	91
9) Propene	2.681	41	60205	0.946	ppbv	94
10) Heptane	7.565	43	136494	0.954	ppbv	99
11) Trichlorofluoromethane	3.545	101	190406	1.071	ppbv	99
12) 1,1,2-Trichlorotrifluo...	4.050	101	135246	1.056	ppbv	98
13) Ethanol	3.246	45	6603m	1.161	ppbv	
14) Bromoethene	3.353	108	48959	0.981	ppbv #	92
15) Acetone	3.462	43	144857	1.095	ppbv	97
16) 1,3-Butadiene	2.970	39	71179	1.007	ppbv	100
17) tert-Butyl alcohol	3.867	59	132423	1.098	ppbv	96
18) 1,1-Dichloroethene	3.861	96	57053	1.042	ppbv	91
19) Isopropyl Alcohol	3.571	45	76655	0.991	ppbv #	95
20) Methylene Chloride	3.915	84	55627	1.128	ppbv	97
21) Allyl Chloride	3.976	41	83509	0.963	ppbv	89
22) trans-1,2-Dichloroethene	4.426	96	53748	0.981	ppbv	96
23) Vinyl Acetate	4.610	43	75873	0.751	ppbv #	93
24) 1,1-Dichloroethane	4.542	63	122972	1.034	ppbv	99
25) Ethyl Acetate	5.192	43	266035	1.013	ppbv	98
26) Hexane	5.195	57	107564	0.987	ppbv	92
27) Carbon Disulfide	4.105	76	109406	0.858	ppbv	96
28) Methyl tert-Butyl Ether	4.574	73	67246	0.978	ppbv	99
29) Chloroform	5.253	83	183121	1.030	ppbv	96
30) Cyclohexane	6.577	84	77992	0.879	ppbv	87
31) cis-1,2-Dichloroethene	5.060	61	103032	0.957	ppbv	98
32) 1,1,1-Trichloroethane	5.979	97	168967	0.976	ppbv	98
34) 2-Butanone	4.774	43	140700	0.904	ppbv	95
35) Carbon Tetrachloride	6.475	117	166616	0.939	ppbv	97
36) Benzene	6.349	78	217922	0.927	ppbv	97
37) 1,2-Dichloroethane	5.787	62	130607	0.992	ppbv	97
38) Trichloroethene	7.269	130	114753	1.065	ppbv	94
39) 1,2-Dichloropropane	7.050	63	89452	0.994	ppbv	92
40) 1,4-Dioxane	7.294	88	32240	0.891	ppbv #	87
41) Tetrahydrofuran	5.552	42	82665m	0.928	ppbv	
42) Bromodichloromethane	7.230	83	179366	0.944	ppbv	99
43) Methyl Methacrylate	7.462	69	72483	0.860	ppbv	93
44) 2,2,4-Trimethylpentane	7.304	57	389526	0.992	ppbv	96
45) t-1,3-Dichloropropene	8.693	75	50543	0.651	ppbv	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091624\
 Data File : VL041594.D
 Acq On : 16 Sep 2024 14:36
 Operator : SY/MD
 Sample : VSTDIC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC001

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/17/2024
 Supervised By :Mahesh Dadoda 09/17/2024

Quant Time: Sep 16 20:15:06 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Mon Sep 16 20:14:05 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	8.118	75	82468	0.736	ppbv	99
47) 1,1,2-Trichloroethane	8.873	97	97460	0.982	ppbv	94
48) Dibromochloromethane	9.687	129	137881	0.862	ppbv	99
49) Bromoform	12.388	173	116706	0.864	ppbv	97
50) 4-Methyl-2-Pentanone	8.172	43	224369	0.913	ppbv	99
51) 2-Hexanone	9.552	43	163976m	0.883	ppbv	
52) Tetrachloroethene	10.593	164	86388	0.951	ppbv	94
53) Toluene	9.198	91	226360	0.864	ppbv	99
54) 1,2-Dibromoethane	9.986	107	121996	0.924	ppbv	96
56) 1,1,1,2-Tetrachloroethane	11.487	131	116306	1.034	ppbv #	1
57) Chlorobenzene	11.503	112	222410	1.077	ppbv #	94
58) Ethyl Benzene	12.072	91	318095	0.951	ppbv	97
59) m/p-Xylene	12.346	91	564895m	1.957	ppbv	
60) o-Xylene	13.037	91	270052	0.989	ppbv	100
61) Styrene	12.879	104	98015	0.796	ppbv	97
62) Isopropylbenzene	14.008	105	423939	1.021	ppbv	98
63) 1,1,2,2-Tetrachloroethane	13.027	83	154402	0.895	ppbv	99
64) n-propylbenzene	14.895	120	103274	0.952	ppbv	80
65) tert-Butylbenzene	16.072	119	388333	1.036	ppbv	98
66) Benzyl Chloride	16.313	91	18633m	0.775	ppbv	
67) sec-Butylbenzene	16.616	105	546134	1.057	ppbv	99
69) p-Isopropyltoluene	16.976	119	428199	0.999	ppbv	97
70) n-Butylbenzene	17.870	91	456221	1.038	ppbv	96
71) 2-Chlorotoluene	14.780	91	303828	0.992	ppbv	100
72) 4-Ethyltoluene	15.172	105	315109	0.980	ppbv #	97
73) 1,3,5-Trimethylbenzene	15.330	105	282986	1.022	ppbv	97
74) 1,2,4-Trimethylbenzene	16.082	105	342646m	1.064	ppbv	
75) 1,3-Dichlorobenzene	16.307	146	206679	1.021	ppbv	96
76) 1,4-Dichlorobenzene	16.452	146	205278	1.047	ppbv	97
77) 1,2-Dichlorobenzene	17.124	146	206345	1.028	ppbv	96
78) Hexachloro-1,3-Butadiene	22.609	225	180258m	1.053	ppbv	
79) Naphthalene	21.368	128	252197m	0.887	ppbv	
80) Naphthalene,2-methyl-	24.358	142	126259m	0.530	ppbv	
81) 1,2,4-Trichlorobenzene	21.124	180	145946	0.863	ppbv	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091624\
 Data File : VL041596.D
 Acq On : 16 Sep 2024 15:56
 Operator : SY/MD
 Sample : VSTDICC0.1
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.1

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/17/2024
 Supervised By :Mahesh Dadoda 09/17/2024

Quant Time: Sep 17 03:12:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Tue Sep 17 02:32:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.176	49	1150062	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.632	114	3120058	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.445	117	2779435	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.754	95	2086516	9.823	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.200%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	2.915	62	6034m	0.084	ppbv	
32) 1,1,1-Trichloroethane	5.986	97	13088m	0.076	ppbv	
35) Carbon Tetrachloride	6.468	117	12312m	0.072	ppbv	
38) Trichloroethene	7.272	130	9519m	0.089	ppbv	
52) Tetrachloroethene	10.590	164	6744m	0.077	ppbv	
54) 1,2-Dibromoethane	9.979	107	8094m	0.064	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.037	83	10963m	0.068	ppbv	
79) Naphthalene	21.374	128	8646m	0.032	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091624\
Data File : VL041596.D
Acq On : 16 Sep 2024 15:56
Operator : SY/MD
Sample : VSTDICC0.1
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 09/17/2024
Supervised By :Mahesh Dadoda 09/17/2024

Quant Time: Sep 17 03:12:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Tue Sep 17 02:32:57 2024
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091624\
 Data File : VL041597.D
 Acq On : 16 Sep 2024 16:38
 Operator : SY/MD
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/17/2024
 Supervised By :Mahesh Dadoda 09/17/2024

Quant Time: Sep 17 05:05:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LFri Aug 2
 QLast Update : Tue Sep 17 04:57:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.173	49	1115440	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.632	114	3030679	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.439	117	2692276	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.751	95	1983531	9.951	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.500%
Target Compounds						
5) Vinyl Chloride	2.906	62	1537m	0.024	ppbv	Qvalue
32) 1,1,1-Trichloroethane	5.976	97	3900m	0.026	ppbv	
35) Carbon Tetrachloride	6.475	117	3299m	0.023	ppbv	
38) Trichloroethene	7.269	130	3158m	0.029	ppbv	
52) Tetrachloroethene	10.581	164	1981m	0.026	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.040	83	3376m	0.027	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091624\
Data File : VL041597.D
Acq On : 16 Sep 2024 16:38
Operator : SY/MD
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 09/17/2024
Supervised By :Mahesh Dadoda 09/17/2024

Quant Time: Sep 17 05:05:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Tue Sep 17 04:57:28 2024
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091624\
 Data File : VL041598.D
 Acq On : 16 Sep 2024 17:20
 Operator : SY/MD
 Sample : VSTDIC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC015

Quant Time: Sep 17 09:14:19 2024
 Quant Title :
 QLast Update : Tue Sep 17 09:12:21 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 09/17/2024
 Supervised By :Mahesh Dadoda 09/17/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.176	49	1175483	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.639	114	3256634	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.449	117	3327610m	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 13.757 95 2332788 9.110 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 91.100%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.719	85	1900049	9.396	ppbv	100
3) Chlorodifluoromethane	2.664	51	1860879	13.298	ppbv	99
4) Chloromethane	2.803	50	1044613m	14.035	ppbv	
5) Vinyl Chloride	2.909	62	1021949	13.930	ppbv	99
6) Bromomethane	3.105	94	462431	12.264	ppbv	99
7) Chloroethane	3.185	64	360774	14.249	ppbv	98
8) Dichlorotetrafluoroethane	2.845	85	2287936	12.769	ppbv	99
9) Propene	2.684	41	862437	13.416	ppbv	100
10) Heptane	7.561	43	2181339	15.103	ppbv	100
11) Trichlorofluoromethane	3.549	101	2384942	13.280	ppbv	99
12) 1,1,2-Trichlorotrifluo...	4.050	101	1737916	13.439	ppbv	99
13) Ethanol	3.234	45	32379m	5.636	ppbv	
14) Bromoethene	3.353	108	722975	14.340	ppbv	100
15) Acetone	3.452	43	1746071	13.073	ppbv	99
16) 1,3-Butadiene	2.973	39	997661	13.981	ppbv	99
17) tert-Butyl alcohol	3.861	59	1826662	14.996	ppbv	99
18) 1,1-Dichloroethene	3.864	96	783989	14.172	ppbv	98
19) Isopropyl Alcohol	3.562	45	998621	12.789	ppbv	100
20) Methylene Chloride	3.918	84	642783	12.906	ppbv	93
21) Allyl Chloride	3.979	41	1225956	13.997	ppbv	99
22) trans-1,2-Dichloroethene	4.430	96	781882	14.136	ppbv	99
23) Vinyl Acetate	4.610	43	1212589	11.888	ppbv #	98
24) 1,1-Dichloroethane	4.549	63	1656513	13.792	ppbv	99
25) Ethyl Acetate	5.185	43	3850317	14.511	ppbv	99
26) Hexane	5.195	57	1577948	14.336	ppbv	99
27) Carbon Disulfide	4.108	76	1898447	14.743	ppbv	99
28) Methyl tert-Butyl Ether	4.568	73	934921	13.469	ppbv	100
29) Chloroform	5.259	83	2472134	13.770	ppbv	98
30) Cyclohexane	6.584	84	1384242	15.443	ppbv	96
31) cis-1,2-Dichloroethene	5.063	61	1572482	14.462	ppbv	99
32) 1,1,1-Trichloroethane	5.986	97	2449235	14.015	ppbv	99
34) 2-Butanone	4.767	43	2226315	14.132	ppbv	99
35) Carbon Tetrachloride	6.478	117	2562607	14.267	ppbv	99
36) Benzene	6.352	78	3400759	14.291	ppbv	98
37) 1,2-Dichloroethane	5.790	62	1894411	14.224	ppbv	99
38) Trichloroethene	7.272	130	1777141	16.301	ppbv	97
39) 1,2-Dichloropropane	7.057	63	1318620	14.474	ppbv	99
40) 1,4-Dioxane	7.266	88	453503	12.392	ppbv #	47
41) Tetrahydrofuran	5.536	42	1391996	15.449	ppbv	100
42) Bromodichloromethane	7.233	83	2707351	14.078	ppbv	97
43) Methyl Methacrylate	7.455	69	1368120	16.033	ppbv	98
44) 2,2,4-Trimethylpentane	7.307	57	5619291	14.138	ppbv	99
45) t-1,3-Dichloropropene	8.693	75	1264444	16.099	ppbv	100
46) cis-1,3-Dichloropropene	8.124	75	1733083	15.281	ppbv	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091624\
 Data File : VL041598.D
 Acq On : 16 Sep 2024 17:20
 Operator : SY/MD
 Sample : VSTDIC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC015

Quant Time: Sep 17 09:14:19 2024
 Quant Title :
 QLast Update : Tue Sep 17 09:12:21 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/17/2024
 Supervised By :Mahesh Dadoda 09/17/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
47) 1,1,2-Trichloroethane	8.880	97	1416240	14.103	ppbv	97
48) Dibromochloromethane	9.693	129	2380715	14.716	ppbv	99
49) Bromoform	12.388	173	2103505	15.393	ppbv	100
50) 4-Methyl-2-Pentanone	8.159	43	3768387	15.162	ppbv	99
51) 2-Hexanone	9.529	43	3039864	16.175	ppbv	98
52) Tetrachloroethene	10.597	164	1298509	14.122	ppbv	99
53) Toluene	9.204	91	3987119	15.042	ppbv	99
54) 1,2-Dibromoethane	9.989	107	1951435	14.599	ppbv	100
56) 1,1,1,2-Tetrachloroethane	11.490	131	1724113	13.360	ppbv #	92
57) Chlorobenzene	11.507	112	3100633	13.087	ppbv	98
58) Ethyl Benzene	12.069	91	5323752	13.880	ppbv	100
59) m/p-Xylene	12.352	91	9021854m	27.249	ppbv	
60) o-Xylene	13.037	91	4380559	13.986	ppbv	99
61) Styrene	12.879	104	2207148	15.620	ppbv	100
62) Isopropylbenzene	14.011	105	6290796	13.212	ppbv	99
63) 1,1,2,2-Tetrachloroethane	13.027	83	2059305	10.412	ppbv	99
64) n-propylbenzene	14.899	120	1733510	13.926	ppbv	94
65) tert-Butylbenzene	16.072	119	5661694	13.165	ppbv	99
66) Benzyl Chloride	16.304	91	343877	12.466	ppbv	97
67) sec-Butylbenzene	16.616	105	7844186	13.238	ppbv	98
69) p-Isopropyltoluene	16.976	119	6613794	13.451	ppbv	99
70) n-Butylbenzene	17.863	91	6825900	13.542	ppbv	99
71) 2-Chlorotoluene	14.780	91	4774971	13.599	ppbv	99
72) 4-Ethyltoluene	15.172	105	5251979	14.243	ppbv	98
73) 1,3,5-Trimethylbenzene	15.330	105	4527850	14.254	ppbv	98
74) 1,2,4-Trimethylbenzene	16.088	105	4937113	13.368	ppbv	98
75) 1,3-Dichlorobenzene	16.313	146	3113298	13.410	ppbv	99
76) 1,4-Dichlorobenzene	16.455	146	3063921	13.630	ppbv	99
77) 1,2-Dichlorobenzene	17.120	146	3015374	13.102	ppbv	100
78) Hexachloro-1,3-Butadiene	22.609	225	2302119	11.730	ppbv	99
79) Naphthalene	21.358	128	4916804	15.084	ppbv	99
80) Naphthalene,2-methyl-	24.355	142	3899900	14.273	ppbv	99
81) 1,2,4-Trichlorobenzene	21.117	180	2616194	13.487	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091624\
 Data File : VL041599.D
 Acq On : 16 Sep 2024 18:46
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICSVVL091624

Quant Time: Sep 17 09:18:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Tue Sep 17 09:12:21 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 09/17/2024
 Supervised By : Mahesh Dadoda 09/17/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.176	49	1211658	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.636	114	3308799	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.449	117	3161469m	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.751	95	2308391	9.919	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.200%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.723	85	1544917	8.829	ppbv	99
3) Chlorodifluoromethane	2.668	51	1323860	9.008	ppbv	99
4) Chloromethane	2.806	50	724772	8.966	ppbv	95
5) Vinyl Chloride	2.912	62	707659	10.111	ppbv	99
6) Bromomethane	3.108	94	319232	9.168	ppbv	99
7) Chloroethane	3.189	64	251821	9.251	ppbv	99
8) Dichlorotetrafluoroethane	2.848	85	1641100	8.941	ppbv	98
9) Propene	2.687	41	640849	10.338	ppbv	100
10) Heptane	7.565	43	1512434	10.420	ppbv	100
11) Trichlorofluoromethane	3.552	101	1652570	8.875	ppbv	99
12) 1,1,2-Trichlorotrifluo...	4.054	101	1197204	8.877	ppbv	99
13) Ethanol	3.240	45	22445m	4.816	ppbv	
14) Bromoethene	3.356	108	497665	9.732	ppbv	98
15) Acetone	3.456	43	1235854	8.124	ppbv	99
16) 1,3-Butadiene	2.977	39	708349	9.574	ppbv	99
17) tert-Butyl alcohol	3.864	59	1338127	10.002	ppbv	100
18) 1,1-Dichloroethene	3.867	96	531734	9.314	ppbv	99
19) Isopropyl Alcohol	3.562	45	679011	8.720	ppbv	99
20) Methylene Chloride	3.919	84	447442	7.830	ppbv	95
21) Allyl Chloride	3.983	41	870229	9.870	ppbv	100
22) trans-1,2-Dichloroethene	4.433	96	544525	9.561	ppbv	99
23) Vinyl Acetate	4.610	43	879864	10.516	ppbv #	97
24) 1,1-Dichloroethane	4.549	63	1155920	9.204	ppbv	99
25) Ethyl Acetate	5.189	43	2756002	10.092	ppbv	99
26) Hexane	5.195	57	1129442	9.917	ppbv	99
27) Carbon Disulfide	4.112	76	1301729	10.753	ppbv	99
28) Methyl tert-Butyl Ether	4.571	73	708243	10.087	ppbv	99
29) Chloroform	5.259	83	1692873	9.084	ppbv	100
30) Cyclohexane	6.587	84	939832	10.706	ppbv	98
31) cis-1,2-Dichloroethene	5.063	61	1109424	10.261	ppbv	98
32) 1,1,1-Trichloroethane	5.986	97	1671780	10.071	ppbv	99
34) 2-Butanone	4.771	43	1595025	10.110	ppbv	99
35) Carbon Tetrachloride	6.475	117	1729977	10.656	ppbv	98
36) Benzene	6.353	78	2377400	10.346	ppbv	100
37) 1,2-Dichloroethane	5.790	62	1303965	9.780	ppbv	99
38) Trichloroethene	7.272	130	1220709	10.316	ppbv	99
39) 1,2-Dichloropropane	7.057	63	924715	9.892	ppbv	97
40) 1,4-Dioxane	7.269	88	268441m	7.618	ppbv	
41) Tetrahydrofuran	5.539	42	998000	11.494	ppbv	100
42) Bromodichloromethane	7.230	83	1853470	9.992	ppbv	99
43) Methyl Methacrylate	7.455	69	952193	11.579	ppbv	98
44) 2,2,4-Trimethylpentane	7.308	57	3942109	9.916	ppbv	99
45) t-1,3-Dichloropropene	8.693	75	854719	12.892	ppbv	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091624\
 Data File : VL041599.D
 Acq On : 16 Sep 2024 18:46
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL091624

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 09/17/2024
 Supervised By : Mahesh Dadoda 09/17/2024

Quant Time: Sep 17 09:18:19 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Sep 17 09:12:21 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	8.124	75	1179959	12.148	ppbv	98
47) 1,1,2-Trichloroethane	8.880	97	965904	9.731	ppbv	97
48) Dibromochloromethane	9.690	129	1596069	10.613	ppbv	100
49) Bromoform	12.388	173	1398574	10.834	ppbv	99
50) 4-Methyl-2-Pentanone	8.163	43	2618515	10.747	ppbv	99
51) 2-Hexanone	9.529	43	2107044	11.433	ppbv #	99
52) Tetrachloroethene	10.594	164	876069	10.529	ppbv	97
53) Toluene	9.201	91	2736931	11.220	ppbv	100
54) 1,2-Dibromoethane	9.989	107	1334237	10.916	ppbv	99
56) 1,1,1,2-Tetrachloroethane	11.487	131	1153971	9.349	ppbv	100
57) Chlorobenzene	11.507	112	2110664	9.260	ppbv	98
58) Ethyl Benzene	12.066	91	3636857	10.746	ppbv	100
59) m/p-Xylene	12.352	91	6119550m	20.317	ppbv	
60) o-Xylene	13.034	91	2941243	10.185	ppbv	99
61) Styrene	12.873	104	1461911	12.117	ppbv	100
62) Isopropylbenzene	14.008	105	4273531	9.768	ppbv	100
63) 1,1,2,2-Tetrachloroethane	13.028	83	1389186	9.409	ppbv	99
64) n-propylbenzene	14.896	120	1144513	10.129	ppbv	99
65) tert-Butylbenzene	16.069	119	3822907	9.532	ppbv	99
66) Benzyl Chloride	16.304	91	212798	10.206	ppbv	100
67) sec-Butylbenzene	16.613	105	5350490	9.616	ppbv	100
69) p-Isopropyltoluene	16.973	119	4497521	9.986	ppbv	100
70) n-Butylbenzene	17.863	91	4595438	9.740	ppbv	100
71) 2-Chlorotoluene	14.780	91	3223465	10.013	ppbv	100
72) 4-Ethyltoluene	15.169	105	3524544	10.441	ppbv	99
73) 1,3,5-Trimethylbenzene	15.327	105	3014196	10.112	ppbv	98
74) 1,2,4-Trimethylbenzene	16.085	105	3320425	9.517	ppbv	99
75) 1,3-Dichlorobenzene	16.307	146	2073354	9.482	ppbv	100
76) 1,4-Dichlorobenzene	16.452	146	2027061	9.675	ppbv	99
77) 1,2-Dichlorobenzene	17.117	146	1996843	9.250	ppbv	100
78) Hexachloro-1,3-Butadiene	22.612	225	1507968m	8.468	ppbv	
79) Naphthalene	21.355	128	3203623	12.476	ppbv	98
80) Naphthalene,2-methyl-	24.349	142	2446937	13.599	ppbv	99
81) 1,2,4-Trichlorobenzene	21.114	180	1671079	10.236	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091624\
 Data File : VL041599.D
 Acq On : 16 Sep 2024 18:46
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL091624

Quant Time: Sep 17 09:18:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Sep 17 09:12:21 2024
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	105	0.00
2 T	Dichlorodifluoromethane	1.444	1.275	11.7	102	0.01
3	Chlorodifluoromethane	1.213	1.093	9.9	104	0.02
4	Chloromethane	0.667	0.598	10.3	103	0.02
5 T	Vinyl Chloride	0.578	0.584	-1.0	105	0.02
6 T	Bromomethane	0.287	0.263	8.4	101	0.01
7	Chloroethane	0.225	0.208	7.6	102	0.02
8 T	Dichlorotetrafluoroethane	1.515	1.354	10.6	100	0.01
9 T	Propene	0.512	0.529	-3.3	108	0.02
10 T	Heptane	1.198	1.248	-4.2	103	0.00
11 T	Trichlorofluoromethane	1.537	1.364	11.3	100	0.01
12 T	1,1,2-Trichlorotrifluoroeth	1.113	0.988	11.2	100	0.01
13	Ethanol	0.038	0.019#	50.0#	82	0.02
14 T	Bromoethene	0.422	0.411	2.6	102	0.01
15 T	Acetone	1.255	1.020	18.7	104	0.01
16 T	1,3-Butadiene	0.611	0.585	4.3	106	0.01
17	tert-Butyl alcohol	1.104	1.104	0.0	106	0.01
18 T	1,1-Dichloroethene	0.471	0.439	6.8	97	0.00
19 T	Isopropyl Alcohol	0.643	0.560	12.9	94	0.00
20 T	Methylene Chloride	0.472	0.369	21.8	100	0.00
21 T	Allyl Chloride	0.728	0.718	1.4	103	0.00
22 T	trans-1,2-Dichloroethene	0.470	0.449	4.5	103	0.01
23 T	Vinyl Acetate	0.691	0.726	-5.1	111	0.00
24 T	1,1-Dichloroethane	1.036	0.954	7.9	101	0.00
25 T	Ethyl Acetate	2.254	2.275	-0.9	105	0.00
26 T	Hexane	0.940	0.932	0.9	106	0.00
27 T	Carbon Disulfide	0.999	1.074	-7.5	102	0.01
28 T	Methyl tert-Butyl Ether	0.579	0.585	-1.0	110	0.00
29 T	Chloroform	1.538	1.397	9.2	101	0.00
30 T	Cyclohexane	0.725	0.776	-7.0	104	0.00
31 T	cis-1,2-Dichloroethene	0.892	0.916	-2.7	107	0.00
32 T	1,1,1-Trichloroethane	1.370	1.380	-0.7	101	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	104	0.00
34 T	2-Butanone	0.477	0.482	-1.0	106	0.00
35 T	Carbon Tetrachloride	0.491	0.523	-6.5	100	0.00
36 T	Benzene	0.694	0.719	-3.6	104	0.00
37 T	1,2-Dichloroethane	0.403	0.394	2.2	102	0.00
38 T	Trichloroethene	0.358	0.369	-3.1	101	0.00
39 T	1,2-Dichloropropane	0.283	0.279	1.4	105	0.00
40 T	1,4-Dioxane	0.107	0.081	24.3	74	0.00
41 T	Tetrahydrofuran	0.262	0.302	-15.3	108	0.00
42 T	Bromodichloromethane	0.561	0.560	0.2	101	0.00
43	Methyl Methacrylate	0.249	0.288	-15.7	106	0.00
44 T	2,2,4-Trimethylpentane	1.202	1.191	0.9	102	0.00
45 T	t-1,3-Dichloropropene	0.200	0.258	-29.0	109	0.00
46 T	cis-1,3-Dichloropropene	0.294	0.357	-21.4	105	0.00
47 T	1,1,2-Trichloroethane	0.300	0.292	2.7	101	0.00
48 T	Dibromochloromethane	0.455	0.482	-5.9	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091624\
 Data File : VL041599.D
 Acq On : 16 Sep 2024 18:46
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL091624

Quant Time: Sep 17 09:18:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Sep 17 09:12:21 2024
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.390	0.423	-8.5	101	0.00
50 T	4-Methyl-2-Pentanone	0.736	0.791	-7.5	104	0.00
51 T	2-Hexanone	0.557	0.637	-14.4	105	0.00
52 T	Tetrachloroethene	0.251	0.265	-5.6	103	0.00
53 T	Toluene	0.737	0.827	-12.2	103	0.00
54 T	1,2-Dibromoethane	0.369	0.403	-9.2	103	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	103	0.00
56	1,1,1,2-Tetrachloroethane	0.390	0.365	6.4	99	0.00
57 T	Chlorobenzene	0.721	0.668	7.4	100	0.00
58 T	Ethyl Benzene	1.070	1.150	-7.5	101	0.00
59 T	m/p-Xylene	0.953	0.968	-1.6	100	0.00
60 T	o-Xylene	0.913	0.930	-1.9	100	0.00
61 T	Styrene	0.382	0.462	-20.9	102	0.00
62	Isopropylbenzene	1.384	1.352	2.3	100	0.00
63 T	1,1,2,2-Tetrachloroethane	0.467	0.439	6.0	99	0.00
64	n-propylbenzene	0.357	0.362	-1.4	100	0.00
65	tert-Butylbenzene	1.269	1.209	4.7	99	0.00
66 T	Benzyl Chloride	0.066	0.067	-1.5	103	0.00
67	sec-Butylbenzene	1.760	1.692	3.9	98	0.00
68 S	1-Bromo-4-Fluorobenzene	0.736	0.730	0.8	101	0.00
69	p-Isopropyltoluene	1.425	1.423	0.1	99	0.00
70	n-Butylbenzene	1.492	1.454	2.5	99	0.00
71	2-Chlorotoluene	1.018	1.020	-0.2	100	0.00
72 T	4-Ethyltoluene	1.068	1.115	-4.4	100	0.00
73 T	1,3,5-Trimethylbenzene	0.943	0.953	-1.1	99	0.00
74 T	1,2,4-Trimethylbenzene	1.104	1.050	4.9	99	0.00
75 T	1,3-Dichlorobenzene	0.692	0.656	5.2	99	0.00
76 T	1,4-Dichlorobenzene	0.663	0.641	3.3	100	0.00
77 T	1,2-Dichlorobenzene	0.683	0.632	7.5	99	0.00
78 T	Hexachloro-1,3-Butadiene	0.563	0.477	15.3	100	0.00
79 T	Naphthalene	0.812	1.013	-24.8	101	0.00
80 T	Naphthalene,2-methyl-	0.569	0.774	-36.0#	106	0.00
81 T	1,2,4-Trichlorobenzene	0.516	0.529	-2.5	100	0.00

(#) = Out of Range


SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091824\
Data File : VL041600.D
Acq On : 18 Sep 2024 8:06
Operator : SY/MD
Sample : BFB
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
Last Update : Tue Sep 17 09:12:21 2024



AutoFind: Scans 3471, 3472, 3473; Background Corrected with Scan 3447

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	8	40	24.0	208555	PASS
75	95	30	66	55.8	485056	PASS
95	95	100	100	100.0	869077	PASS
96	95	5	9	7.0	60467	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	68.9	599040	PASS
175	174	4	9	7.6	45741	PASS
176	174	93	101	98.5	589888	PASS
177	176	5	9	6.7	39323	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091824\
 Data File : VL041601.D
 Acq On : 18 Sep 2024 8:46
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Quant Time: Sep 19 01:28:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Tue Sep 17 09:12:21 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 09/19/2024
 Supervised By : Mahesh Dadoda 09/19/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.185	49	1176243	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	6.645	114	3208665	10.000	ppbv	0.02
55) Chlorobenzene-d5	11.458	117	3110660	10.000	ppbv	0.01
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.767	95	2275993	9.939	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.400%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.726	85	1344207	7.913	ppbv	99
3) Chlorodifluoromethane	2.671	51	1296141	9.085	ppbv	99
4) Chloromethane	2.809	50	708455	9.028	ppbv	97
5) Vinyl Chloride	2.915	62	705961	10.390	ppbv	98
6) Bromomethane	3.115	94	311784	9.223	ppbv	97
7) Chloroethane	3.192	64	251005	9.499	ppbv	98
8) Dichlorotetrafluoroethane	2.851	85	1646174	9.238	ppbv	100
9) Propene	2.690	41	612014	10.170	ppbv	98
10) Heptane	7.571	43	1478682	10.495	ppbv	100
11) Trichlorofluoromethane	3.555	101	1639385	9.069	ppbv	100
12) 1,1,2-Trichlorotrifluo...	4.060	101	1214036	9.273	ppbv	100
13) Ethanol	3.243	45	26533m	5.865	ppbv	
14) Bromoethene	3.359	108	499377	10.060	ppbv	100
15) Acetone	3.462	43	1208787	8.186	ppbv	100
16) 1,3-Butadiene	2.980	39	673454	9.376	ppbv	98
17) tert-Butyl alcohol	3.870	59	1261877	9.716	ppbv	100
18) 1,1-Dichloroethene	3.873	96	530425	9.571	ppbv	97
19) Isopropyl Alcohol	3.568	45	715591	9.466	ppbv #	98
20) Methylene Chloride	3.925	84	450417	8.119	ppbv	91
21) Allyl Chloride	3.989	41	858301	10.028	ppbv	100
22) trans-1,2-Dichloroethene	4.436	96	534697	9.671	ppbv	98
23) Vinyl Acetate	4.619	43	863997	10.637	ppbv	99
24) 1,1-Dichloroethane	4.555	63	1158084	9.499	ppbv	100
25) Ethyl Acetate	5.198	43	2686598	10.134	ppbv	100
26) Hexane	5.205	57	1094009	9.896	ppbv	99
27) Carbon Disulfide	4.118	76	1309474	11.142	ppbv	99
28) Methyl tert-Butyl Ether	4.578	73	673643	9.883	ppbv	98
29) Chloroform	5.266	83	1702540	9.411	ppbv	99
30) Cyclohexane	6.594	84	911409	10.695	ppbv	98
31) cis-1,2-Dichloroethene	5.073	61	1082135	10.310	ppbv	100
32) 1,1,1-Trichloroethane	5.996	97	1681151	10.433	ppbv	99
34) 2-Butanone	4.780	43	1525364	9.970	ppbv	99
35) Carbon Tetrachloride	6.484	117	1762225	11.193	ppbv	97
36) Benzene	6.362	78	2328102	10.448	ppbv	99
37) 1,2-Dichloroethane	5.799	62	1282907	9.922	ppbv	100
38) Trichloroethene	7.285	130	1247903	10.874	ppbv	99
39) 1,2-Dichloropropane	7.063	63	902201	9.952	ppbv	98
40) 1,4-Dioxane	7.278	88	375300	10.982	ppbv #	97
41) Tetrahydrofuran	5.549	42	940519	11.170	ppbv	99
42) Bromodichloromethane	7.240	83	1854610	10.310	ppbv	100
43) Methyl Methacrylate	7.465	69	916154	11.489	ppbv	99
44) 2,2,4-Trimethylpentane	7.320	57	3916909	10.160	ppbv	100
45) t-1,3-Dichloropropene	8.700	75	816302	12.697	ppbv	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091824\
 Data File : VL041601.D
 Acq On : 18 Sep 2024 8:46
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 09/19/2024
 Supervised By : Mahesh Dadoda 09/19/2024

Quant Time: Sep 19 01:28:02 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Sep 17 09:12:21 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	8.137	75	1139357	12.096	ppbv	97
47) 1,1,2-Trichloroethane	8.893	97	973665	10.116	ppbv	99
48) Dibromochloromethane	9.703	129	1597346	10.953	ppbv	99
49) Bromoform	12.400	173	1390368	11.106	ppbv	100
50) 4-Methyl-2-Pentanone	8.172	43	2573409	10.892	ppbv	100
51) 2-Hexanone	9.542	43	2058868	11.520	ppbv	100
52) Tetrachloroethene	10.606	164	868142	10.759	ppbv	98
53) Toluene	9.214	91	2687707	11.362	ppbv	99
54) 1,2-Dibromoethane	10.002	107	1304646	11.007	ppbv	99
56) 1,1,1,2-Tetrachloroethane	11.500	131	1174008	9.667	ppbv	98
57) Chlorobenzene	11.519	112	2106729	9.393	ppbv	99
58) Ethyl Benzene	12.079	91	3644079	10.943	ppbv	98
59) m/p-Xylene	12.365	91	6150817m	20.754	ppbv	
60) o-Xylene	13.050	91	2949502	10.380	ppbv	99
61) Styrene	12.889	104	1446412	12.185	ppbv	100
62) Isopropylbenzene	14.018	105	4295725	9.979	ppbv	100
63) 1,1,2,2-Tetrachloroethane	13.037	83	1399331	9.632	ppbv	99
64) n-propylbenzene	14.908	120	1163729	10.467	ppbv	99
65) tert-Butylbenzene	16.082	119	3881352	9.836	ppbv	99
66) Benzyl Chloride	16.317	91	210631	10.268	ppbv	100
67) sec-Butylbenzene	16.629	105	5423308	9.906	ppbv	100
69) p-Isopropyltoluene	16.985	119	4552021	10.272	ppbv	99
70) n-Butylbenzene	17.876	91	4680037	10.082	ppbv	99
71) 2-Chlorotoluene	14.789	91	3243345	10.240	ppbv	99
72) 4-Ethyltoluene	15.182	105	3543918	10.670	ppbv	100
73) 1,3,5-Trimethylbenzene	15.342	105	3032687	10.340	ppbv	97
74) 1,2,4-Trimethylbenzene	16.098	105	3382442	9.853	ppbv	99
75) 1,3-Dichlorobenzene	16.320	146	2085381	9.693	ppbv	100
76) 1,4-Dichlorobenzene	16.465	146	2059008	9.987	ppbv	99
77) 1,2-Dichlorobenzene	17.133	146	2033227	9.573	ppbv	99
78) Hexachloro-1,3-Butadiene	22.619	225	1515419	8.649	ppbv	100
79) Naphthalene	21.371	128	3186137	12.611	ppbv	99
80) Naphthalene,2-methyl-	24.361	142	2359812	13.329	ppbv	99
81) 1,2,4-Trichlorobenzene	21.127	180	1691433	10.530	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091824\
 Data File : VL041601.D
 Acq On : 18 Sep 2024 8:46
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Sep 19 01:28:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Sep 17 09:12:21 2024
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	102	0.02
2 T	Dichlorodifluoromethane	1.444	1.143	20.8	89	0.02
3	Chlorodifluoromethane	1.213	1.102	9.2	102	0.02
4	Chloromethane	0.667	0.602	9.7	100	0.02
5 T	Vinyl Chloride	0.578	0.600	-3.8	104	0.02
6 T	Bromomethane	0.287	0.265	7.7	99	0.02
7	Chloroethane	0.225	0.213	5.3	102	0.02
8 T	Dichlorotetrafluoroethane	1.515	1.400	7.6	100	0.02
9 T	Propene	0.512	0.520	-1.6	103	0.02
10 T	Heptane	1.198	1.257	-4.9	101	0.02
11 T	Trichlorofluoromethane	1.537	1.394	9.3	99	0.02
12 T	1,1,2-Trichlorotrifluoroeth	1.113	1.032	7.3	101	0.02
13	Ethanol	0.038	0.023#	39.5#	97	0.02
14 T	Bromoethene	0.422	0.425	-0.7	102	0.02
15 T	Acetone	1.255	1.028	18.1	102	0.02
16 T	1,3-Butadiene	0.611	0.573	6.2	101	0.02
17	tert-Butyl alcohol	1.104	1.073	2.8	100	0.02
18 T	1,1-Dichloroethene	0.471	0.451	4.2	97	0.02
19 T	Isopropyl Alcohol	0.643	0.608	5.4	99	0.02
20 T	Methylene Chloride	0.472	0.383	18.9	101	0.02
21 T	Allyl Chloride	0.728	0.730	-0.3	101	0.02
22 T	trans-1,2-Dichloroethene	0.470	0.455	3.2	101	0.02
23 T	Vinyl Acetate	0.691	0.735	-6.4	109	0.02
24 T	1,1-Dichloroethane	1.036	0.985	4.9	101	0.02
25 T	Ethyl Acetate	2.254	2.284	-1.3	102	0.02
26 T	Hexane	0.940	0.930	1.1	102	0.02
27 T	Carbon Disulfide	0.999	1.113	-11.4	102	0.02
28 T	Methyl tert-Butyl Ether	0.579	0.573	1.0	105	0.02
29 T	Chloroform	1.538	1.447	5.9	101	0.02
30 T	Cyclohexane	0.725	0.775	-6.9	101	0.02
31 T	cis-1,2-Dichloroethene	0.892	0.920	-3.1	104	0.02
32 T	1,1,1-Trichloroethane	1.370	1.429	-4.3	101	0.02
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	101	0.02
34 T	2-Butanone	0.477	0.475	0.4	102	0.02
35 T	Carbon Tetrachloride	0.491	0.549	-11.8	102	0.01
36 T	Benzene	0.694	0.726	-4.6	102	0.02
37 T	1,2-Dichloroethane	0.403	0.400	0.7	101	0.02
38 T	Trichloroethene	0.358	0.389	-8.7	103	0.02
39 T	1,2-Dichloropropane	0.283	0.281	0.7	103	0.01
40 T	1,4-Dioxane	0.107	0.117	-9.3	104	0.02
41 T	Tetrahydrofuran	0.262	0.293	-11.8	102	0.02
42 T	Bromodichloromethane	0.561	0.578	-3.0	101	0.01
43	Methyl Methacrylate	0.249	0.286	-14.9	102	0.02
44 T	2,2,4-Trimethylpentane	1.202	1.221	-1.6	101	0.02
45 T	t-1,3-Dichloropropene	0.200	0.254	-27.0	104	0.01
46 T	cis-1,3-Dichloropropene	0.294	0.355	-20.7	101	0.02
47 T	1,1,2-Trichloroethane	0.300	0.303	-1.0	102	0.02
48 T	Dibromochloromethane	0.455	0.498	-9.5	102	0.02

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091824\
 Data File : VL041601.D
 Acq On : 18 Sep 2024 8:46
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Sep 19 01:28:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Sep 17 09:12:21 2024
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.390	0.433	-11.0	101	0.02
50 T	4-Methyl-2-Pentanone	0.736	0.802	-9.0	102	0.02
51 T	2-Hexanone	0.557	0.642	-15.3	102	0.02
52 T	Tetrachloroethene	0.251	0.271	-8.0	102	0.02
53 T	Toluene	0.737	0.838	-13.7	101	0.02
54 T	1,2-Dibromoethane	0.369	0.407	-10.3	100	0.02
55 I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.01
56	1,1,1,2-Tetrachloroethane	0.390	0.377	3.3	101	0.01
57 T	Chlorobenzene	0.721	0.677	6.1	100	0.02
58 T	Ethyl Benzene	1.070	1.171	-9.4	101	0.01
59 T	m/p-Xylene	0.953	0.989	-3.8	101	0.02
60 T	o-Xylene	0.913	0.948	-3.8	100	0.02
61 T	Styrene	0.382	0.465	-21.7	101	0.02
62	Isopropylbenzene	1.384	1.381	0.2	100	0.01
63 T	1,1,2,2-Tetrachloroethane	0.467	0.450	3.6	100	0.01
64	n-propylbenzene	0.357	0.374	-4.8	101	0.01
65	tert-Butylbenzene	1.269	1.248	1.7	101	0.02
66 T	Benzyl Chloride	0.066	0.068	-3.0	102	0.01
67	sec-Butylbenzene	1.760	1.743	1.0	100	0.02
68 S	1-Bromo-4-Fluorobenzene	0.736	0.732	0.5	100	0.02
69	p-Isopropyltoluene	1.425	1.463	-2.7	100	0.01
70	n-Butylbenzene	1.492	1.505	-0.9	100	0.01
71	2-Chlorotoluene	1.018	1.043	-2.5	101	0.01
72 T	4-Ethyltoluene	1.068	1.139	-6.6	100	0.01
73 T	1,3,5-Trimethylbenzene	0.943	0.975	-3.4	100	0.02
74 T	1,2,4-Trimethylbenzene	1.104	1.087	1.5	101	0.01
75 T	1,3-Dichlorobenzene	0.692	0.670	3.2	100	0.01
76 T	1,4-Dichlorobenzene	0.663	0.662	0.2	101	0.01
77 T	1,2-Dichlorobenzene	0.683	0.654	4.2	101	0.01
78 T	Hexachloro-1,3-Butadiene	0.563	0.487	13.5	100	0.00
79 T	Naphthalene	0.812	1.024	-26.1	101	0.01
80 T	Naphthalene,2-methyl-	0.569	0.759	-33.4#	103	0.00
81 T	1,2,4-Trichlorobenzene	0.516	0.544	-5.4	101	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091824\
 Data File : VL041602.D
 Acq On : 18 Sep 2024 9:36
 Operator : SY/MD
 Sample : VL0918ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0918ABL01

Quant Time: Sep 19 01:29:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Sep 17 09:12:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.185	49	1177899	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	6.645	114	3241896	10.000	ppbv	0.02
55) Chlorobenzene-d5	11.452	117	2826139	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	13.761	95	2076380	9.981	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.800%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091824\
Data File : VL041602.D
Acq On : 18 Sep 2024 9:36
Operator : SY/MD
Sample : VL0918ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0918ABL01

Quant Time: Sep 19 01:29:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Tue Sep 17 09:12:21 2024
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091824\
 Data File : VL041603.D
 Acq On : 18 Sep 2024 10:16
 Operator : SY/MD
 Sample : VL0918ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0918ABS01

Quant Time: Sep 19 01:29:22 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Sep 17 09:12:21 2024

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 09/19/2024
 Supervised By : Mahesh Dadoda 09/19/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.189	49	1156371	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	6.648	114	3163270	10.000	ppbv	0.02
55) Chlorobenzene-d5	11.462	117	3156402	10.000	ppbv	0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 13.770 95 2259237 9.723 ppbv 0.02
 Spiked Amount 10.000 Range 65 - 135 Recovery = 97.200%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.732	85	1884134	11.282	ppbv	99
3) Chlorodifluoromethane	2.677	51	1281112	9.134	ppbv	99
4) Chloromethane	2.816	50	707023	9.165	ppbv	98
5) Vinyl Chloride	2.922	62	686540	10.278	ppbv	99
6) Bromomethane	3.121	94	316275	9.517	ppbv	99
7) Chloroethane	3.198	64	236020	9.085	ppbv	98
8) Dichlorotetrafluoroethane	2.858	85	1659459	9.473	ppbv	99
9) Propene	2.697	41	588244	9.943	ppbv	99
10) Heptane	7.574	43	1457487	10.522	ppbv	99
11) Trichlorofluoromethane	3.562	101	1615013	9.088	ppbv	96
12) 1,1,2-Trichlorotrifluo...	4.063	101	1190198	9.248	ppbv	99
13) Ethanol	3.250	45	26862m	6.039	ppbv	
14) Bromoethene	3.366	108	487861	9.997	ppbv	98
15) Acetone	3.468	43	1192968	8.217	ppbv	99
16) 1,3-Butadiene	2.986	39	693352	9.819	ppbv	96
17) tert-Butyl alcohol	3.877	59	1224256	9.589	ppbv	99
18) 1,1-Dichloroethene	3.880	96	535313	9.825	ppbv	99
19) Isopropyl Alcohol	3.575	45	656028	8.827	ppbv #	98
20) Methylene Chloride	3.931	84	441509	8.095	ppbv	98
21) Allyl Chloride	3.993	41	813252	9.665	ppbv	98
22) trans-1,2-Dichloroethene	4.443	96	522406	9.611	ppbv	99
23) Vinyl Acetate	4.623	43	835177	10.459	ppbv	98
24) 1,1-Dichloroethane	4.562	63	1142671	9.534	ppbv	99
25) Ethyl Acetate	5.201	43	2647300	10.158	ppbv	99
26) Hexane	5.208	57	1074230	9.884	ppbv	100
27) Carbon Disulfide	4.121	76	1270273	10.995	ppbv	100
28) Methyl tert-Butyl Ether	4.584	73	657790	9.816	ppbv	99
29) Chloroform	5.272	83	1685499	9.477	ppbv	98
30) Cyclohexane	6.597	84	909681	10.858	ppbv	100
31) cis-1,2-Dichloroethene	5.076	61	1045127	10.129	ppbv	98
32) 1,1,1-Trichloroethane	5.999	97	1637346	10.336	ppbv	98
34) 2-Butanone	4.784	43	1503091	9.965	ppbv	99
35) Carbon Tetrachloride	6.491	117	1732518	11.163	ppbv	98
36) Benzene	6.365	78	2307599	10.504	ppbv	99
37) 1,2-Dichloroethane	5.803	62	1290133	10.122	ppbv	100
38) Trichloroethene	7.288	130	1229916	10.871	ppbv	98
39) 1,2-Dichloropropane	7.070	63	887630	9.932	ppbv	99
40) 1,4-Dioxane	7.282	88	364111	10.808	ppbv #	98
41) Tetrahydrofuran	5.552	42	920914	11.094	ppbv	99
42) Bromodichloromethane	7.246	83	1840578	10.379	ppbv	97
43) Methyl Methacrylate	7.468	69	905817	11.522	ppbv	99
44) 2,2,4-Trimethylpentane	7.320	57	3875181	10.196	ppbv	100
45) t-1,3-Dichloropropene	8.703	75	809504	12.771	ppbv	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091824\
 Data File : VL041603.D
 Acq On : 18 Sep 2024 10:16
 Operator : SY/MD
 Sample : VL0918ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0918ABS01

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/19/2024
 Supervised By :Mahesh Dadoda 09/19/2024

Quant Time: Sep 19 01:29:22 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Tue Sep 17 09:12:21 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	8.140	75	1131086	12.181	ppbv	99
47) 1,1,2-Trichloroethane	8.893	97	963708	10.156	ppbv	98
48) Dibromochloromethane	9.706	129	1586368	11.034	ppbv	99
49) Bromoform	12.404	173	1395106	11.304	ppbv	100
50) 4-Methyl-2-Pentanone	8.176	43	2558165	10.982	ppbv	100
51) 2-Hexanone	9.542	43	1996601	11.332	ppbv #	99
52) Tetrachloroethene	10.610	164	870327	10.941	ppbv	98
53) Toluene	9.214	91	2660302	11.408	ppbv	100
54) 1,2-Dibromoethane	10.002	107	1300569	11.130	ppbv	98
56) 1,1,1,2-Tetrachloroethane	11.500	131	1174089	9.527	ppbv	98
57) Chlorobenzene	11.523	112	2083591	9.156	ppbv	99
58) Ethyl Benzene	12.082	91	3610725	10.686	ppbv	98
59) m/p-Xylene	12.365	91	6126839m	20.373	ppbv	
60) o-Xylene	13.053	91	2952785	10.241	ppbv	100
61) Styrene	12.889	104	1434081	11.906	ppbv	100
62) Isopropylbenzene	14.024	105	4303632	9.852	ppbv	100
63) 1,1,2,2-Tetrachloroethane	13.044	83	1395823	9.469	ppbv	99
64) n-propylbenzene	14.912	120	1163609	10.315	ppbv	98
65) tert-Butylbenzene	16.085	119	3844766	9.602	ppbv	100
66) Benzyl Chloride	16.320	91	212184	10.193	ppbv	99
67) sec-Butylbenzene	16.632	105	5382674	9.689	ppbv	100
69) p-Isopropyltoluene	16.989	119	4506080	10.021	ppbv	100
70) n-Butylbenzene	17.879	91	4641102	9.853	ppbv	100
71) 2-Chlorotoluene	14.796	91	3239678	10.080	ppbv	100
72) 4-Ethyltoluene	15.188	105	3572445	10.600	ppbv	99
73) 1,3,5-Trimethylbenzene	15.346	105	3028398	10.176	ppbv	97
74) 1,2,4-Trimethylbenzene	16.101	105	3355900	9.634	ppbv	99
75) 1,3-Dichlorobenzene	16.323	146	2087035	9.560	ppbv	100
76) 1,4-Dichlorobenzene	16.468	146	2006487	9.592	ppbv	99
77) 1,2-Dichlorobenzene	17.137	146	2006971	9.312	ppbv	100
78) Hexachloro-1,3-Butadiene	22.625	225	1497828	8.425	ppbv	99
79) Naphthalene	21.374	128	3150966	12.291	ppbv	98
80) Naphthalene,2-methyl-	24.368	142	2210947	12.308	ppbv	99
81) 1,2,4-Trichlorobenzene	21.133	180	1664970	10.215	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091824\
 Data File : VL041615.D
 Acq On : 18 Sep 2024 20:38
 Operator : SY/MD
 Sample : QC090524 CAN10153
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC090524 CAN10153

Quant Time: Sep 19 03:40:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Tue Sep 17 09:12:21 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/19/2024
 Supervised By :Mahesh Dadoda 09/19/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.198	49	1000091	10.000	ppbv	0.03
33) 1,4-Difluorobenzene	6.655	114	2767195	10.000	ppbv	0.03
55) Chlorobenzene-d5	11.465	117	2435639	10.000	ppbv	0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	13.773	95	1757273	9.801	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091824\
Data File : VL041615.D
Acq On : 18 Sep 2024 20:38
Operator : SY/MD
Sample : QC090524 CAN10153
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC090524 CAN10153

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 09/19/2024
Supervised By :Mahesh Dadoda 09/19/2024

Quant Time: Sep 19 03:40:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Tue Sep 17 09:12:21 2024
Response via : Initial Calibration





Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL101424\
Data File : VL041659.D
Acq On : 14 Oct 2024 8:23
Operator : SY/MD
Sample : BFB
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
Last Update : Tue Sep 17 09:12:21 2024



AutoFind: Scans 3481, 3482, 3483; Background Corrected with Scan 3456

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	8	40	24.0	194560	PASS
75	95	30	66	56.3	456469	PASS
95	95	100	100	100.0	811243	PASS
96	95	5	9	7.0	56400	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	68.4	554837	PASS
175	174	4	9	7.1	39429	PASS
176	174	93	101	94.4	523797	PASS
177	176	5	9	6.7	35344	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL101424\
 Data File : VL041660.D
 Acq On : 14 Oct 2024 9:43
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/15/2024
 Supervised By :Mahesh Dadoda 10/15/2024

Quant Time: Oct 15 02:38:21 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Tue Sep 17 09:12:21 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.201	49	1139050	10.000	ppbv	0.03
33) 1,4-Difluorobenzene	6.671	114	3098402	10.000	ppbv	0.04
55) Chlorobenzene-d5	11.491	117	3033567	10.000	ppbv	0.05

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 13.799 95 2208698 9.891 ppbv 0.05
 Spiked Amount 10.000 Range 65 - 135 Recovery = 98.900%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.726	85	1287579	7.827	ppbv	99
3) Chlorodifluoromethane	2.668	51	1359492	9.840	ppbv	99
4) Chloromethane	2.809	50	754633	9.931	ppbv	96
5) Vinyl Chloride	2.915	62	733569	11.149	ppbv	100
6) Bromomethane	3.115	94	281384	8.596	ppbv	98
7) Chloroethane	3.195	64	259953	10.159	ppbv	98
8) Dichlorotetrafluoroethane	2.851	85	1730383	10.028	ppbv	99
9) Propene	2.687	41	653601	11.216	ppbv	100
10) Heptane	7.597	43	1554594	11.394	ppbv	100
11) Trichlorofluoromethane	3.562	101	1689077	9.649	ppbv	99
12) 1,1,2-Trichlorotrifluo...	4.070	101	1265757	9.984	ppbv	99
13) Ethanol	3.243	45	25933m	5.919	ppbv	
14) Bromoethene	3.365	108	516546	10.745	ppbv	99
15) Acetone	3.468	43	1299921	9.090	ppbv	100
16) 1,3-Butadiene	2.983	39	734455	10.559	ppbv	99
17) tert-Butyl alcohol	3.880	59	1332556	10.596	ppbv	99
18) 1,1-Dichloroethene	3.883	96	567989	10.583	ppbv	98
19) Isopropyl Alcohol	3.578	45	754520	10.307	ppbv	99
20) Methylene Chloride	3.935	84	454891	8.467	ppbv	96
21) Allyl Chloride	3.999	41	908373	10.960	ppbv	99
22) trans-1,2-Dichloroethene	4.452	96	565315	10.559	ppbv	100
23) Vinyl Acetate	4.635	43	966869	12.292	ppbv #	98
24) 1,1-Dichloroethane	4.571	63	1246647	10.559	ppbv	99
25) Ethyl Acetate	5.214	43	2782394	10.838	ppbv	100
26) Hexane	5.221	57	1163508	10.868	ppbv	97
27) Carbon Disulfide	4.127	76	1356524	11.920	ppbv	99
28) Methyl tert-Butyl Ether	4.594	73	721141	10.925	ppbv	100
29) Chloroform	5.288	83	1749654	9.987	ppbv	97
30) Cyclohexane	6.616	84	959853	11.631	ppbv	99
31) cis-1,2-Dichloroethene	5.089	61	1094693	10.771	ppbv	98
32) 1,1,1-Trichloroethane	6.015	97	1715729	10.995	ppbv	99
34) 2-Butanone	4.796	43	1483922	10.044	ppbv	99
35) Carbon Tetrachloride	6.510	117	1753896	11.537	ppbv	98
36) Benzene	6.385	78	2426961	11.279	ppbv	99
37) 1,2-Dichloroethane	5.822	62	1320644	10.578	ppbv	99
38) Trichloroethene	7.311	130	1180521	10.653	ppbv	98
39) 1,2-Dichloropropane	7.092	63	948760	10.838	ppbv	100
40) 1,4-Dioxane	7.304	88	376944	11.423	ppbv #	100
41) Tetrahydrofuran	5.568	42	1015739	12.493	ppbv	98
42) Bromodichloromethane	7.269	83	1881537	10.832	ppbv	99
43) Methyl Methacrylate	7.491	69	943145	12.248	ppbv	100
44) 2,2,4-Trimethylpentane	7.346	57	4090203	10.987	ppbv	100
45) t-1,3-Dichloropropene	8.729	75	720472	11.605	ppbv	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL101424\
 Data File : VL041660.D
 Acq On : 14 Oct 2024 9:43
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 10/15/2024
 Supervised By : Mahesh Dadoda 10/15/2024

Quant Time: Oct 15 02:38:21 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Sep 17 09:12:21 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	8.163	75	1040833	11.444	ppbv	99
47) 1,1,2-Trichloroethane	8.918	97	977523	10.517	ppbv	97
48) Dibromochloromethane	9.732	129	1583619	11.245	ppbv	99
49) Bromoform	12.429	173	1364123	11.284	ppbv	99
50) 4-Methyl-2-Pentanone	8.198	43	2676056	11.729	ppbv	100
51) 2-Hexanone	9.568	43	2120806	12.289	ppbv	99
52) Tetrachloroethene	10.635	164	876267	11.246	ppbv	98
53) Toluene	9.243	91	2746651	12.025	ppbv	100
54) 1,2-Dibromoethane	10.031	107	1315299	11.491	ppbv	97
56) 1,1,1,2-Tetrachloroethane	11.532	131	1179827	9.961	ppbv	98
57) Chlorobenzene	11.552	112	2095164	9.579	ppbv	99
58) Ethyl Benzene	12.111	91	3656350	11.259	ppbv	99
59) m/p-Xylene	12.394	91	6168384m	21.342	ppbv	
60) o-Xylene	13.085	91	2959365	10.680	ppbv	99
61) Styrene	12.921	104	1423375	12.295	ppbv	100
62) Isopropylbenzene	14.053	105	4313712	10.275	ppbv	100
63) 1,1,2,2-Tetrachloroethane	13.069	83	1527720	10.783	ppbv	99
64) n-propylbenzene	14.941	120	1151763	10.623	ppbv	99
65) tert-Butylbenzene	16.114	119	3846358	9.995	ppbv	100
66) Benzyl Chloride	16.349	91	202012	10.098	ppbv	99
67) sec-Butylbenzene	16.661	105	5377834	10.073	ppbv	100
69) p-Isopropyltoluene	17.021	119	4463259	10.328	ppbv	100
70) n-Butylbenzene	17.911	91	4587597	10.134	ppbv	100
71) 2-Chlorotoluene	14.828	91	3227785	10.450	ppbv	99
72) 4-Ethyltoluene	15.217	105	3522903	10.876	ppbv	99
73) 1,3,5-Trimethylbenzene	15.378	105	3015731	10.544	ppbv	96
74) 1,2,4-Trimethylbenzene	16.133	105	3335754	9.964	ppbv	97
75) 1,3-Dichlorobenzene	16.355	146	2005347	9.558	ppbv	100
76) 1,4-Dichlorobenzene	16.497	146	1924339	9.571	ppbv	100
77) 1,2-Dichlorobenzene	17.166	146	1941181	9.372	ppbv	99
78) Hexachloro-1,3-Butadiene	22.660	225	1360072	7.960	ppbv	99
79) Naphthalene	21.410	128	2595328	10.533	ppbv	99
80) Naphthalene,2-methyl-	24.394	142	895271	5.185	ppbv	99
81) 1,2,4-Trichlorobenzene	21.165	180	1390542	8.877	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL101424\
 Data File : VL041660.D
 Acq On : 14 Oct 2024 9:43
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Quant Time: Oct 15 02:38:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
 QLast Update : Tue Sep 17 09:12:21 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 10/15/2024
 Supervised By : Mahesh Dadoda 10/15/2024



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL101424\
 Data File : VL041660.D
 Acq On : 14 Oct 2024 9:43
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Oct 15 02:38:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Sep 17 09:12:21 2024
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	99	0.03
2 T	Dichlorodifluoromethane	1.444	1.130	21.7	85	0.02
3	Chlorodifluoromethane	1.213	1.194	1.6	107	0.02
4	Chloromethane	0.667	0.663	0.6	107	0.02
5 T	Vinyl Chloride	0.578	0.644	-11.4	108	0.02
6 T	Bromomethane	0.287	0.247	13.9	89	0.02
7	Chloroethane	0.225	0.228	-1.3	106	0.02
8 T	Dichlorotetrafluoroethane	1.515	1.519	-0.3	105	0.02
9 T	Propene	0.512	0.574	-12.1	110	0.02
10 T	Heptane	1.198	1.365	-13.9	106	0.04
11 T	Trichlorofluoromethane	1.537	1.483	3.5	102	0.02
12 T	1,1,2-Trichlorotrifluoroeth	1.113	1.111	0.2	105	0.03
13	Ethanol	0.038	0.023#	39.5#	95	0.02
14 T	Bromoethene	0.422	0.453	-7.3	106	0.02
15 T	Acetone	1.255	1.141	9.1	110	0.03
16 T	1,3-Butadiene	0.611	0.645	-5.6	110	0.02
17	tert-Butyl alcohol	1.104	1.170	-6.0	106	0.03
18 T	1,1-Dichloroethene	0.471	0.499	-5.9	103	0.03
19 T	Isopropyl Alcohol	0.643	0.662	-3.0	105	0.03
20 T	Methylene Chloride	0.472	0.399	15.5	102	0.03
21 T	Allyl Chloride	0.728	0.797	-9.5	107	0.03
22 T	trans-1,2-Dichloroethene	0.470	0.496	-5.5	106	0.03
23 T	Vinyl Acetate	0.691	0.849	-22.9	121	0.03
24 T	1,1-Dichloroethane	1.036	1.094	-5.6	109	0.03
25 T	Ethyl Acetate	2.254	2.443	-8.4	106	0.04
26 T	Hexane	0.940	1.021	-8.6	109	0.03
27 T	Carbon Disulfide	0.999	1.191	-19.2	106	0.03
28 T	Methyl tert-Butyl Ether	0.579	0.633	-9.3	112	0.03
29 T	Chloroform	1.538	1.536	0.1	104	0.04
30 T	Cyclohexane	0.725	0.843	-16.3	106	0.04
31 T	cis-1,2-Dichloroethene	0.892	0.961	-7.7	105	0.04
32 T	1,1,1-Trichloroethane	1.370	1.506	-9.9	103	0.04
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	97	0.04
34 T	2-Butanone	0.477	0.479	-0.4	99	0.04
35 T	Carbon Tetrachloride	0.491	0.566	-15.3	101	0.04
36 T	Benzene	0.694	0.783	-12.8	107	0.04
37 T	1,2-Dichloroethane	0.403	0.426	-5.7	104	0.04
38 T	Trichloroethene	0.358	0.381	-6.4	98	0.05
39 T	1,2-Dichloropropane	0.283	0.306	-8.1	108	0.04
40 T	1,4-Dioxane	0.107	0.122	-14.0	104	0.04
41 T	Tetrahydrofuran	0.262	0.328	-25.2	110	0.04
42 T	Bromodichloromethane	0.561	0.607	-8.2	103	0.04
43	Methyl Methacrylate	0.249	0.304	-22.1	105	0.04
44 T	2,2,4-Trimethylpentane	1.202	1.320	-9.8	106	0.04
45 T	t-1,3-Dichloropropene	0.200	0.233	-16.5	92	0.04
46 T	cis-1,3-Dichloropropene	0.294	0.336	-14.3	92	0.04
47 T	1,1,2-Trichloroethane	0.300	0.315	-5.0	103	0.05
48 T	Dibromochloromethane	0.455	0.511	-12.3	101	0.05

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL101424\
 Data File : VL041660.D
 Acq On : 14 Oct 2024 9:43
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Oct 15 02:38:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Sep 17 09:12:21 2024
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.390	0.440	-12.8	99	0.05
50 T	4-Methyl-2-Pentanone	0.736	0.864	-17.4	106	0.04
51 T	2-Hexanone	0.557	0.684	-22.8	105	0.04
52 T	Tetrachloroethene	0.251	0.283	-12.7	103	0.05
53 T	Toluene	0.737	0.886	-20.2	103	0.05
54 T	1,2-Dibromoethane	0.369	0.425	-15.2	101	0.05
55 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.05
56	1,1,1,2-Tetrachloroethane	0.390	0.389	0.3	102	0.05
57 T	Chlorobenzene	0.721	0.691	4.2	100	0.05
58 T	Ethyl Benzene	1.070	1.205	-12.6	102	0.05
59 T	m/p-Xylene	0.953	1.017	-6.7	101	0.05
60 T	o-Xylene	0.913	0.976	-6.9	100	0.05
61 T	Styrene	0.382	0.469	-22.8	99	0.05
62	Isopropylbenzene	1.384	1.422	-2.7	101	0.05
63 T	1,1,2,2-Tetrachloroethane	0.467	0.504	-7.9	109	0.05
64	n-propylbenzene	0.357	0.380	-6.4	100	0.05
65	tert-Butylbenzene	1.269	1.268	0.1	100	0.05
66 T	Benzyl Chloride	0.066	0.067	-1.5	98	0.05
67	sec-Butylbenzene	1.760	1.773	-0.7	99	0.05
68 S	1-Bromo-4-Fluorobenzene	0.736	0.728	1.1	97	0.05
69	p-Isopropyltoluene	1.425	1.471	-3.2	98	0.05
70	n-Butylbenzene	1.492	1.512	-1.3	98	0.05
71	2-Chlorotoluene	1.018	1.064	-4.5	100	0.05
72 T	4-Ethyltoluene	1.068	1.161	-8.7	99	0.05
73 T	1,3,5-Trimethylbenzene	0.943	0.994	-5.4	99	0.05
74 T	1,2,4-Trimethylbenzene	1.104	1.100	0.4	99	0.05
75 T	1,3-Dichlorobenzene	0.692	0.661	4.5	96	0.05
76 T	1,4-Dichlorobenzene	0.663	0.634	4.4	95	0.05
77 T	1,2-Dichlorobenzene	0.683	0.640	6.3	96	0.05
78 T	Hexachloro-1,3-Butadiene	0.563	0.448	20.4	90	0.05
79 T	Naphthalene	0.812	0.856	-5.4	82	0.05
80 T	Naphthalene,2-methyl-	0.569	0.295	48.2#	39	0.04
81 T	1,2,4-Trichlorobenzene	0.516	0.458	11.2	83	0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL101424\
 Data File : VL041663.D
 Acq On : 14 Oct 2024 11:41
 Operator : SY/MD
 Sample : VIBLK
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Quant Time: Oct 15 02:40:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Sep 17 09:12:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.208	49	1006897	10.000	ppbv	0.04
33) 1,4-Difluorobenzene	6.671	114	2688648	10.000	ppbv	0.04
55) Chlorobenzene-d5	11.491	117	2332708	10.000	ppbv	0.05

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.802	95	1624426	9.460	ppbv	0.05
Spiked Amount	10.000	Range	65 - 135	Recovery	=	94.600%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL101424\
Data File : VL041663.D
Acq On : 14 Oct 2024 11:41
Operator : SY/MD
Sample : VIBLK
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: Oct 15 02:40:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Tue Sep 17 09:12:21 2024
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL101424\
 Data File : VL041662.D
 Acq On : 14 Oct 2024 11:02
 Operator : SY/MD
 Sample : VL1014ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1014ABS01

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/15/2024
 Supervised By :Mahesh Dadoda 10/15/2024

Quant Time: Oct 15 02:39:33 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Tue Sep 17 09:12:21 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.205	49	1108335	10.000	ppbv	0.04
33) 1,4-Difluorobenzene	6.668	114	2963163	10.000	ppbv	0.04
55) Chlorobenzene-d5	11.491	117	2950594m	10.000	ppbv	0.05

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 13.799 95 2157989 9.935 ppbv 0.05
 Spiked Amount 10.000 Range 65 - 135 Recovery = 99.400%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.722	85	1145692	7.158	ppbv	100
3) Chlorodifluoromethane	2.668	51	1310086	9.745	ppbv	99
4) Chloromethane	2.809	50	719788	9.735	ppbv	99
5) Vinyl Chloride	2.915	62	702662	10.975	ppbv	99
6) Bromomethane	3.118	94	255661	8.027	ppbv	98
7) Chloroethane	3.195	64	243123	9.764	ppbv	100
8) Dichlorotetrafluoroethane	2.851	85	1660268	9.888	ppbv	100
9) Propene	2.687	41	601524	10.608	ppbv	100
10) Heptane	7.597	43	1497343	11.278	ppbv	100
11) Trichlorofluoromethane	3.562	101	1638422	9.620	ppbv	98
12) 1,1,2-Trichlorotrifluo...	4.070	101	1209387	9.804	ppbv	99
13) Ethanol	3.250	45	25964m	6.091	ppbv	
14) Bromoethene	3.362	108	490120	10.478	ppbv	99
15) Acetone	3.468	43	1216859	8.745	ppbv	99
16) 1,3-Butadiene	2.980	39	697703	10.309	ppbv	97
17) tert-Butyl alcohol	3.880	59	1247539	10.195	ppbv	99
18) 1,1-Dichloroethene	3.883	96	536176	10.267	ppbv	98
19) Isopropyl Alcohol	3.578	45	669012	9.392	ppbv #	98
20) Methylene Chloride	3.935	84	432714	8.278	ppbv	93
21) Allyl Chloride	3.999	41	835610	10.361	ppbv	99
22) trans-1,2-Dichloroethene	4.449	96	529705	10.168	ppbv	99
23) Vinyl Acetate	4.632	43	878250	11.475	ppbv	99
24) 1,1-Dichloroethane	4.571	63	1187574	10.338	ppbv	100
25) Ethyl Acetate	5.214	43	2658560	10.643	ppbv	99
26) Hexane	5.221	57	1083358	10.400	ppbv	99
27) Carbon Disulfide	4.128	76	1283235	11.588	ppbv	99
28) Methyl tert-Butyl Ether	4.594	73	665747	10.366	ppbv	99
29) Chloroform	5.285	83	1702002	9.984	ppbv	97
30) Cyclohexane	6.619	84	915770	11.404	ppbv	99
31) cis-1,2-Dichloroethene	5.089	61	1035844	10.474	ppbv	99
32) 1,1,1-Trichloroethane	6.015	97	1684176	11.092	ppbv	99
34) 2-Butanone	4.793	43	1385663	9.807	ppbv	99
35) Carbon Tetrachloride	6.507	117	1746265	12.011	ppbv	100
36) Benzene	6.385	78	2302124	11.187	ppbv	99
37) 1,2-Dichloroethane	5.822	62	1275743	10.685	ppbv	99
38) Trichloroethene	7.307	130	1161316	10.958	ppbv	99
39) 1,2-Dichloropropane	7.089	63	896567	10.709	ppbv	98
40) 1,4-Dioxane	7.307	88	351904	11.151	ppbv #	99
41) Tetrahydrofuran	5.568	42	919479	11.825	ppbv	100
42) Bromodichloromethane	7.266	83	1834480	11.043	ppbv	99
43) Methyl Methacrylate	7.494	69	902852	12.260	ppbv	99
44) 2,2,4-Trimethylpentane	7.343	57	3928081	11.033	ppbv	100
45) t-1,3-Dichloropropene	8.729	75	686411	11.561	ppbv	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL101424\
 Data File : VL041662.D
 Acq On : 14 Oct 2024 11:02
 Operator : SY/MD
 Sample : VL1014ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1014ABS01

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 10/15/2024
 Supervised By : Mahesh Dadoda 10/15/2024

Quant Time: Oct 15 02:39:33 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Sep 17 09:12:21 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	8.163	75	986486	11.341	ppbv	99
47) 1,1,2-Trichloroethane	8.918	97	953812	10.730	ppbv	99
48) Dibromochloromethane	9.732	129	1556080	11.554	ppbv	100
49) Bromoform	12.433	173	1337606	11.570	ppbv	99
50) 4-Methyl-2-Pentanone	8.198	43	2599545	11.914	ppbv	99
51) 2-Hexanone	9.571	43	2038924	12.354	ppbv	100
52) Tetrachloroethene	10.639	164	845842	11.351	ppbv	97
53) Toluene	9.240	91	2657525	12.166	ppbv	100
54) 1,2-Dibromoethane	10.031	107	1251918	11.437	ppbv	98
56) 1,1,1,2-Tetrachloroethane	11.529	131	1161496	10.082	ppbv	99
57) Chlorobenzene	11.552	112	2072182	9.741	ppbv	98
58) Ethyl Benzene	12.111	91	3564325	11.285	ppbv	99
59) m/p-Xylene	12.394	91	6050276m	21.522	ppbv	
60) o-Xylene	13.079	91	2940555	10.910	ppbv	100
61) Styrene	12.918	104	1389742	12.342	ppbv	100
62) Isopropylbenzene	14.053	105	4251841	10.413	ppbv	100
63) 1,1,2,2-Tetrachloroethane	13.069	83	1519552	11.027	ppbv	99
64) n-propylbenzene	14.937	120	1134283	10.756	ppbv	99
65) tert-Butylbenzene	16.114	119	3818025	10.201	ppbv	100
66) Benzyl Chloride	16.349	91	206302	10.602	ppbv	95
67) sec-Butylbenzene	16.661	105	5353060	10.308	ppbv	100
69) p-Isopropyltoluene	17.018	119	4451290	10.590	ppbv	100
70) n-Butylbenzene	17.908	91	4571641	10.382	ppbv	100
71) 2-Chlorotoluene	14.825	91	3185904	10.604	ppbv	99
72) 4-Ethyltoluene	15.214	105	3484380	11.060	ppbv	99
73) 1,3,5-Trimethylbenzene	15.375	105	2987131	10.737	ppbv	98
74) 1,2,4-Trimethylbenzene	16.134	105	3325802	10.214	ppbv	95
75) 1,3-Dichlorobenzene	16.355	146	1996430	9.783	ppbv	100
76) 1,4-Dichlorobenzene	16.497	146	1937729	9.909	ppbv	100
77) 1,2-Dichlorobenzene	17.169	146	1943634	9.647	ppbv	99
78) Hexachloro-1,3-Butadiene	22.664	225	1356495	8.162	ppbv	100
79) Naphthalene	21.413	128	2594745	10.827	ppbv	99
80) Naphthalene,2-methyl-	24.394	142	833531	4.964	ppbv	99
81) 1,2,4-Trichlorobenzene	21.165	180	1375748	9.030	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL101424\
 Data File : VL041664.D
 Acq On : 14 Oct 2024 12:20
 Operator : SY/MD
 Sample : CAN 10258
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 CAN 10258

Quant Time: Oct 15 02:40:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Sep 17 09:12:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.201	49	1000184	10.000	ppbv	0.03
33) 1,4-Difluorobenzene	6.667	114	2684060	10.000	ppbv	0.04
55) Chlorobenzene-d5	11.487	117	2368428	10.000	ppbv	0.04

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	13.799	95	1743192	9.998	ppbv	0.05
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.000%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL101424\
Data File : VL041664.D
Acq On : 14 Oct 2024 12:20
Operator : SY/MD
Sample : CAN 10258
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10258

Quant Time: Oct 15 02:40:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Tue Sep 17 09:12:21 2024
Response via : Initial Calibration



Instrument ID: **MSVOA_L**

Daily Analysis Runlog For Sequence/QC Batch ID # VL101424

Review By	Semsettin Yesilyurt	Review On	10/15/2024 1:33:56 PM		
Supervise By	Mahesh Dadoda	Supervise On	10/15/2024 6:29:46 PM		
SubDirectory	VL101424	HP Acquire Method	AIR-TO15	HP Processing Method	VL091624AIR.M
STD. NAME		STD REF.#			
Tune/Reschk		AP2543			
Initial Calibration Stds		AP2546,AP2547,AP2548			
CCC		AP2546			
Internal Standard/PEM		AP2543			
ICV/I.BLK		AP2544			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VL041659.D	14 Oct 2024 8:23	SY/MD	Ok
2	VSTDCCC010	VL041660.D	14 Oct 2024 9:43	SY/MD	Ok,M
3	VL1014ABL01	VL041661.D	14 Oct 2024 10:23	SY/MD	Ok
4	VL1014ABS01	VL041662.D	14 Oct 2024 11:02	SY/MD	Ok,M
5	VIBLK	VL041663.D	14 Oct 2024 11:41	SY/MD	Ok
6	CAN 10258	VL041664.D	14 Oct 2024 12:20	SY/MD	Ok
7	CAN 10590	VL041665.D	14 Oct 2024 13:00	SY/MD	Ok,M
8	CAN 10334	VL041666.D	14 Oct 2024 13:40	SY/MD	Ok,M
9	P4368-15 0.45	VL041667.D	14 Oct 2024 14:20	SY/MD	Ok,M
10	P4368-15 0.40	VL041668.D	14 Oct 2024 15:00	SY/MD	Ok,M
11	P4368-15 0.1	VL041669.D	14 Oct 2024 15:40	SY/MD	Ok,M
12	P4368-15 0.03	VL041670.D	14 Oct 2024 16:20	SY/MD	Ok,M
13	VSTDCCC010	VL041671.D	14 Oct 2024 17:00	SY/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_L**

Daily Analysis Runlog For Sequence/QC Batch ID # VL072524

Review By	Semsettin Yesilyurt	Review On	7/29/2024 10:58:46 AM
Supervise By	Mahesh Dadoda	Supervise On	7/29/2024 5:57:16 PM
SubDirectory	VL072524	HP Acquire Method	AIR-TO15
		HP Processing Method	VL072524AIR.M
STD. NAME	STD REF.#		
Tune/Reschk	AP2535		
Initial Calibration Stds	AP2539,AP2540,AP2541		
CCC	AP2539		
Internal Standard/PEM	AP2535		
ICV/I.BLK	AP2537		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VL041397.D	25 Jul 2024 12:29	SY/MD	Ok
2	VSTDICC0.5	VL041398.D	25 Jul 2024 13:19	SY/MD	Ok,M
3	VSTDICCC010	VL041399.D	25 Jul 2024 13:59	SY/MD	Ok,M
4	VSTDICC002	VL041400.D	25 Jul 2024 14:40	SY/MD	Ok,M
5	VSTDICC001	VL041401.D	25 Jul 2024 15:19	SY/MD	Ok,M
6	VSTDICC0.5	VL041402.D	25 Jul 2024 16:00	SY/MD	Not Ok
7	VSTDICC0.1	VL041403.D	25 Jul 2024 16:42	SY/MD	Ok,M
8	VSTDICC0.03	VL041404.D	25 Jul 2024 17:24	SY/MD	Ok,M
9	VSTDICC015	VL041405.D	25 Jul 2024 18:06	SY/MD	Ok,M
10	VSTDICV010	VL041406.D	25 Jul 2024 19:31	SY/MD	Ok,M
11	VL0725ABL01	VL041407.D	25 Jul 2024 20:56	SY/MD	Ok
12	VL0725ABS01	VL041408.D	25 Jul 2024 21:38	SY/MD	Ok,M
13	VIBLK	VL041409.D	25 Jul 2024 22:20	SY/MD	Ok
14	P3320-01	VL041410.D	26 Jul 2024 7:22	SY/MD	Ok,M
15	P3320-01DUP	VL041411.D	26 Jul 2024 8:02	SY/MD	Ok,M
16	P3320-02	VL041412.D	26 Jul 2024 8:43	SY/MD	Dilution
17	P3342-04	VL041413.D	26 Jul 2024 9:24	SY/MD	Ok,M
18	P3356-01	VL041414.D	26 Jul 2024 10:04	SY/MD	Dilution
19	P3356-02	VL041415.D	26 Jul 2024 10:45	SY/MD	Ok,M
20	P3320-02DL	VL041416.D	26 Jul 2024 11:25	SY/MD	Ok,M
21	P3356-01DL	VL041417.D	26 Jul 2024 12:05	SY/MD	Ok,M

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL072524

Review By	Semsettin Yesilyurt	Review On	7/29/2024 10:58:46 AM		
Supervise By	Mahesh Dadoda	Supervise On	7/29/2024 5:57:16 PM		
SubDirectory	VL072524	HP Acquire Method	AIR-TO15	HP Processing Method	VL072524AIR.M
STD. NAME	STD REF.#				
Tune/Reschk	AP2535				
Initial Calibration Stds	AP2539,AP2540,AP2541				
CCC	AP2539				
Internal Standard/PEM	AP2535				
ICV/I.BLK	AP2537				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

22	VSTDCCC010	VL041418.D	26 Jul 2024 12:45	SY/MD	Not Ok
----	------------	------------	-------------------	-------	--------

M : Manual Integration

Instrument ID: **MSVOA_L**

Daily Analysis Runlog For Sequence/QC Batch ID # VL081924

Review By	Semsettin Yesilyurt	Review On	8/20/2024 2:35:55 PM
Supervise By	Maresh Dadoda	Supervise On	8/20/2024 4:35:44 PM
SubDirectory	VL081924	HP Acquire Method	AIR-TO15
		HP Processing Method	VL072524AIR.M
STD. NAME	STD REF.#		
Tune/Reschk	AP2536		
Initial Calibration Stds	AP2539,AP2540,AP2541		
CCC	AP2539		
Internal Standard/PEM	AP2536		
ICV/I.BLK	AP2537		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VL041539.D	19 Aug 2024 8:10	SY/MD	Ok
2	VSTDCCC010	VL041540.D	19 Aug 2024 8:59	SY/MD	Ok,M
3	VL0819ABL01	VL041541.D	19 Aug 2024 10:07	SY/MD	Ok
4	VL0819ABS01	VL041542.D	19 Aug 2024 10:48	SY/MD	Ok,M
5	QC081324 CAN 10604	VL041543.D	19 Aug 2024 11:29	SY/MD	Ok
6	QC081424 CAN 10280	VL041544.D	19 Aug 2024 12:10	SY/MD	Ok
7	QC081424 CAN 10054	VL041545.D	19 Aug 2024 12:51	SY/MD	Ok
8	P3658-01	VL041546.D	19 Aug 2024 13:33	SY/MD	Ok,M
9	P3658-01DUP	VL041547.D	19 Aug 2024 14:14	SY/MD	Ok,M
10	P3658-02	VL041548.D	19 Aug 2024 14:56	SY/MD	Ok,M
11	P3659-01	VL041549.D	19 Aug 2024 15:40	SY/MD	Ok,M
12	P3658-03	VL041550.D	19 Aug 2024 16:25	SY/MD	Ok,M
13	P3658-03DL	VL041551.D	19 Aug 2024 17:11	SY/MD	Not Ok
14	P3658-04	VL041552.D	19 Aug 2024 17:56	SY/MD	Ok,M
15	P3658-04DL	VL041553.D	19 Aug 2024 18:40	SY/MD	Not Ok
16	P3658-05	VL041554.D	19 Aug 2024 19:24	SY/MD	Ok,M
17	P3658-05DL	VL041555.D	19 Aug 2024 20:09	SY/MD	Not Ok
18	P3659-02	VL041556.D	19 Aug 2024 20:53	SY/MD	Dilution
19	P3659-02DL	VL041557.D	19 Aug 2024 21:36	SY/MD	Dilution

M : Manual Integration

Instrument ID: **MSVOA_L**

Daily Analysis Runlog For Sequence/QC Batch ID # VL091624

Review By	Semsettin Yesilyurt	Review On	9/17/2024 11:23:42 AM		
Supervise By	Maresh Dadoda	Supervise On	9/17/2024 6:52:14 PM		
SubDirectory	VL091624	HP Acquire Method	AIR-TO15	HP Processing Method	VL091624AIR.M
STD. NAME		STD REF.#			
Tune/Reschk	AP2542				
Initial Calibration Stds	AP2546,AP2547,AP2548				
CCC	AP2546				
Internal Standard/PEM	AP2542				
ICV/I.BLK	AP2544				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VL041590.D	16 Sep 2024 11:50	SY/MD	Ok
2	VSTDICC0.5	VL041591.D	16 Sep 2024 12:36	SY/MD	Ok,M
3	VSTDICCC010	VL041592.D	16 Sep 2024 13:15	SY/MD	Ok,M
4	VSTDICC002	VL041593.D	16 Sep 2024 13:56	SY/MD	Ok,M
5	VSTDICC001	VL041594.D	16 Sep 2024 14:36	SY/MD	Ok,M
6	VSTDICC0.5	VL041595.D	16 Sep 2024 15:16	SY/MD	Not Ok
7	VSTDICC0.1	VL041596.D	16 Sep 2024 15:56	SY/MD	Ok,M
8	VSTDICC0.03	VL041597.D	16 Sep 2024 16:38	SY/MD	Ok,M
9	VSTDICC015	VL041598.D	16 Sep 2024 17:20	SY/MD	Ok
10	VSTDICV010	VL041599.D	16 Sep 2024 18:46	SY/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_L**

Daily Analysis Runlog For Sequence/QC Batch ID # VL091824

Review By	Semsettin Yesilyurt	Review On	9/19/2024 10:46:15 AM		
Supervise By	Maresh Dadoda	Supervise On	9/19/2024 2:44:52 PM		
SubDirectory	VL091824	HP Acquire Method	AIR-TO15	HP Processing Method	VL091624AIR.M
STD. NAME		STD REF.#			
Tune/Reschk		AP2542			
Initial Calibration Stds		AP2546,AP2547,AP2548			
CCC		AP2546			
Internal Standard/PEM		AP2542			
ICV/I.BLK		AP2544			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VL041600.D	18 Sep 2024 8:06	SY/MD	Ok
2	VSTDCCC010	VL041601.D	18 Sep 2024 8:46	SY/MD	Ok,M
3	VL0918ABL01	VL041602.D	18 Sep 2024 9:36	SY/MD	Ok
4	VL0918ABS01	VL041603.D	18 Sep 2024 10:16	SY/MD	Ok,M
5	VIBLK	VL041604.D	18 Sep 2024 13:01	SY/MD	Ok
6	P4028-02	VL041605.D	18 Sep 2024 13:42	SY/MD	Ok,M
7	P4028-02DUP	VL041606.D	18 Sep 2024 14:23	SY/MD	Ok,M
8	P4028-04	VL041607.D	18 Sep 2024 15:03	SY/MD	Ok,M
9	P4028-01	VL041608.D	18 Sep 2024 15:43	SY/MD	Ok,M
10	P4028-01DL	VL041609.D	18 Sep 2024 16:24	SY/MD	Not Ok
11	P4028-03	VL041610.D	18 Sep 2024 17:06	SY/MD	Ok,M
12	P4028-03DL	VL041611.D	18 Sep 2024 17:49	SY/MD	Not Ok
13	QC090324 CAN10197	VL041612.D	18 Sep 2024 18:31	SY/MD	Ok
14	QC090324 CAN10197	VL041613.D	18 Sep 2024 19:14	SY/MD	Not Ok
15	QC090424 CAN10401	VL041614.D	18 Sep 2024 19:56	SY/MD	Ok,M
16	QC090524 CAN10153	VL041615.D	18 Sep 2024 20:38	SY/MD	Ok,M
17	VIBLK	VL041616.D	18 Sep 2024 21:20	SY/MD	Ok,M
18	P4028-03DUP	VL041617.D	18 Sep 2024 22:03	SY/MD	Ok,M
19	VSTDCCC010	VL041618.D	18 Sep 2024 22:44	SY/MD	Ok,M

M : Manual Integration