

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>CHEMTECH</u>	Contract:	<u>TETR16</u>
Lab Code:	<u>CHEM</u>	Case No.:	<u>P5293</u>
Instrument ID:	<u>MSVOA_L</u>	SAS No.:	<u>P5293</u>
Heated Purge: (Y/N)	<u>N</u>	SDG No.:	<u>P5293</u>
GC Column:	<u>RTX-1</u>	Calibration Date(s):	<u>12/16/2024</u>
ID:	<u>0.32 (mm)</u>	Calibration Time(s):	<u>13:30</u>
			<u>16:41</u>

LAB FILE ID:	RRF010 = VL041752.D	RRF002 = VL041753.D	RRF001 = VL041754.D					
	RRF0.5 = VL041755.D	RRF0.1 = VL041756.D	RRF.03 = VL041757.D					
COMPOUND	RRF010	RRF002	RRF001	RRF0.5	RRF0.1	RRF.03	RRF	% RSD
Dichlorodifluoromethane	1.682	1.840	1.927	1.886			1.779	8.6
Chloromethane	0.630	0.736	0.729	0.737			0.696	7.5
Vinyl Chloride	0.625	0.689	0.695	0.682	0.673	0.909	0.701	13.6
Bromomethane	0.295	0.343	0.363	0.343			0.331	8.3
Chloroethane	0.254	0.297	0.267	0.282			0.272	6.3
Tetrahydrofuran	0.427	0.455	0.488	0.455			0.451	5.5
Trichlorofluoromethane	1.791	1.943	1.915	1.981			1.890	4.3
1,1,2-Trichlorotrifluoroethane	1.341	1.466	1.504	1.491			1.438	4.9
Dichlorotetrafluoroethane	1.461	1.616	1.659	1.712			1.593	6.5
tert-Butyl alcohol	1.833	1.975	2.024	1.999			1.930	5
Heptane	1.705	1.875	1.821	1.950			1.816	5.6
1,1-Dichloroethene	0.565	0.622	0.628	0.608			0.603	4.2
Acetone	1.461	2.047	2.284	2.474			1.946	24
Carbon Disulfide	1.346	1.202	1.237	1.024			1.238	11.3
Methyl tert-Butyl Ether	1.098	1.115	1.295	1.033			1.114	9.7
Methylene Chloride	0.477	0.562	0.587	0.653			0.555	12.9
trans-1,2-Dichloroethene	0.640	0.697	0.698	0.749			0.690	6
1,1-Dichloroethane	1.311	1.414	1.459	1.458			1.396	4.9
Cyclohexane	1.096	1.215	1.187	1.217			1.169	4.6
2-Butanone	0.736	0.860	0.820	0.869			0.805	8.1
Carbon Tetrachloride	0.756	0.758	0.716	0.634	0.657	0.671	0.710	7.9
cis-1,2-Dichloroethene	1.331	1.460	1.519	1.430			1.419	5.4
Chloroform	1.969	2.138	2.182	2.155			2.093	4.4
1,1,1-Trichloroethane	2.028	2.101	2.186	2.158	2.113	2.218	2.128	3
2,2,4-Trimethylpentane	1.674	1.842	1.809	1.826			1.767	4.6
Benzene	0.975	1.059	1.080	1.079			1.034	5.2
1,2-Dichloroethane	0.608	0.650	0.642	0.690			0.644	4.7
Trichloroethene	0.388	0.428	0.437	0.408	0.469	0.551	0.439	12.9
1,2-Dichloropropane	0.353	0.388	0.395	0.382			0.374	5.4
Bromodichloromethane	0.777	0.730	0.735	0.664			0.740	6.9

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TETR16
Lab Code: CHEM Case No.: P5293 SAS No.: P5293 SDG No.: P5293
Instrument ID: MSVOA_L Calibration Date(s): 12/16/2024 12/16/2024
Heated Purge: (Y/N) N Calibration Time(s): 13:30 16:41
GC Column: RTX-1 ID: 0.32 (mm)

LAB FILE ID:		RRF010 = VL041752.D		RRF002 = VL041753.D		RRF001 = VL041754.D		RRF0.5 = VL041755.D		RRF0.1 = VL041756.D		RRF.03 = VL041757.D	
COMPOUND	RRF010	RRF002	RRF001	RRF0.5	RRF0.1	RRF.03	RRF	% RSD					
4-Methyl-2-Pentanone	1.060	1.149	1.130	1.062			1.095	3.7					
Toluene	1.123	1.215	1.207	1.183			1.173	3.5					
t-1,3-Dichloropropene	0.423	0.407	0.393	0.382			0.408	5.3					
cis-1,3-Dichloropropene	0.533	0.522	0.493	0.500			0.518	4					
1,1,2-Trichloroethane	0.368	0.406	0.402	0.408			0.391	5.1					
Dibromochloromethane	0.593	0.540	0.493	0.433			0.534	13.6					
1,2-Dibromoethane	0.512	0.542	0.545	0.540	0.565		0.538	3.3					
Tetrachloroethene	0.343	0.372	0.395	0.346	0.333	0.555	0.385	20.2					
Chlorobenzene	0.951	1.088	1.078	1.053			1.023	6.8					
Ethyl Benzene	1.774	1.931	1.834	1.694			1.800	4.9					
m/p-Xylene	1.417	1.522	1.467	1.423			1.448	3.3					
o-Xylene	1.421	1.560	1.492	1.382			1.452	5					
Styrene	0.674	0.636	0.598	0.544			0.628	9.3					
Bromoform	0.498	0.443	0.395	0.328			0.437	17.8					
1,1,2,2-Tetrachloroethane	0.841	0.907	0.909	0.953	0.936	0.979	0.909	6					
2-Chlorotoluene	1.616	1.760	1.699	1.640			1.665	3.8					
1,3,5-Trimethylbenzene	1.465	1.483	1.378	1.324			1.422	4.8					
1,2,4-Trimethylbenzene	1.566	1.619	1.519	1.396			1.529	5.4					
1,3-Dichlorobenzene	0.929	0.994	1.015	1.023			0.975	5.2					
1,4-Dichlorobenzene	0.931	0.992	0.972	0.996			0.963	3.5					
1,2-Dichlorobenzene	0.865	0.983	0.978	1.009			0.938	7.7					
1,2,4-Trichlorobenzene	0.621	0.592	0.553	0.536			0.582	6.3					
Hexachloro-1,3-Butadiene	0.586	0.773	0.756	0.786			0.695	15.1					
1,3-Butadiene	0.685	0.777	0.779	0.891			0.766	10.7					
Naphthalene	1.258	1.154	1.046	0.928	0.754		1.065	18.6					
4-Ethyltoluene	1.731	1.794	1.659	1.681			1.713	3.1					
1-Bromo-4-Fluorobenzene	0.848	0.844	0.819	0.834	0.829	0.827	0.833	1.2					
Hexane	1.303	1.423	1.403	1.423			1.374	4.3					
Allyl Chloride	1.012	1.092	1.126	1.008			1.055	4.9					
1,4-Dioxane	166.814	186.511	181.136	212.045			182.757	10.1					

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TETR16
 Lab Code: CHEM Case No.: P5293 SAS No.: P5293 SDG No.: P5293
 Instrument ID: MSVOA_L Calibration Date(s): 12/16/2024 12/16/2024
 Heated Purge: (Y/N) N Calibration Time(s): 13:30 16:41
 GC Column: RTX-1 ID: 0.32 (mm)

LAB FILE ID:								
RRF010 = VL041752.D			RRF002 = VL041753.D			RRF001 = VL041754.D		
RRF0.5 = VL041755.D			RRF0.1 = VL041756.D			RRF.03 = VL041757.D		
COMPOUND	RRF010	RRF002	RRF001	RRF0.5	RRF0.1	RRF.03	RRF	% RSD
Methyl Methacrylate	0.395	0.406	0.394	0.372			0.394	3.4

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\

Method File : VL121624AIR.M

Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022

Last Update : Thu Dec 19 07:32:09 2024

Response Via : Initial Calibration

Calibration Files

0.03=VL041757.D 0.1 =VL041756.D 0.5 =VL041755.D 1 =VL041754.D 2 =VL041753.D 10 =VL041752.D 15 =VL041758.D

Compound	0.03	0.1	0.5	1	2	10	15	Avg	%RSD

1) I Bromochloromethane	-----ISTD-----								
2) T Dichlorodifluo...			1.886	1.927	1.840	1.682	1.562	1.779	8.61
3) Chlorodifluoro...			1.679	1.697	1.653	1.566	1.559	1.631	3.94
4) Chloromethane			0.737	0.729	0.736	0.630	0.650	0.696	7.48
5) T Vinyl Chloride	0.909	0.673	0.682	0.695	0.689	0.625	0.637	0.701	13.60
6) T Bromomethane			0.343	0.363	0.343	0.295	0.310	0.331	8.32
7) Chloroethane			0.282	0.267	0.297	0.254	0.262	0.272	6.26
8) T Dichlorotetra...			1.712	1.659	1.616	1.461	1.516	1.593	6.49
9) T Propene			0.788	0.734	0.750	0.633	0.629	0.707	10.16
10) T Heptane			1.950	1.821	1.875	1.705	1.726	1.816	5.64
11) T Trichlorofluor...			1.981	1.915	1.943	1.791	1.820	1.890	4.31
12) T 1,1,2-Trichlor...			1.491	1.504	1.466	1.341	1.386	1.437	4.92
13) Ethanol			0.081	0.080	0.059	0.036	0.037	0.059	37.17
14) T Bromoethene			0.509	0.477	0.484	0.474	0.482	0.485	2.84
15) T Acetone			2.474	2.284	2.047	1.461	1.462	1.946	24.00
16) T 1,3-Butadiene			0.891	0.779	0.777	0.685	0.699	0.766	10.74
17) tert-Butyl alc...			1.999	2.024	1.975	1.833	1.819	1.930	5.00
18) T 1,1-Dichloroet...			0.608	0.628	0.622	0.565	0.591	0.603	4.20
19) T Isopropyl Alcohol			1.709	1.320	1.107	0.860	0.841	1.167	30.91
20) T Methylene Chlo...			0.653	0.587	0.562	0.477	0.495	0.555	12.88
21) T Allyl Chloride			1.008	1.126	1.092	1.012	1.038	1.055	4.93
22) T trans-1,2-Dich...			0.749	0.698	0.697	0.640	0.665	0.690	5.96
23) T Vinyl Acetate			0.781	0.811	1.309	0.790	0.826	0.903	25.15
24) T 1,1-Dichloroet...			1.458	1.459	1.414	1.311	1.340	1.396	4.89
25) T Ethyl Acetate			3.618	3.813	3.745	3.430	3.492	3.620	4.48
26) T Hexane			1.423	1.403	1.423	1.303	1.316	1.374	4.32
27) T Carbon Disulfide			1.024	1.237	1.202	1.346	1.380	1.238	11.34
28) T Methyl tert-Bu...			1.033	1.295	1.115	1.098	1.030	1.114	9.69
29) T Chloroform			2.155	2.182	2.138	1.969	2.022	2.093	4.41
30) T Cyclohexane			1.217	1.187	1.215	1.096	1.130	1.169	4.62
31) T cis-1,2-Dichlo...			1.430	1.519	1.460	1.331	1.358	1.419	5.37
32) T 1,1,1-Trichlor...	2.218	2.113	2.158	2.186	2.101	2.028	2.091	2.128	3.03

33) I 1,4-Difluorobenzene	-----ISTD-----								
34) T 2-Butanone			0.869	0.820	0.860	0.736	0.737	0.805	8.05
35) T Carbon Tetrach...	0.671	0.657	0.634	0.716	0.758	0.756	0.777	0.710	7.94
36) T Benzene			1.079	1.080	1.059	0.975	0.979	1.034	5.16
37) T 1,2-Dichloroet...			0.690	0.642	0.650	0.608	0.628	0.644	4.71
38) T Trichloroethene	0.551	0.469	0.408	0.437	0.428	0.388	0.393	0.439	12.92
39) T 1,2-Dichloropr...			0.382	0.395	0.388	0.353	0.353	0.374	5.36

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\										
Method File : VL121624AIR.M										
40)	T	1,4-Dioxane		0.212	0.181	0.187	0.167	0.167	0.183	10.12
41)	T	Tetrahydrofuran		0.455	0.488	0.455	0.427	0.428	0.451	5.52
42)	T	Bromodichlorom...		0.664	0.735	0.730	0.777	0.796	0.740	6.88
43)		Methyl Methacr...		0.372	0.394	0.406	0.395	0.404	0.394	3.38
44)	T	2,2,4-Trimethy...		1.826	1.809	1.842	1.674	1.681	1.767	4.64
45)	T	t-1,3-Dichloro...		0.382	0.393	0.407	0.423	0.434	0.408	5.25
46)	T	cis-1,3-Dichlo...		0.500	0.493	0.522	0.533	0.540	0.518	3.96
47)	T	1,1,2-Trichlor...		0.408	0.402	0.406	0.368	0.371	0.391	5.06
48)	T	Dibromochlorom...		0.433	0.493	0.540	0.593	0.609	0.534	13.58
49)	T	Bromoform		0.328	0.395	0.443	0.498	0.519	0.437	17.80
50)	T	4-Methyl-2-Pen...		1.062	1.130	1.149	1.060	1.076	1.095	3.75
51)	T	2-Hexanone		0.831	0.923	0.953	0.857	0.876	0.888	5.59
52)	T	Tetrachloroethene	0.555 0.333	0.346	0.395	0.372	0.343	0.351	0.385	20.18
53)	T	Toluene		1.183	1.207	1.215	1.123	1.136	1.173	3.54
54)	T	1,2-Dibromoethane	0.565	0.540	0.545	0.542	0.512	0.527	0.538	3.32
-----ISTD-----										
55)	I	Chlorobenzene-d5								
56)		1,1,1,2-Tetrac...		0.487	0.533	0.568	0.532	0.532	0.530	5.43
57)	T	Chlorobenzene		1.053	1.078	1.088	0.951	0.946	1.023	6.78
58)	T	Ethyl Benzene		1.694	1.834	1.931	1.774	1.769	1.800	4.90
59)	T	m/p-Xylene		1.423	1.467	1.522	1.417	1.410	1.448	3.26
60)	T	o-Xylene		1.382	1.492	1.560	1.421	1.404	1.452	5.04
61)	T	Styrene		0.544	0.598	0.636	0.674	0.687	0.628	9.27
62)		Isopropylbenzene		2.138	2.234	2.293	2.135	2.105	2.181	3.64
63)	T	1,1,2,2-Tetrac...	0.979 0.936	0.953	0.909	0.907	0.841	0.835	0.909	5.96
64)		n-propylbenzene		0.520	0.581	0.584	0.535	0.537	0.551	5.28
65)		tert-Butylbenzene		1.964	2.027	2.100	1.892	1.856	1.968	5.04
66)	T	Benzyl Chloride		0.169	0.155	0.165	0.190	0.188	0.173	8.70
67)		sec-Butylbenzene		2.682	2.774	2.926	2.638	2.595	2.723	4.84
68)	S	1-Bromo-4-Fluo...	0.827 0.829	0.834	0.819	0.844	0.848	0.833	0.833	1.18
69)		p-Isopropyltol...		2.149	2.237	2.416	2.223	2.195	2.244	4.54
70)		n-Butylbenzene		1.977	2.184	2.300	2.198	2.183	2.168	5.43
71)		2-Chlorotoluene		1.640	1.698	1.760	1.616	1.609	1.665	3.84
72)	T	4-Ethyltoluene		1.681	1.659	1.794	1.731	1.703	1.713	3.06
73)	T	1,3,5-Trimethy...		1.324	1.378	1.483	1.465	1.458	1.422	4.78
74)	T	1,2,4-Trimethy...		1.396	1.519	1.619	1.566	1.543	1.529	5.41
75)	T	1,3-Dichlorobe...		1.023	1.015	0.994	0.929	0.912	0.975	5.22
76)	T	1,4-Dichlorobe...		0.996	0.972	0.992	0.931	0.924	0.963	3.50
77)	T	1,2-Dichlorobe...		1.009	0.978	0.983	0.865	0.855	0.938	7.68
78)	T	Hexachloro-1,3...		0.786	0.756	0.773	0.586	0.575	0.695	15.13
79)	T	Naphthalene	0.754	0.928	1.046	1.154	1.258	1.251	1.065	18.57
80)	T	Naphthalene,2- ...		0.094	0.093	0.137	0.651	0.673	0.329	92.27
81)	T	1,2,4-Trichlor...		0.536	0.553	0.592	0.621	0.608	0.582	6.25

(#) = Out of Range