

Surrogate Summary

SDG No.: P5293

Client: Tetra Tech, EMI

Analytical Method: SWTO-15

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P5293-01	C0AB9	1-Bromo-4-Fluorobenzene	10	10.2	102	65	135
P5293-01DUP	C0AB9DUP	1-Bromo-4-Fluorobenzene	10	10.1	101	65	135
P5293-02	C0AC0	1-Bromo-4-Fluorobenzene	10	10.4	104	65	135
P5293-03	C0AC1	1-Bromo-4-Fluorobenzene	10	10.9	109	65	135
P5293-03DL	C0AC1DL	1-Bromo-4-Fluorobenzene	10	10.7	107	65	135
VL1218ABL02	VL1218ABL02	1-Bromo-4-Fluorobenzene	10	10.1	101	65	135
VL1218ABS01	VL1218ABS01	1-Bromo-4-Fluorobenzene	10	10.8	108	65	135