

Surrogate Summary

SDG No.: P5294

Client: Tetra Tech, EMI

Analytical Method: SWTO-15

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P5293-01DUP	C0AB9DUP	1-Bromo-4-Fluorobenzene	10	10.1	101	65	135
P5294-01	C0AB6	1-Bromo-4-Fluorobenzene	10	9.72	97	65	135
P5294-02	C0AB7	1-Bromo-4-Fluorobenzene	10	10.2	102	65	135
P5294-03	C0AB8	1-Bromo-4-Fluorobenzene	10	10.1	101	65	135
VL1218ABL02	VL1218ABL02	1-Bromo-4-Fluorobenzene	10	10.1	101	65	135
VL1218ABS01	VL1218ABS01	1-Bromo-4-Fluorobenzene	10	10.8	108	65	135