

Surrogate Summary

SDG No.: P5295

Client: Tetra Tech, EMI

Analytical Method: SWTO-15

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P5293-01DUP	C0AB9DUP	1-Bromo-4-Fluorobenzene	10	10.1	101	65	135
P5295-01	C0AC8	1-Bromo-4-Fluorobenzene	10	9.57	96	65	135
P5295-01DL	C0AC8DL	1-Bromo-4-Fluorobenzene	10	9.67	97	65	135
P5295-02	C0AC9	1-Bromo-4-Fluorobenzene	10	9.86	99	65	135
P5295-02DL	C0AC9DL	1-Bromo-4-Fluorobenzene	10	10.0	100	65	135
VL1218ABL02	VL1218ABL02	1-Bromo-4-Fluorobenzene	10	10.1	101	65	135
VL1218ABS01	VL1218ABS01	1-Bromo-4-Fluorobenzene	10	10.8	108	65	135