

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB134109

Review By	jaswal	Review On	12/27/2024 8:36:01 PM
Supervise By	mohan	Supervise On	12/27/2024 8:37:47 PM

STD. NAME	STD REF.#
ICAL Standard	MP83786,MP83787,MP83788,MP83789,MP83790,MP83791,MP83792
ICV Standard	MP83793
CCV Standard	MP83795
ICSA Standard	
CRI Standard	MP83797
LCS Standard	
Chk Standard	MP83794,MP83796,MP83798,MP83800

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	12/27/24 13:44		Mohan	OK
2	S0.05	S0.05	CAL2	12/27/24 13:46		Mohan	OK
3	S0.2	S0.2	CAL3	12/27/24 13:48		Mohan	OK
4	S2.5	S2.5	CAL4	12/27/24 13:50		Mohan	OK
5	S5	S5	CAL5	12/27/24 13:53		Mohan	OK
6	S7.5	S7.5	CAL6	12/27/24 13:55		Mohan	OK
7	S10	S10	CAL7	12/27/24 14:08		Mohan	OK
8	ICV34	ICV34	ICV	12/27/24 14:13		Mohan	OK
9	ICB34	ICB34	ICB	12/27/24 14:18		Mohan	OK
10	CCV76	CCV76	CCV	12/27/24 14:21		Mohan	OK
11	CCB76	CCB76	CCB	12/27/24 14:23		Mohan	OK
12	CRA	CRA	CRDL	12/27/24 14:30		Mohan	OK
13	HighStd	HighStd	HIGH STD	12/27/24 14:33		Mohan	OK
14	ChkStd	ChkStd	SAM	12/27/24 14:35		Mohan	OK
15	PB165897BL	PB165897BL	MB	12/27/24 14:47		Mohan	OK
16	PB165897BS	PB165897BS	LCS	12/27/24 14:55		Mohan	OK
17	P4368-12	MDL-WATER-06-QT4	SAM	12/27/24 15:08		Mohan	OK
18	P5297-02	SEPT-23	SAM	12/27/24 15:11		Mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB134109

Review By	jaswal	Review On	12/27/2024 8:36:01 PM
Supervise By	mohan	Supervise On	12/27/2024 8:37:47 PM

STD. NAME	STD REF.#
ICAL Standard	MP83786,MP83787,MP83788,MP83789,MP83790,MP83791,MP83792
ICV Standard	MP83793
CCV Standard	MP83795
ICSA Standard	
CRI Standard	MP83797
LCS Standard	
Chk Standard	MP83794,MP83796,MP83798,MP83800

19	P5297-02DUP	SEPT-23DUP	DUP	12/27/24 15:16		Mohan	OK
20	P5297-02MS	SEPT-23MS	MS	12/27/24 15:18		Mohan	OK
21	P5297-02MSD	SEPT-23MSD	MSD	12/27/24 15:20		Mohan	OK
22	CCV77	CCV77	CCV	12/27/24 15:22		Mohan	OK
23	CCB77	CCB77	CCB	12/27/24 15:25		Mohan	OK
24	P4368-10	LOD-MDL-WATER-04	SAM	12/27/24 15:30		Mohan	OK
25	P4368-10RE	LOD-MDL-WATER-04	SAM	12/27/24 15:32		Mohan	OK
26	P5297-02L	SEPT-23L	SD	12/27/24 15:34		Mohan	OK
27	P5297-02A	SEPT-23A	PS	12/27/24 15:36		Mohan	OK
28	CCV78	CCV78	CCV	12/27/24 15:39		Mohan	OK
29	CCB78	CCB78	CCB	12/27/24 15:43		Mohan	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134113

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	12/27/24 12:42		Kareem	OK
2	S1	S1	CAL2	12/27/24 12:47		Kareem	OK
3	S2	S2	CAL3	12/27/24 12:51		Kareem	OK
4	S3	S3	CAL4	12/27/24 12:55		Kareem	OK
5	S4	S4	CAL5	12/27/24 12:59		Kareem	OK
6	S5	S5	CAL6	12/27/24 13:04		Kareem	OK
7	ICV01	ICV01	ICV	12/27/24 13:54	ICV fail for As,Tl (200.7)	Kareem	OK
8	LLICV01	LLICV01	LLICV	12/27/24 14:08		Kareem	OK
9	ICB01	ICB01	ICB	12/27/24 14:59		Kareem	OK
10	CRI01	CRI01	CRDL	12/27/24 15:03		Kareem	OK
11	ICSA01	ICSA01	ICSA	12/27/24 15:16		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	12/27/24 15:20		Kareem	OK
13	ICSADL	ICSADL	ICSA	12/27/24 15:25		Kareem	OK
14	ICSABDL	ICSABDL	ICSAB	12/27/24 15:30		Kareem	OK
15	CCV01	CCV01	CCV	12/27/24 15:34		Kareem	OK
16	CCB01	CCB01	CCB	12/27/24 15:40		Kareem	OK
17	P5297-02	SEPT-23	SAM	12/27/24 15:44		Kareem	OK
18	P5297-02DUP	SEPT-23DUP	DUP	12/27/24 15:49		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134113

Review By	Review On
Supervise By	Supervise On
STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

19	P5297-02L	SEPT-23L	SD	12/27/24 15:53		Kareem	OK
20	P5297-02MS	SEPT-23MS	MS	12/27/24 15:57		Kareem	OK
21	P5297-02MSD	SEPT-23MSD	MSD	12/27/24 16:01		Kareem	OK
22	P5297-02A	SEPT-23A	PS	12/27/24 16:06		Kareem	OK
23	PB165878BL	PB165878BL	MB	12/27/24 16:10		Kareem	OK
24	PB165832BL	PB165832BL	MB	12/27/24 16:20		Kareem	OK
25	PB165832BS	PB165832BS	LCS	12/27/24 16:24		Kareem	OK
26	PB165878BS	PB165878BS	LCS	12/27/24 16:28		Kareem	OK
27	CCV02	CCV02	CCV	12/27/24 16:32		Kareem	OK
28	CCB02	CCB02	CCB	12/27/24 16:37		Kareem	OK
29	LR2	LR2	HIGH STD	12/27/24 16:47		Kareem	OK
30	LR1	LR1	HIGH STD	12/27/24 16:52		Kareem	OK
31	CCV03	CCV03	CCV	12/27/24 16:59		Kareem	OK
32	CCB03	CCB03	CCB	12/27/24 17:17		Kareem	OK