



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Elegant Jewelers Mfg. Co. Inc.

SDG No.: P5297

Contract: ELEG01

Lab Code: CHEM

Case No.: P5297

SAS No.: P5297

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB34	Mercury	0.20	+/-0.20	U	0.20	CV	12/27/2024	14:18	LB134109

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Elegant Jewelers Mfg. Co. Inc.

SDG No.: P5297

Contract: ELEG01

Lab Code: CHEM

Case No.: P5297

SAS No.: P5297

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB76	Mercury	0.20	+/-0.20	U	0.20	CV	12/27/2024	14:23	LB134109
CCB77	Mercury	0.20	+/-0.20	U	0.20	CV	12/27/2024	15:25	LB134109
CCB78	Mercury	0.20	+/-0.20	U	0.20	CV	12/27/2024	15:43	LB134109

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Elegant Jewelers Mfg. Co. Inc.

SDG No.: P5297

Contract: ELEG01

Lab Code: CHEM

Case No.: P5297

SAS No.: P5297

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Cadmium	6.00	+/-6.00	U	6.00	P	12/27/2024	14:59	LB134113
	Chromium	10.0	+/-10.0	U	10.0	P	12/27/2024	14:59	LB134113
	Copper	20.0	+/-20.0	U	20.0	P	12/27/2024	14:59	LB134113
	Lead	12.0	+/-12.0	U	12.0	P	12/27/2024	14:59	LB134113
	Molybdenum	200	+/-200	U	200	P	12/27/2024	14:59	LB134113
	Nickel	40.0	+/-40.0	U	40.0	P	12/27/2024	14:59	LB134113
	Silver	10.0	+/-10.0	U	10.0	P	12/27/2024	14:59	LB134113
	Zinc	40.0	+/-40.0	U	40.0	P	12/27/2024	14:59	LB134113

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Elegant Jewelers Mfg. Co. Inc. **SDG No.:** P5297
Contract: ELEG01 **Lab Code:** CHEM **Case No.:** P5297 **SAS No.:** P5297

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Cadmium	6.00	+/-6.00	U	6.00	P	12/27/2024	15:40	LB134113
	Chromium	10.0	+/-10.0	U	10.0	P	12/27/2024	15:40	LB134113
	Copper	20.0	+/-20.0	U	20.0	P	12/27/2024	15:40	LB134113
	Lead	12.0	+/-12.0	U	12.0	P	12/27/2024	15:40	LB134113
	Molybdenum	200	+/-200	U	200	P	12/27/2024	15:40	LB134113
	Nickel	40.0	+/-40.0	U	40.0	P	12/27/2024	15:40	LB134113
	Silver	10.0	+/-10.0	U	10.0	P	12/27/2024	15:40	LB134113
	Zinc	40.0	+/-40.0	U	40.0	P	12/27/2024	15:40	LB134113
CCB02	Cadmium	6.00	+/-6.00	U	6.00	P	12/27/2024	16:37	LB134113
	Chromium	10.0	+/-10.0	U	10.0	P	12/27/2024	16:37	LB134113
	Copper	20.0	+/-20.0	U	20.0	P	12/27/2024	16:37	LB134113
	Lead	12.0	+/-12.0	U	12.0	P	12/27/2024	16:37	LB134113
	Molybdenum	200	+/-200	U	200	P	12/27/2024	16:37	LB134113
	Nickel	40.0	+/-40.0	U	40.0	P	12/27/2024	16:37	LB134113
	Silver	10.0	+/-10.0	U	10.0	P	12/27/2024	16:37	LB134113
	Zinc	40.0	+/-40.0	U	40.0	P	12/27/2024	16:37	LB134113
CCB03	Cadmium	6.00	+/-6.00	U	6.00	P	12/27/2024	17:17	LB134113
	Chromium	10.0	+/-10.0	U	10.0	P	12/27/2024	17:17	LB134113
	Copper	20.0	+/-20.0	U	20.0	P	12/27/2024	17:17	LB134113
	Lead	12.0	+/-12.0	U	12.0	P	12/27/2024	17:17	LB134113
	Molybdenum	200	+/-200	U	200	P	12/27/2024	17:17	LB134113
	Nickel	40.0	+/-40.0	U	40.0	P	12/27/2024	17:17	LB134113
	Silver	10.0	+/-10.0	U	10.0	P	12/27/2024	17:17	LB134113
	Zinc	40.0	+/-40.0	U	40.0	P	12/27/2024	17:17	LB134113