

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

SEPT-23L

Lab Name: Chemtech Consulting Group

Contract: ELEG01

Lab Code: CHEM **Lb No.:** lb134113 **Lab Sample ID :** P5297-02L **SDG No.:** P5297

Matrix (soil/water): Water **Level (low/med):** LOW

Concentration Units: ug/L

Analyte	Initial Sample Result (I)		Serial Dilution Result (S)		% Difference	Q	M
		C		C			
Cadmium	0.27	J	15.0	U	100.0		P
Chromium	14.5		13.2	J	9		P
Copper	562		579		3		P
Lead	2.53	J	30.0	U	100.0		P
Mercury	0.17	J	0.25	J	42		CV
Molybdenum	100	U	500	U			P
Nickel	24.2		23.3	J	4		P
Silver	78.6		76.9		2		P
Zinc	70.2		35.6	J	49		P