



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/18/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 12/17/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:12
Out Date: 12/18/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133976

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P5245-03	72-12016	1	1.15	8.37	9.52	8.99	93.7	
P5299-01	SB-01	2	1.15	8.40	9.55	7.74	78.5	
P5299-02	SB-02	3	1.16	8.70	9.86	7.73	75.5	
P5299-03	SB-01	33	1.13	8.61	9.74	7.05	68.8	
P5299-04	SB-02	4	1.15	8.75	9.9	8.08	79.2	
P5306-01	OU4-VSL-07-121224	5	1.16	8.52	9.68	8.9	90.8	
P5306-03	OU4-VSL-08-121224	6	1.17	8.73	9.9	9.1	90.8	
P5306-05	OU4-VSL-09-121224	7	1.19	8.45	9.64	8.8	90.1	
P5306-07	OU4-VSL-10-121224	8	1.15	8.65	9.8	9.37	95.0	
P5306-09	OU4-VSL-11-121224	9	1.11	8.77	9.88	9.32	93.6	
P5306-11	OU4-VSL-12-121224	10	1.12	8.65	9.77	8.97	90.8	
P5306-13	OU4-VSL-13-121224	11	1.13	8.72	9.85	8.98	90.0	
P5306-15	OU4-VSL-14-121224	12	1.18	8.46	9.64	9.29	95.9	
P5306-17	OU4-VSL-06R-121224	13	1.15	8.80	9.95	9.22	91.7	
P5307-01	1A-1B-1C-ROOF-2	14	1.00	1.00	2.00	2.00	100.0	caluk
P5307-02	2A-2B-2C-ROOF-2	15	1.00	1.00	2.00	2.00	100.0	caluk
P5307-03	3A-3B-3C-1907	16	1.00	1.00	2.00	2.00	100.0	caluk
P5307-04	4A-4B-4C-1907	17	1.00	1.00	2.00	2.00	100.0	caluk
P5307-05	5A-5B-5C-1907	18	1.00	1.00	2.00	2.00	100.0	caluk
P5307-06	6A-6B-6C-1952	19	1.00	1.00	2.00	2.00	100.0	caluk
P5307-07	1907-BLDG-GRAY	20	1.00	1.00	2.00	2.00	100.0	caluk
P5307-08	1952-BLDG	21	1.00	1.00	2.00	2.00	100.0	caluk
P5307-09	9A-9B-9C-1907	22	1.00	1.00	2.00	2.00	100.0	caluk
P5307-10	1907-BLDG-OFF-WHITE	23	1.00	1.00	2.00	2.00	100.0	caluk
P5307-11	11A-11B-11C-1952-BLDG	24	1.00	1.00	2.00	2.00	100.0	caluk
P5307-12	12A-12B-12C-1952	25	1.00	1.00	2.00	2.00	100.0	caluk
P5307-13	13A-13B-13C-1952	26	1.00	1.00	2.00	2.00	100.0	caluk
P5307-14	14A-14B-14C-1907	27	1.00	1.00	2.00	2.00	100.0	caluk



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/18/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 12/17/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:12
Out Date: 12/18/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133976

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P5307-15	15A-15B-15C-ROOF-7	28	1.00	1.00	2.00	2.00	100.0	caluk
P5312-01	SOIL-VNJ-222	29	1.15	8.43	9.58	8.55	87.8	
P5312-02	SOIL-VNJ-222	30	1.12	8.66	9.78	9.35	95.0	
P5312-03	CONCRETE-VNJ-222	31	1.00	1.00	2.00	2.00	100.0	CONCRETE sample
P5312-04	CONCRETE-VNJ-222	32	1.00	1.00	2.00	2.00	100.0	CONCRETE sample

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$