



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Portal Partners Tri-Venture **SDG No.:** P5299
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P5299 **SAS No.:** P5299
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.16	0.2	82	40 - 160	CV	12/18/2024	13:45	LB133991
CRI01	Aluminum	102	100	102	40 - 160	P	12/20/2024	16:26	LB134054
	Antimony	49.3	50.0	99	40 - 160	P	12/20/2024	16:26	LB134054
	Arsenic	21.1	20.0	106	40 - 160	P	12/20/2024	16:26	LB134054
	Barium	87.8	100	88	40 - 160	P	12/20/2024	16:26	LB134054
	Beryllium	5.71	6.0	95	40 - 160	P	12/20/2024	16:26	LB134054
	Cadmium	5.88	6.0	98	40 - 160	P	12/20/2024	16:26	LB134054
	Calcium	1880	2000	94	40 - 160	P	12/20/2024	16:26	LB134054
	Chromium	10.2	10.0	102	40 - 160	P	12/20/2024	16:26	LB134054
	Cobalt	29.0	30.0	97	40 - 160	P	12/20/2024	16:26	LB134054
	Copper	21.8	20.0	109	40 - 160	P	12/20/2024	16:26	LB134054
	Iron	96.6	100	97	40 - 160	P	12/20/2024	16:26	LB134054
	Lead	11.1	12.0	93	40 - 160	P	12/20/2024	16:26	LB134054
	Magnesium	2050	2000	103	40 - 160	P	12/20/2024	16:26	LB134054
	Manganese	18.9	20.0	95	40 - 160	P	12/20/2024	16:26	LB134054
	Nickel	39.0	40.0	98	40 - 160	P	12/20/2024	16:26	LB134054
	Potassium	2100	2000	105	40 - 160	P	12/20/2024	16:26	LB134054
	Selenium	18.3	20.0	92	40 - 160	P	12/20/2024	16:26	LB134054
	Silver	10.3	10.0	103	40 - 160	P	12/20/2024	16:26	LB134054
	Sodium	1950	2000	97	40 - 160	P	12/20/2024	16:26	LB134054
	Thallium	38.5	40.0	96	40 - 160	P	12/20/2024	16:26	LB134054
	Vanadium	33.3	40.0	83	40 - 160	P	12/20/2024	16:26	LB134054
	Zinc	38.8	40.0	97	40 - 160	P	12/20/2024	16:26	LB134054