

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client:	Portal Partners Tri-Venture	SDG No.:	P5299
Contract:	PORT06	Lab Code:	CHEM
ICS Source:	EPA	Case No.:	P5299
		SAS No.:	P5299
		Instrument ID:	P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	239000	255000	94	216000	294000	12/20/2024	16:30	LB134054
	Antimony	0.15			-50	50	12/20/2024	16:30	LB134054
	Arsenic	-1.51			-20	20	12/20/2024	16:30	LB134054
	Barium	2.79	6.0	46	-94	106	12/20/2024	16:30	LB134054
	Beryllium	1.43			-6	6	12/20/2024	16:30	LB134054
	Cadmium	-0.99	1.0	99	-5	7	12/20/2024	16:30	LB134054
	Calcium	225000	245000	92	208000	282000	12/20/2024	16:30	LB134054
	Chromium	60.0	52.0	115	42	62	12/20/2024	16:30	LB134054
	Cobalt	1.46			-30	30	12/20/2024	16:30	LB134054
	Copper	16.4	2.0	820	-18	22	12/20/2024	16:30	LB134054
	Iron	103000	101000	102	85600	116500	12/20/2024	16:30	LB134054
	Lead	6.97			-12	12	12/20/2024	16:30	LB134054
	Magnesium	247000	255000	97	216000	294000	12/20/2024	16:30	LB134054
	Manganese	4.79	7.0	68	-13	27	12/20/2024	16:30	LB134054
	Nickel	2.13	2.0	106	-38	42	12/20/2024	16:30	LB134054
	Potassium	136			0	0	12/20/2024	16:30	LB134054
	Selenium	-16.6			-20	20	12/20/2024	16:30	LB134054
	Silver	-3.17			-10	10	12/20/2024	16:30	LB134054
	Sodium	5.98			0	0	12/20/2024	16:30	LB134054
	Thallium	11.1			-40	40	12/20/2024	16:30	LB134054
	Vanadium	4.97			-40	40	12/20/2024	16:30	LB134054
	Zinc	0.10			-40	40	12/20/2024	16:30	LB134054
ICSAB01	Aluminum	236000	247000	96	209000	285000	12/20/2024	16:34	LB134054
	Antimony	632	618	102	525	711	12/20/2024	16:34	LB134054
	Arsenic	113	104	109	88.4	120	12/20/2024	16:34	LB134054
	Barium	460	537	86	437	637	12/20/2024	16:34	LB134054
	Beryllium	492	495	99	420	570	12/20/2024	16:34	LB134054
	Cadmium	1040	972	107	826	1120	12/20/2024	16:34	LB134054
	Calcium	222000	235000	94	199000	271000	12/20/2024	16:34	LB134054
	Chromium	604	542	111	460	624	12/20/2024	16:34	LB134054
	Cobalt	523	476	110	404	548	12/20/2024	16:34	LB134054
	Copper	520	511	102	434	588	12/20/2024	16:34	LB134054
	Iron	101000	99300	102	84400	114500	12/20/2024	16:34	LB134054
	Lead	56.2	49.0	115	37	61	12/20/2024	16:34	LB134054
	Magnesium	244000	248000	98	210000	286000	12/20/2024	16:34	LB134054
	Manganese	487	507	96	430	584	12/20/2024	16:34	LB134054
	Nickel	1040	954	109	810	1100	12/20/2024	16:34	LB134054
	Potassium	163			0	0	12/20/2024	16:34	LB134054
	Selenium	34.2	46.0	74	26	66	12/20/2024	16:34	LB134054
	Silver	201	201	100	170	232	12/20/2024	16:34	LB134054
	Sodium	5.70			0	0	12/20/2024	16:34	LB134054
	Thallium	95.5	108	88	68	148	12/20/2024	16:34	LB134054

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Portal Partners Tri-Venture **SDG No.:** P5299
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P5299 **SAS No.:** P5299
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Vanadium	474	491	96	417	565	12/20/2024	16:34	LB134054
	Zinc	1080	952	113	809	1095	12/20/2024	16:34	LB134054