

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Portal Partners Tri-Venture **SDG No.:** P5299
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P5299 **SAS No.:** P5299

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165722BS							
Aluminum	mg/Kg	88.1	79.5		90	80 - 120	P
Antimony	mg/Kg	35.2	33.9		96	80 - 120	P
Arsenic	mg/Kg	35.2	35.1		100	80 - 120	P
Barium	mg/Kg	8.8	7.36		84	80 - 120	P
Beryllium	mg/Kg	8.8	8.30		94	80 - 120	P
Cadmium	mg/Kg	8.8	8.38		95	80 - 120	P
Calcium	mg/Kg	44.1	40.7	J	92	80 - 120	P
Chromium	mg/Kg	17.6	17.8		101	80 - 120	P
Cobalt	mg/Kg	8.8	8.42		96	80 - 120	P
Copper	mg/Kg	13.2	13.4		102	80 - 120	P
Iron	mg/Kg	130	127		98	80 - 120	P
Lead	mg/Kg	44.1	42.0		95	80 - 120	P
Magnesium	mg/Kg	88.1	84.3	J	96	80 - 120	P
Manganese	mg/Kg	8.8	8.44		96	80 - 120	P
Nickel	mg/Kg	22.0	21.3		97	80 - 120	P
Potassium	mg/Kg	440	416		94	80 - 120	P
Selenium	mg/Kg	88.1	85.7		97	80 - 120	P
Silver	mg/Kg	3.3	3.28		99	80 - 120	P
Sodium	mg/Kg	130	128		98	80 - 120	P
Thallium	mg/Kg	88.1	85.7		97	80 - 120	P
Vanadium	mg/Kg	13.2	11.9		90	80 - 120	P
Zinc	mg/Kg	8.8	8.69		99	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: P5299

Contract: PORT06

Lab Code: CHEM

Case No.: P5299

SAS No.: P5299

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165728BS Mercury	mg/Kg	0.26	0.22		86	80 - 124	CV