

Continuing Calibration Report for SequenceID : FC122024AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068059.D

Aliphatic C28-C40	12271986.000	120.000	16.792	21.609	102266.550	105964.639	3.490
Aliphatic EPH	46593571.000	360.000	3.117	21.609	129426.586	135215.008	4.281
Aliphatic C9-C12	9123272.000	60.000	3.117	6.389	152054.533	158302.459	3.947
Aliphatic C12-C16	6147282.000	40.000	6.390	9.779	153682.050	160275.290	4.114
Aliphatic C16-C21	8675091.000	60.000	9.780	13.137	144584.850	152362.625	5.105
Aliphatic C21-C28	10375940.000	80.000	13.138	16.791	129699.250	136384.122	4.902