

Continuing Calibration Report for SequenceID : FC122024AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068071.D

Aliphatic C28-C40	12785603.000	120.000	16.792	21.609	106546.692	105964.639	-0.549
Aliphatic EPH	42958637.000	360.000	3.117	21.609	119329.547	135215.008	11.748
Aliphatic C9-C12	7756045.000	60.000	3.117	6.389	129267.417	158302.459	18.341
Aliphatic C12-C16	5307951.000	40.000	6.390	9.779	132698.775	160275.290	17.206
Aliphatic C16-C21	7636984.000	60.000	9.780	13.137	127283.067	152362.625	16.460
Aliphatic C21-C28	9472054.000	80.000	13.138	16.791	118400.675	136384.122	13.186