

Continuing Calibration Report for SequenceID : FD122024AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048897.D

Aromatic C10-C12	7444754.000	40.000	4.083	5.800	186118.850	190209.437	2.151
Aromatic C12-C16	10696363.000	60.000	5.801	8.405	178272.717	185883.358	4.094
Aromatic C16-C21	12508773.000	80.000	8.406	12.669	156359.663	169519.943	7.763
Aromatic C21-C36	24222747.000	180.000	12.670	18.077	134570.817	141588.666	4.957
Aromatic EPH	54872637.000	360.000	4.083	18.077	152423.992	160580.373	5.079