

Continuing Calibration Report for SequenceID : FD122024AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048909.D

Aromatic C10-C12	7122880.000	40.000	4.083	5.800	178072.000	190209.437	6.381
Aromatic C12-C16	10357272.000	60.000	5.801	8.405	172621.200	185883.358	7.135
Aromatic C16-C21	11644106.000	80.000	8.406	12.669	145551.325	169519.943	14.139
Aromatic C21-C36	23719010.000	180.000	12.670	18.077	131772.278	141588.666	6.933
Aromatic EPH	52843268.000	360.000	4.083	18.077	146786.856	160580.373	8.590