

**SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY**

**Lab Name:** Chemtech **Client:** Portal Partners Tri-Venture  
**Lab Code:** CHEM **Cas No:** P5299 **SAS No :** P5299 **SDG No:** P5299  
**Sample No :** PB165775BS **Datafile:** FC068061.D  
**Client ID :** PB165775BS

| COMPOUND          | SPIKE<br>ADDED<br>mg/kg | LCS/LCSD<br>CONCENTRATION<br>mg/kg | % REC | Qual | QC LIMITS |
|-------------------|-------------------------|------------------------------------|-------|------|-----------|
| Aliphatic C9-C12  | 10.0                    | 5.50                               | 55    |      | (40-140)  |
| Aliphatic C12-C16 | 6.7                     | 5.33                               | 80    |      | (40-140)  |
| Aliphatic C16-C21 | 10.0                    | 8.99                               | 90    |      | (40-140)  |
| Aliphatic C21-C28 | 13.3                    | 12.4                               | 93    |      | (40-140)  |
| Aliphatic C28-C40 | 20.0                    | 22.4                               | 112   |      | (40-140)  |