

### CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P5299 SAS No.: P5299 SDG NO.: P5299

Continuing Calib Date: 12/18/2024 Initial Calibration Date(s): 12/06/2024 12/06/2024

Continuing Calib Time: 15:41 Initial Calibration Time(s): 14:19 22:34

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1254-1       | (1) | 5.61       | 5.62      | 5.52      | 5.72 | 0.01       |
| Aroclor-1254-2       | (2) | 5.76       | 5.77      | 5.67      | 5.87 | 0.01       |
| Aroclor-1254-3       | (3) | 6.17       | 6.17      | 6.07      | 6.27 | 0.00       |
| Aroclor-1254-4       | (4) | 6.40       | 6.40      | 6.30      | 6.50 | 0.00       |
| Aroclor-1254-5       | (5) | 6.82       | 6.83      | 6.73      | 6.93 | 0.01       |
| Tetrachloro-m-xylene |     | 3.71       | 3.71      | 3.61      | 3.81 | 0.00       |
| Decachlorobiphenyl   |     | 8.77       | 8.79      | 8.69      | 8.89 | 0.02       |

### CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P5299 SAS No.: P5299 SDG NO.: P5299

Continuing Calib Date: 12/18/2024 Initial Calibration Date(s): 12/06/2024 12/06/2024

Continuing Calib Time: 15:41 Initial Calibration Time(s): 14:19 22:34

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1254-1 (1)   | 5.59       | 5.60      | 5.50      | 5.70 | 0.01       |
| Aroclor-1254-2 (2)   | 5.74       | 5.75      | 5.65      | 5.85 | 0.01       |
| Aroclor-1254-3 (3)   | 6.14       | 6.15      | 6.05      | 6.25 | 0.01       |
| Aroclor-1254-4 (4)   | 6.37       | 6.38      | 6.28      | 6.48 | 0.01       |
| Aroclor-1254-5 (5)   | 6.79       | 6.80      | 6.70      | 6.90 | 0.01       |
| Tetrachloro-m-xylene | 3.70       | 3.71      | 3.61      | 3.81 | 0.01       |
| Decachlorobiphenyl   | 8.72       | 8.74      | 8.64      | 8.84 | 0.02       |