

Surrogate Summary

SDG No.: **P5299**

Client: **Portal Partners Tri-Venture**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO108361.D	PIBLK-PO108361.D	Tetrachloro-m-xylene	1	20	23.0	115		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	23.6	118		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	21.6	108		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	23.6	118		70 (60)	130 (140)
I.BLK-PO108624.D	PIBLK-PO108624.D	Tetrachloro-m-xylene	1	20	20.4	102		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	20.2	101		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	21.3	107		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	23.4	117		70 (60)	130 (140)
PB165703BL	PB165703BL	Tetrachloro-m-xylene	1	20	18.9	94		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	19.2	96		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	19.5	98		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	22.0	110		30 (32)	150 (175)
PB165703BS	PB165703BS	Tetrachloro-m-xylene	1	20	17.7	89		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	19.3	96		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	18.4	92		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	22.3	112		30 (32)	150 (175)
P5299-01	SB-01	Tetrachloro-m-xylene	1	20	21.6	108		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	19.0	95		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	22.7	113		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	21.9	109		30 (32)	150 (175)
P5299-02	SB-02	Tetrachloro-m-xylene	1	20	21.2	106		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	18.9	94		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	22.3	111		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	21.8	109		30 (32)	150 (175)
P5299-03	SB-01	Tetrachloro-m-xylene	1	20	19.8	99		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	15.2	76		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	20.8	104		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	17.5	88		30 (32)	150 (175)
P5299-04	SB-02	Tetrachloro-m-xylene	1	20	19.8	99		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	17.1	85		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	20.7	103		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	19.7	99		30 (32)	150 (175)
P5306-01MS	OU4-VSL-07-121224MS	Tetrachloro-m-xylene	1	20	22.4	112		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	23.8	119		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	23.1	115		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	27.4	137		30 (32)	150 (175)
P5306-01MSD	OU4-VSL-07-121224MSD	Tetrachloro-m-xylene	1	20	22.5	112		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	20.6	103		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	23.3	117		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	23.9	119		30 (32)	150 (175)
I.BLK-PO108639.D	PIBLK-PO108639.D	Tetrachloro-m-xylene	1	20	20.5	103		70 (60)	130 (140)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: P5299

Client: Portal Partners Tri-Venture

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO108639.D	PIBLK-PO108639.D	Decachlorobiphenyl	1	20	20.3	102		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	21.5	108		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	24.1	121		70 (60)	130 (140)