

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: PORT06

Lab Code: CHEM

SAS No.: P5299 SDG NO.: P5299

Lab File ID: BF140993.D

DFTPP Injection Date: 12/27/2024

Instrument ID: BNA_F

DFTPP Injection Time: 13:33

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	39.9
68	Less than 2.0% of mass 69	0.7 (1.8) 1
69	Mass 69 relative abundance	37.1
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	49
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
275	10.0 - 60.0% of mass 198	27.3
365	Greater than 1% of mass 198	3.5
441	Present, but less than mass 443	13.2
442	Greater than 50% of mass 198	84.7
443	15.0 - 24.0% of mass 442	16.2 (19.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF140994.D	12/27/2024	13:59
SB-02	P5299-04	BF140996.D	12/27/2024	15:00
SB-01	P5299-03	BF140997.D	12/27/2024	15:26
SB-02	P5299-02	BF140998.D	12/27/2024	16:03
SB-01	P5299-01	BF140999.D	12/27/2024	16:29