

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: P5299 SAS No.: P5299 SDG No.: P5299

Instrument ID: MSVOA\_Y Calibration Date(s): 12/17/2024 12/17/2024

Heated Purge: (Y/N) Y Calibration Time(s): 10:01 14:51

GC Column: RXI-624 ID: 0.25 (mm)

|                                |                     |        |        |                     |        |                     |       |                     |  |                     |  |                     |  |
|--------------------------------|---------------------|--------|--------|---------------------|--------|---------------------|-------|---------------------|--|---------------------|--|---------------------|--|
| LAB FILE ID:                   | RRF010 = VY020613.D |        |        | RRF020 = VY020614.D |        | RRF050 = VY020615.D |       | RRF100 = VY020616.D |  | RRF150 = VY020617.D |  | RRF005 = VY020619.D |  |
| COMPOUND                       | RRF010              | RRF020 | RRF050 | RRF100              | RRF150 | RRF005              | RRF   | % RSD               |  |                     |  |                     |  |
| Dichlorodifluoromethane        | 0.678               | 0.633  | 0.617  | 0.601               | 0.552  | 0.665               | 0.625 | 7.3                 |  |                     |  |                     |  |
| Chloromethane                  | 0.490               | 0.506  | 0.428  | 0.426               | 0.372  | 0.555               | 0.463 | 14.3                |  |                     |  |                     |  |
| Vinyl Chloride                 | 0.443               | 0.469  | 0.371  | 0.440               | 0.374  | 0.487               | 0.431 | 11.3                |  |                     |  |                     |  |
| Bromomethane                   | 0.291               | 0.279  | 0.229  | 0.282               | 0.254  | 0.327               | 0.277 | 12.1                |  |                     |  |                     |  |
| Chloroethane                   | 0.273               | 0.287  | 0.212  | 0.272               | 0.251  | 0.302               | 0.266 | 11.8                |  |                     |  |                     |  |
| Trichlorofluoromethane         | 0.981               | 0.954  | 0.791  | 0.878               | 0.810  | 0.993               | 0.901 | 9.7                 |  |                     |  |                     |  |
| 1,1,2-Trichlorotrifluoroethane | 0.715               | 0.659  | 0.603  | 0.649               | 0.537  | 0.712               | 0.646 | 10.5                |  |                     |  |                     |  |
| 1,1-Dichloroethene             | 0.701               | 0.637  | 0.589  | 0.638               | 0.528  | 0.667               | 0.627 | 9.7                 |  |                     |  |                     |  |
| Acetone                        | 0.141               | 0.107  | 0.098  | 0.110               |        | 0.168               | 0.125 | 23.2                |  |                     |  |                     |  |
| Carbon Disulfide               | 2.315               | 2.083  | 1.939  | 2.092               | 1.669  | 2.142               | 2.040 | 10.7                |  |                     |  |                     |  |
| Methyl tert-butyl Ether        | 1.799               | 1.584  | 1.663  | 1.652               | 1.466  | 1.542               | 1.618 | 7.1                 |  |                     |  |                     |  |
| Methyl Acetate                 | 0.474               | 0.351  | 0.350  | 0.391               | 0.292  | 0.378               | 0.373 | 16.1                |  |                     |  |                     |  |
| Methylene Chloride             | 0.758               | 0.683  | 0.617  | 0.665               | 0.596  | 0.766               | 0.681 | 10.3                |  |                     |  |                     |  |
| trans-1,2-Dichloroethene       | 0.771               | 0.703  | 0.710  | 0.692               | 0.636  | 0.707               | 0.703 | 6.2                 |  |                     |  |                     |  |
| 1,1-Dichloroethane             | 1.401               | 1.307  | 1.323  | 1.300               | 1.173  | 1.333               | 1.306 | 5.7                 |  |                     |  |                     |  |
| Cyclohexane                    | 1.388               | 1.247  | 1.198  | 1.160               | 1.056  | 1.443               | 1.249 | 11.6                |  |                     |  |                     |  |
| 2-Butanone                     | 0.245               | 0.191  | 0.197  | 0.195               | 0.166  | 0.228               | 0.204 | 13.9                |  |                     |  |                     |  |
| Carbon Tetrachloride           | 0.768               | 0.715  | 0.701  | 0.697               | 0.656  | 0.730               | 0.711 | 5.2                 |  |                     |  |                     |  |
| cis-1,2-Dichloroethene         | 0.875               | 0.815  | 0.817  | 0.797               | 0.724  | 0.827               | 0.809 | 6.1                 |  |                     |  |                     |  |
| Bromochloromethane             | 0.436               | 0.353  | 0.495  | 0.501               | 0.432  | 0.427               | 0.441 | 12.3                |  |                     |  |                     |  |
| Chloroform                     | 1.768               | 1.465  | 1.378  | 1.302               | 1.186  | 2.069               | 1.528 | 21.6                |  |                     |  |                     |  |
| 1,1,1-Trichloroethane          | 1.298               | 1.215  | 1.198  | 1.183               | 1.086  | 1.219               | 1.200 | 5.7                 |  |                     |  |                     |  |
| Methylcyclohexane              | 0.883               | 0.848  | 0.832  | 0.812               | 0.766  | 0.883               | 0.837 | 5.3                 |  |                     |  |                     |  |
| Benzene                        | 2.055               | 1.900  | 1.859  | 1.847               | 1.713  | 1.927               | 1.883 | 5.9                 |  |                     |  |                     |  |
| 1,2-Dichloroethane             | 0.546               | 0.499  | 0.500  | 0.498               | 0.452  | 0.502               | 0.500 | 6                   |  |                     |  |                     |  |
| Trichloroethene                | 0.501               | 0.462  | 0.458  | 0.453               | 0.422  | 0.481               | 0.463 | 5.8                 |  |                     |  |                     |  |
| 1,2-Dichloropropane            | 0.473               | 0.450  | 0.434  | 0.440               | 0.401  | 0.442               | 0.440 | 5.3                 |  |                     |  |                     |  |
| Bromodichloromethane           | 0.672               | 0.632  | 0.632  | 0.627               | 0.581  | 0.626               | 0.628 | 4.6                 |  |                     |  |                     |  |
| 4-Methyl-2-Pentanone           | 0.364               | 0.290  | 0.299  | 0.301               | 0.261  | 0.297               | 0.302 | 11.1                |  |                     |  |                     |  |
| Toluene                        | 1.247               | 1.185  | 1.167  | 1.165               | 1.082  | 1.176               | 1.170 | 4.5                 |  |                     |  |                     |  |

\* Compounds with required minimum RRF and maximum %RSD values.  
All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

|                     |                  |                      |                   |
|---------------------|------------------|----------------------|-------------------|
| Lab Name:           | <u>CHEMTECH</u>  | Contract:            | <u>PORT06</u>     |
| Lab Code:           | <u>CHEM</u>      | Case No.:            | <u>P5299</u>      |
| Instrument ID:      | <u>MSVOA_Y</u>   | SAS No.:             | <u>P5299</u>      |
| Heated Purge: (Y/N) | <u>Y</u>         | SDG No.:             | <u>P5299</u>      |
| GC Column:          | <u>RXI-624</u>   | Calibration Date(s): | <u>12/17/2024</u> |
| ID:                 | <u>0.25 (mm)</u> | Calibration Time(s): | <u>10:01</u>      |
|                     |                  |                      | <u>14:51</u>      |

| LAB FILE ID:                | RRF010 = VY020613.D | RRF020 = VY020614.D | RRF050 = VY020615.D |        |        |        |       |       |
|-----------------------------|---------------------|---------------------|---------------------|--------|--------|--------|-------|-------|
|                             | RRF100 = VY020616.D | RRF150 = VY020617.D | RRF005 = VY020619.D |        |        |        |       |       |
| COMPOUND                    | RRF010              | RRF020              | RRF050              | RRF100 | RRF150 | RRF005 | RRF   | % RSD |
| t-1,3-Dichloropropene       | 0.638               | 0.572               | 0.602               | 0.611  | 0.556  | 0.534  | 0.586 | 6.5   |
| cis-1,3-Dichloropropene     | 0.742               | 0.694               | 0.706               | 0.714  | 0.646  | 0.661  | 0.694 | 5.1   |
| 1,1,2-Trichloroethane       | 0.344               | 0.306               | 0.303               | 0.309  | 0.276  | 0.308  | 0.308 | 7.1   |
| 2-Hexanone                  | 0.225               | 0.183               | 0.199               | 0.202  | 0.175  | 0.180  | 0.194 | 9.5   |
| Dibromochloromethane        | 0.447               | 0.402               | 0.420               | 0.421  | 0.381  | 0.396  | 0.411 | 5.7   |
| 1,2-Dibromoethane           | 0.319               | 0.281               | 0.288               | 0.290  | 0.261  | 0.277  | 0.286 | 6.7   |
| Tetrachloroethene           | 0.505               | 0.464               | 0.452               | 0.446  | 0.422  | 0.533  | 0.470 | 8.7   |
| Chlorobenzene               | 1.501               | 1.397               | 1.374               | 1.381  | 1.293  | 1.441  | 1.398 | 5     |
| Ethyl Benzene               | 2.747               | 2.583               | 2.556               | 2.543  | 2.414  | 2.633  | 2.579 | 4.2   |
| m/p-Xylenes                 | 1.031               | 0.951               | 0.953               | 0.942  | 0.892  | 0.984  | 0.959 | 4.8   |
| o-Xylene                    | 0.956               | 0.892               | 0.898               | 0.887  | 0.842  | 0.894  | 0.895 | 4.1   |
| Styrene                     | 1.586               | 1.496               | 1.502               | 1.501  | 1.406  | 1.478  | 1.495 | 3.8   |
| Bromoform                   | 0.302               | 0.256               | 0.270               | 0.272  | 0.250  | 0.262  | 0.269 | 6.9   |
| Isopropylbenzene            | 4.636               | 4.451               | 4.356               | 4.284  | 4.051  | 4.401  | 4.363 | 4.4   |
| 1,1,2,2-Tetrachloroethane   | 0.804               | 0.702               | 0.710               | 0.710  | 0.635  | 0.689  | 0.708 | 7.7   |
| 1,3-Dichlorobenzene         | 1.982               | 1.912               | 1.864               | 1.832  | 1.702  | 1.966  | 1.876 | 5.5   |
| 1,4-Dichlorobenzene         | 1.980               | 1.852               | 1.832               | 1.797  | 1.680  | 2.008  | 1.858 | 6.5   |
| 1,2-Dichlorobenzene         | 1.735               | 1.614               | 1.606               | 1.602  | 1.501  | 1.694  | 1.626 | 5     |
| 1,2-Dibromo-3-Chloropropane | 0.133               | 0.107               | 0.113               | 0.114  | 0.107  | 0.116  | 0.115 | 8.4   |
| 1,2,4-Trichlorobenzene      | 0.975               | 0.927               | 1.014               | 1.030  | 0.990  | 0.964  | 0.983 | 3.8   |
| 1,2,3-Trichlorobenzene      | 0.798               | 0.791               | 0.864               | 0.874  | 0.846  | 0.813  | 0.831 | 4.2   |
| 1,2-Dichloroethane-d4       | 0.619               | 0.487               | 0.569               | 0.507  | 0.467  | 0.636  | 0.548 | 13    |
| Dibromofluoromethane        | 0.385               | 0.328               | 0.360               | 0.325  | 0.313  | 0.404  | 0.352 | 10.4  |
| Toluene-d8                  | 1.314               | 1.127               | 1.271               | 1.143  | 1.100  | 1.445  | 1.233 | 10.9  |
| 4-Bromofluorobenzene        | 0.538               | 0.448               | 0.480               | 0.433  | 0.412  | 0.585  | 0.482 | 13.9  |

\* Compounds with required minimum RRF and maximum %RSD values.  
All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.