

Surrogate Summary

SDG No.: **P5299**

Client: **Portal Partners Tri-Venture**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P5299-01	SB-01	1,2-Dichloroethane-d4	50	49.0	98	70 (50)	130 (163)
		Dibromofluoromethane	50	45.8	92	70 (54)	130 (147)
		Toluene-d8	50	48.4	97	70 (58)	130 (134)
		4-Bromofluorobenzene	50	38.7	77	70 (29)	130 (146)
P5299-02	SB-02	1,2-Dichloroethane-d4	50	49.8	100	70 (50)	130 (163)
		Dibromofluoromethane	50	45.2	90	70 (54)	130 (147)
		Toluene-d8	50	48.9	98	70 (58)	130 (134)
		4-Bromofluorobenzene	50	40.5	81	70 (29)	130 (146)
P5299-03	SB-01	1,2-Dichloroethane-d4	50	48.6	97	70 (50)	130 (163)
		Dibromofluoromethane	50	43.6	87	70 (54)	130 (147)
		Toluene-d8	50	49.2	98	70 (58)	130 (134)
		4-Bromofluorobenzene	50	40.0	80	70 (29)	130 (146)
P5299-04	SB-02	1,2-Dichloroethane-d4	50	58.5	117	70 (50)	130 (163)
		Dibromofluoromethane	50	40.3	81	70 (54)	130 (147)
		Toluene-d8	50	48.7	97	70 (58)	130 (134)
		4-Bromofluorobenzene	50	41.0	82	70 (29)	130 (146)
VY1218SBL01	VY1218SBL01	1,2-Dichloroethane-d4	50	49.7	99	70 (50)	130 (163)
		Dibromofluoromethane	50	44.7	89	70 (54)	130 (147)
		Toluene-d8	50	48.4	97	70 (58)	130 (134)
		4-Bromofluorobenzene	50	37.4	75	70 (29)	130 (146)
VY1218SBS01	VY1218SBS01	1,2-Dichloroethane-d4	50	48.4	97	70 (50)	130 (163)
		Dibromofluoromethane	50	48.3	97	70 (54)	130 (147)
		Toluene-d8	50	48.7	97	70 (58)	130 (134)
		4-Bromofluorobenzene	50	46.3	93	70 (29)	130 (146)
VY1218SBSD01	VY1218SBSD01	1,2-Dichloroethane-d4	50	51.7	103	70 (50)	130 (163)
		Dibromofluoromethane	50	50.1	100	70 (54)	130 (147)
		Toluene-d8	50	51.5	103	70 (58)	130 (134)
		4-Bromofluorobenzene	50	48.3	97	70 (29)	130 (146)

