



#### **SDG NARRATIVE**

LAB NAME: Alliance Technical Group, LLC

CASE: 51914 SDG: C0CQ2

CONTRACT: 68HERH20D0011

LAB CODE: ACE

**LAB ORDER ID: P5303** 

**MODIFICATION REF. NUMBER: NA** 

Sample ID	<b>EPA Sample ID</b>	pН
P5303-01	C0CQ2	1.0
P5303-02	C0CQ3	1.0
P5303-03	C0CQ4	1.0
P5303-04	C0CQ5	1.0

04 Water samples were delivered to the laboratory intact on 12/17/2024.

Test requested on the Chain of Custody was Trace Volatile Organic by Method SFAM01.1.

The temperature of the samples was measured using an I R Gun. The samples temperature was 2.1 degree Celsius for the samples received on 12/17/2024.

#### **Trace Volatiles:**

The analysis performed on instrument MSVOA\_U were done using GC column DB-624UI 20m 0.18mm 1.0 um. Cat#121-1324UI.

The analysis of VOC-SFAM was based on method SFAM01.1\_Trace.

Holding Times were met requirement.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

Instrument Performance Check met requirements.

The Retention Times met requirements.

The Tuning criteria met requirements.

The initial Calibration criteria met requirements.

The Continuing Calibration (VSTD005087) file ID VU062588.D met the requirements except for Dibromochloromethane (23.2%). As per method, up to two target analyte in opening and closing CCV are allowed to exceed the %D values. Therefore no further corrective action was taken.



The Blank analysis did not indicate the presence of lab contamination. The storage blank analysis did not indicate the presence of lab contamination.

See **Manual Integration report** for the manual integration information at the end of the case narrative.

## **Calculation:**

## Low/Med Water Level Calculation

Concentration in ug/L = 
$$(Ax) (Is) (DF)$$
  
(Ais) (RRF) (Vo)

Where,

Ax = Area of the characteristic ion (EICP) for the compound to be measured.

Ais = Area of the characteristic ion (EICP) for the internal standard.

Amount of internal standard added in ng.

RRF = Mean Relative Response Factor from the initial calibration standard.

Vo = Total volume of water purged, in mL.

DF = Dilution Factor

## Example calculation of **C0CQ2** for **Acetone**:

```
Ax= 8937

Is = 125

RRF= 0.044

DF= 1

Ais= 77965

Vo. = 25

Concentration in ug/L = (8937) (125) (1)

(77965)(0.044)(25)
```

Reported Result = 13.03 ug/L

Final Reported Result = 13 ug/L

# Relative Response Factor = **Dichlorodifluoromethane**: RUN **VU120924** for **0.5** ppb

RRF= 0.475





I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature	Name: Nimisha Pandya.	
Date:	Title: Document Control Officer.	