



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Nobis Group</u>				SDG No.:	<u>P5306</u>				
Contract:	<u>NOBI03</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>P5306</u>	SAS No.:	<u>P5306</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB24	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	12/18/2024	13:39	LB133991

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>P5306</u>								
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P5306</u>	SAS No.: <u>P5306</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB42	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	12/18/2024	13:43	LB133991
CCB43	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	12/18/2024	14:11	LB133991
CCB44	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	12/18/2024	14:48	LB133991
CCB45	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	12/18/2024	15:09	LB133991

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group SDG No.: P5306
Contract: NOBI03 Lab Code: CHEM Case No.: P5306 SAS No.: P5306

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	80.0	100	P	12/18/2024	15:03	LB134001
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	12/18/2024	15:03	LB134001
	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	12/18/2024	15:03	LB134001
	Barium	100	+/-100	U	25.0	100	P	12/18/2024	15:03	LB134001
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	12/18/2024	15:03	LB134001
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	12/18/2024	15:03	LB134001
	Calcium	2000	+/-2000	U	500	2000	P	12/18/2024	15:03	LB134001
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	12/18/2024	15:03	LB134001
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	12/18/2024	15:03	LB134001
	Copper	20.0	+/-20.0	U	16.0	20.0	P	12/18/2024	15:03	LB134001
	Iron	100	+/-100	U	80.0	100	P	12/18/2024	15:03	LB134001
	Lead	12.0	+/-12.0	U	9.60	12.0	P	12/18/2024	15:03	LB134001
	Magnesium	2000	+/-2000	U	500	2000	P	12/18/2024	15:03	LB134001
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	12/18/2024	15:03	LB134001
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	12/18/2024	15:03	LB134001
	Potassium	2000	+/-2000	U	1600	2000	P	12/18/2024	15:03	LB134001
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	12/18/2024	15:03	LB134001
	Silver	10.0	+/-10.0	U	5.00	10.0	P	12/18/2024	15:03	LB134001
	Sodium	2000	+/-2000	U	1000	2000	P	12/18/2024	15:03	LB134001
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	12/18/2024	15:03	LB134001
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	12/18/2024	15:03	LB134001
	Zinc	40.0	+/-40.0	U	10.0	40.0	P	12/18/2024	15:03	LB134001

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>P5306</u>								
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P5306</u>	SAS No.: <u>P5306</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	80.0	100	P	12/18/2024	15:31	LB134001
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	12/18/2024	15:31	LB134001
	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	12/18/2024	15:31	LB134001
	Barium	100	+/-100	U	25.0	100	P	12/18/2024	15:31	LB134001
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	12/18/2024	15:31	LB134001
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	12/18/2024	15:31	LB134001
	Calcium	73.6	+/-2000	J	500	2000	P	12/18/2024	15:31	LB134001
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	12/18/2024	15:31	LB134001
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	12/18/2024	15:31	LB134001
	Copper	20.0	+/-20.0	U	16.0	20.0	P	12/18/2024	15:31	LB134001
	Iron	100	+/-100	U	80.0	100	P	12/18/2024	15:31	LB134001
	Lead	12.0	+/-12.0	U	9.60	12.0	P	12/18/2024	15:31	LB134001
	Magnesium	2000	+/-2000	U	500	2000	P	12/18/2024	15:31	LB134001
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	12/18/2024	15:31	LB134001
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	12/18/2024	15:31	LB134001
	Potassium	2000	+/-2000	U	1600	2000	P	12/18/2024	15:31	LB134001
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	12/18/2024	15:31	LB134001
	Silver	10.0	+/-10.0	U	5.00	10.0	P	12/18/2024	15:31	LB134001
	Sodium	2000	+/-2000	U	1000	2000	P	12/18/2024	15:31	LB134001
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	12/18/2024	15:31	LB134001
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	12/18/2024	15:31	LB134001
	Zinc	7.56	+/-40.0	J	10.0	40.0	P	12/18/2024	15:31	LB134001
CCB02	Aluminum	100	+/-100	U	80.0	100	P	12/18/2024	16:14	LB134001
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	12/18/2024	16:14	LB134001
	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	12/18/2024	16:14	LB134001
	Barium	100	+/-100	U	25.0	100	P	12/18/2024	16:14	LB134001
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	12/18/2024	16:14	LB134001
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	12/18/2024	16:14	LB134001
	Calcium	2000	+/-2000	U	500	2000	P	12/18/2024	16:14	LB134001
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	12/18/2024	16:14	LB134001
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	12/18/2024	16:14	LB134001
	Copper	20.0	+/-20.0	U	16.0	20.0	P	12/18/2024	16:14	LB134001
	Iron	100	+/-100	U	80.0	100	P	12/18/2024	16:14	LB134001
	Lead	12.0	+/-12.0	U	9.60	12.0	P	12/18/2024	16:14	LB134001
	Magnesium	2000	+/-2000	U	500	2000	P	12/18/2024	16:14	LB134001
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	12/18/2024	16:14	LB134001
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	12/18/2024	16:14	LB134001
	Potassium	2000	+/-2000	U	1600	2000	P	12/18/2024	16:14	LB134001
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	12/18/2024	16:14	LB134001

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>P5306</u>								
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P5306</u>	SAS No.: <u>P5306</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	5.00	10.0	P	12/18/2024	16:14	LB134001
	Sodium	2000	+/-2000	U	1000	2000	P	12/18/2024	16:14	LB134001
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	12/18/2024	16:14	LB134001
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	12/18/2024	16:14	LB134001
	Zinc	40.0	+/-40.0	U	10.0	40.0	P	12/18/2024	16:14	LB134001
CCB03	Aluminum	100	+/-100	U	80.0	100	P	12/18/2024	17:05	LB134001
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	12/18/2024	17:05	LB134001
	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	12/18/2024	17:05	LB134001
	Barium	100	+/-100	U	25.0	100	P	12/18/2024	17:05	LB134001
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	12/18/2024	17:05	LB134001
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	12/18/2024	17:05	LB134001
	Calcium	2000	+/-2000	U	500	2000	P	12/18/2024	17:05	LB134001
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	12/18/2024	17:05	LB134001
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	12/18/2024	17:05	LB134001
	Copper	20.0	+/-20.0	U	16.0	20.0	P	12/18/2024	17:05	LB134001
	Iron	100	+/-100	U	80.0	100	P	12/18/2024	17:05	LB134001
	Lead	12.0	+/-12.0	U	9.60	12.0	P	12/18/2024	17:05	LB134001
	Magnesium	2000	+/-2000	U	500	2000	P	12/18/2024	17:05	LB134001
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	12/18/2024	17:05	LB134001
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	12/18/2024	17:05	LB134001
	Potassium	2000	+/-2000	U	1600	2000	P	12/18/2024	17:05	LB134001
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	12/18/2024	17:05	LB134001
	Silver	10.0	+/-10.0	U	5.00	10.0	P	12/18/2024	17:05	LB134001
	Sodium	2000	+/-2000	U	1000	2000	P	12/18/2024	17:05	LB134001
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	12/18/2024	17:05	LB134001
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	12/18/2024	17:05	LB134001
	Zinc	40.0	+/-40.0	U	10.0	40.0	P	12/18/2024	17:05	LB134001
CCB04	Aluminum	100	+/-100	U	80.0	100	P	12/18/2024	17:47	LB134001
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	12/18/2024	17:47	LB134001
	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	12/18/2024	17:47	LB134001
	Barium	100	+/-100	U	25.0	100	P	12/18/2024	17:47	LB134001
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	12/18/2024	17:47	LB134001
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	12/18/2024	17:47	LB134001
	Calcium	2000	+/-2000	U	500	2000	P	12/18/2024	17:47	LB134001
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	12/18/2024	17:47	LB134001
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	12/18/2024	17:47	LB134001
	Copper	20.0	+/-20.0	U	16.0	20.0	P	12/18/2024	17:47	LB134001
	Iron	100	+/-100	U	80.0	100	P	12/18/2024	17:47	LB134001
	Lead	12.0	+/-12.0	U	9.60	12.0	P	12/18/2024	17:47	LB134001

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>P5306</u>								
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P5306</u>	SAS No.: <u>P5306</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	500	2000	P	12/18/2024	17:47	LB134001
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	12/18/2024	17:47	LB134001
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	12/18/2024	17:47	LB134001
	Potassium	2000	+/-2000	U	1600	2000	P	12/18/2024	17:47	LB134001
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	12/18/2024	17:47	LB134001
	Silver	10.0	+/-10.0	U	5.00	10.0	P	12/18/2024	17:47	LB134001
	Sodium	2000	+/-2000	U	1000	2000	P	12/18/2024	17:47	LB134001
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	12/18/2024	17:47	LB134001
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	12/18/2024	17:47	LB134001
	Zinc	40.0	+/-40.0	U	10.0	40.0	P	12/18/2024	17:47	LB134001
CCB05	Aluminum	100	+/-100	U	80.0	100	P	12/18/2024	18:37	LB134001
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	12/18/2024	18:37	LB134001
	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	12/18/2024	18:37	LB134001
	Barium	100	+/-100	U	25.0	100	P	12/18/2024	18:37	LB134001
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	12/18/2024	18:37	LB134001
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	12/18/2024	18:37	LB134001
	Calcium	2000	+/-2000	U	500	2000	P	12/18/2024	18:37	LB134001
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	12/18/2024	18:37	LB134001
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	12/18/2024	18:37	LB134001
	Copper	20.0	+/-20.0	U	16.0	20.0	P	12/18/2024	18:37	LB134001
	Iron	100	+/-100	U	80.0	100	P	12/18/2024	18:37	LB134001
	Lead	12.0	+/-12.0	U	9.60	12.0	P	12/18/2024	18:37	LB134001
	Magnesium	2000	+/-2000	U	500	2000	P	12/18/2024	18:37	LB134001
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	12/18/2024	18:37	LB134001
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	12/18/2024	18:37	LB134001
	Potassium	2000	+/-2000	U	1600	2000	P	12/18/2024	18:37	LB134001
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	12/18/2024	18:37	LB134001
	Silver	10.0	+/-10.0	U	5.00	10.0	P	12/18/2024	18:37	LB134001
	Sodium	2000	+/-2000	U	1000	2000	P	12/18/2024	18:37	LB134001
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	12/18/2024	18:37	LB134001
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	12/18/2024	18:37	LB134001
	Zinc	40.0	+/-40.0	U	10.0	40.0	P	12/18/2024	18:37	LB134001

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Nobis Group				SDG No.:	P5306				
Contract:	NOBI03	Lab Code:	CHEM	Case No.:	P5306	SAS No.: P5306				
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	80.0	100	P	12/23/2024	11:44	LB134066
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	12/23/2024	11:44	LB134066
	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	12/23/2024	11:44	LB134066
	Barium	100	+/-100	U	25.0	100	P	12/23/2024	11:44	LB134066
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	12/23/2024	11:44	LB134066
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	12/23/2024	11:44	LB134066
	Calcium	2000	+/-2000	U	500	2000	P	12/23/2024	11:44	LB134066
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	12/23/2024	11:44	LB134066
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	12/23/2024	11:44	LB134066
	Copper	20.0	+/-20.0	U	16.0	20.0	P	12/23/2024	11:44	LB134066
	Iron	100	+/-100	U	80.0	100	P	12/23/2024	11:44	LB134066
	Lead	12.0	+/-12.0	U	9.60	12.0	P	12/23/2024	11:44	LB134066
	Magnesium	2000	+/-2000	U	500	2000	P	12/23/2024	11:44	LB134066
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	12/23/2024	11:44	LB134066
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	12/23/2024	11:44	LB134066
	Potassium	2000	+/-2000	U	1600	2000	P	12/23/2024	11:44	LB134066
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	12/23/2024	11:44	LB134066
	Silver	10.0	+/-10.0	U	5.00	10.0	P	12/23/2024	11:44	LB134066
	Sodium	2000	+/-2000	U	1000	2000	P	12/23/2024	11:44	LB134066
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	12/23/2024	11:44	LB134066
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	12/23/2024	11:44	LB134066
	Zinc	40.0	+/-40.0	U	10.0	40.0	P	12/23/2024	11:44	LB134066

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>P5306</u>								
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P5306</u>		SAS No.: <u>P5306</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	80.0	100	P	12/23/2024	12:35	LB134066
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	12/23/2024	12:35	LB134066
	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	12/23/2024	12:35	LB134066
	Barium	100	+/-100	U	25.0	100	P	12/23/2024	12:35	LB134066
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	12/23/2024	12:35	LB134066
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	12/23/2024	12:35	LB134066
	Calcium	2000	+/-2000	U	500	2000	P	12/23/2024	12:35	LB134066
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	12/23/2024	12:35	LB134066
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	12/23/2024	12:35	LB134066
	Copper	20.0	+/-20.0	U	16.0	20.0	P	12/23/2024	12:35	LB134066
	Iron	100	+/-100	U	80.0	100	P	12/23/2024	12:35	LB134066
	Lead	12.0	+/-12.0	U	9.60	12.0	P	12/23/2024	12:35	LB134066
	Magnesium	2000	+/-2000	U	500	2000	P	12/23/2024	12:35	LB134066
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	12/23/2024	12:35	LB134066
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	12/23/2024	12:35	LB134066
	Potassium	2000	+/-2000	U	1600	2000	P	12/23/2024	12:35	LB134066
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	12/23/2024	12:35	LB134066
	Silver	10.0	+/-10.0	U	5.00	10.0	P	12/23/2024	12:35	LB134066
	Sodium	2000	+/-2000	U	1000	2000	P	12/23/2024	12:35	LB134066
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	12/23/2024	12:35	LB134066
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	12/23/2024	12:35	LB134066
	Zinc	40.0	+/-40.0	U	10.0	40.0	P	12/23/2024	12:35	LB134066
CCB02	Aluminum	100	+/-100	U	80.0	100	P	12/23/2024	13:47	LB134066
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	12/23/2024	13:47	LB134066
	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	12/23/2024	13:47	LB134066
	Barium	100	+/-100	U	25.0	100	P	12/23/2024	13:47	LB134066
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	12/23/2024	13:47	LB134066
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	12/23/2024	13:47	LB134066
	Calcium	2000	+/-2000	U	500	2000	P	12/23/2024	13:47	LB134066
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	12/23/2024	13:47	LB134066
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	12/23/2024	13:47	LB134066
	Copper	20.0	+/-20.0	U	16.0	20.0	P	12/23/2024	13:47	LB134066
	Iron	100	+/-100	U	80.0	100	P	12/23/2024	13:47	LB134066
	Lead	12.0	+/-12.0	U	9.60	12.0	P	12/23/2024	13:47	LB134066
	Magnesium	2000	+/-2000	U	500	2000	P	12/23/2024	13:47	LB134066
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	12/23/2024	13:47	LB134066
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	12/23/2024	13:47	LB134066
	Potassium	2000	+/-2000	U	1600	2000	P	12/23/2024	13:47	LB134066
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	12/23/2024	13:47	LB134066

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>P5306</u>								
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P5306</u>	SAS No.: <u>P5306</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	5.00	10.0	P	12/23/2024	13:47	LB134066
	Sodium	2000	+/-2000	U	1000	2000	P	12/23/2024	13:47	LB134066
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	12/23/2024	13:47	LB134066
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	12/23/2024	13:47	LB134066
	Zinc	40.0	+/-40.0	U	10.0	40.0	P	12/23/2024	13:47	LB134066
CCB03	Aluminum	100	+/-100	U	80.0	100	P	12/23/2024	14:11	LB134066
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	12/23/2024	14:11	LB134066
	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	12/23/2024	14:11	LB134066
	Barium	100	+/-100	U	25.0	100	P	12/23/2024	14:11	LB134066
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	12/23/2024	14:11	LB134066
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	12/23/2024	14:11	LB134066
	Calcium	2000	+/-2000	U	500	2000	P	12/23/2024	14:11	LB134066
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	12/23/2024	14:11	LB134066
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	12/23/2024	14:11	LB134066
	Copper	20.0	+/-20.0	U	16.0	20.0	P	12/23/2024	14:11	LB134066
	Iron	100	+/-100	U	80.0	100	P	12/23/2024	14:11	LB134066
	Lead	12.0	+/-12.0	U	9.60	12.0	P	12/23/2024	14:11	LB134066
	Magnesium	2000	+/-2000	U	500	2000	P	12/23/2024	14:11	LB134066
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	12/23/2024	14:11	LB134066
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	12/23/2024	14:11	LB134066
	Potassium	2000	+/-2000	U	1600	2000	P	12/23/2024	14:11	LB134066
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	12/23/2024	14:11	LB134066
	Silver	10.0	+/-10.0	U	5.00	10.0	P	12/23/2024	14:11	LB134066
	Sodium	2000	+/-2000	U	1000	2000	P	12/23/2024	14:11	LB134066
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	12/23/2024	14:11	LB134066
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	12/23/2024	14:11	LB134066
	Zinc	40.0	+/-40.0	U	10.0	40.0	P	12/23/2024	14:11	LB134066
CCB04	Aluminum	100	+/-100	U	80.0	100	P	12/23/2024	15:10	LB134066
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	12/23/2024	15:10	LB134066
	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	12/23/2024	15:10	LB134066
	Barium	100	+/-100	U	25.0	100	P	12/23/2024	15:10	LB134066
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	12/23/2024	15:10	LB134066
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	12/23/2024	15:10	LB134066
	Calcium	2000	+/-2000	U	500	2000	P	12/23/2024	15:10	LB134066
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	12/23/2024	15:10	LB134066
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	12/23/2024	15:10	LB134066
	Copper	20.0	+/-20.0	U	16.0	20.0	P	12/23/2024	15:10	LB134066
	Iron	100	+/-100	U	80.0	100	P	12/23/2024	15:10	LB134066
	Lead	12.0	+/-12.0	U	9.60	12.0	P	12/23/2024	15:10	LB134066

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>P5306</u>								
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P5306</u>	SAS No.: <u>P5306</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	500	2000	P	12/23/2024	15:10	LB134066
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	12/23/2024	15:10	LB134066
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	12/23/2024	15:10	LB134066
	Potassium	2000	+/-2000	U	1600	2000	P	12/23/2024	15:10	LB134066
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	12/23/2024	15:10	LB134066
	Silver	10.0	+/-10.0	U	5.00	10.0	P	12/23/2024	15:10	LB134066
	Sodium	2000	+/-2000	U	1000	2000	P	12/23/2024	15:10	LB134066
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	12/23/2024	15:10	LB134066
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	12/23/2024	15:10	LB134066
	Zinc	40.0	+/-40.0	U	10.0	40.0	P	12/23/2024	15:10	LB134066
CCB05	Aluminum	100	+/-100	U	80.0	100	P	12/23/2024	16:10	LB134066
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	12/23/2024	16:10	LB134066
	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	12/23/2024	16:10	LB134066
	Barium	100	+/-100	U	25.0	100	P	12/23/2024	16:10	LB134066
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	12/23/2024	16:10	LB134066
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	12/23/2024	16:10	LB134066
	Calcium	2000	+/-2000	U	500	2000	P	12/23/2024	16:10	LB134066
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	12/23/2024	16:10	LB134066
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	12/23/2024	16:10	LB134066
	Copper	20.0	+/-20.0	U	16.0	20.0	P	12/23/2024	16:10	LB134066
	Iron	100	+/-100	U	80.0	100	P	12/23/2024	16:10	LB134066
	Lead	12.0	+/-12.0	U	9.60	12.0	P	12/23/2024	16:10	LB134066
	Magnesium	2000	+/-2000	U	500	2000	P	12/23/2024	16:10	LB134066
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	12/23/2024	16:10	LB134066
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	12/23/2024	16:10	LB134066
	Potassium	2000	+/-2000	U	1600	2000	P	12/23/2024	16:10	LB134066
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	12/23/2024	16:10	LB134066
	Silver	1.18	+/-10.0	J	5.00	10.0	P	12/23/2024	16:10	LB134066
	Sodium	2000	+/-2000	U	1000	2000	P	12/23/2024	16:10	LB134066
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	12/23/2024	16:10	LB134066
CCB06	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	12/23/2024	16:10	LB134066
	Zinc	40.0	+/-40.0	U	10.0	40.0	P	12/23/2024	16:10	LB134066
	Aluminum	100	+/-100	U	80.0	100	P	12/23/2024	16:58	LB134066
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	12/23/2024	16:58	LB134066
	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	12/23/2024	16:58	LB134066
	Barium	100	+/-100	U	25.0	100	P	12/23/2024	16:58	LB134066
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	12/23/2024	16:58	LB134066
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	12/23/2024	16:58	LB134066
	Calcium	2000	+/-2000	U	500	2000	P	12/23/2024	16:58	LB134066

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>P5306</u>								
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P5306</u>	SAS No.: <u>P5306</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	5.00	10.0	P	12/23/2024	16:58	LB134066
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	12/23/2024	16:58	LB134066
	Copper	20.0	+/-20.0	U	16.0	20.0	P	12/23/2024	16:58	LB134066
	Iron	100	+/-100	U	80.0	100	P	12/23/2024	16:58	LB134066
	Lead	12.0	+/-12.0	U	9.60	12.0	P	12/23/2024	16:58	LB134066
	Magnesium	2000	+/-2000	U	500	2000	P	12/23/2024	16:58	LB134066
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	12/23/2024	16:58	LB134066
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	12/23/2024	16:58	LB134066
	Potassium	2000	+/-2000	U	1600	2000	P	12/23/2024	16:58	LB134066
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	12/23/2024	16:58	LB134066
	Silver	10.0	+/-10.0	U	5.00	10.0	P	12/23/2024	16:58	LB134066
	Sodium	2000	+/-2000	U	1000	2000	P	12/23/2024	16:58	LB134066
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	12/23/2024	16:58	LB134066
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	12/23/2024	16:58	LB134066
	Zinc	40.0	+/-40.0	U	10.0	40.0	P	12/23/2024	16:58	LB134066