

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Nobis Group **SDG No.:** P5306
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** P5306 **SAS No.:** P5306

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165698BS							
Aluminum	mg/Kg	95.7	91.4		96	74 - 119	P
Antimony	mg/Kg	38.3	36.8		96	79 - 114	P
Arsenic	mg/Kg	38.3	37.6		98	82 - 111	P
Barium	mg/Kg	9.6	9.05		94	83 - 113	P
Beryllium	mg/Kg	9.6	9.25		96	83 - 113	P
Cadmium	mg/Kg	9.6	9.05		94	82 - 113	P
Calcium	mg/Kg	47.8	47.6	J	100	81 - 116	P
Chromium	mg/Kg	19.1	19.4		102	85 - 113	P
Cobalt	mg/Kg	9.6	9.19		96	85 - 112	P
Copper	mg/Kg	14.4	14.7		102	81 - 117	P
Iron	mg/Kg	140	144		103	81 - 118	P
Lead	mg/Kg	47.8	45.7		96	81 - 112	P
Magnesium	mg/Kg	95.7	90.9	J	95	78 - 115	P
Manganese	mg/Kg	9.6	9.56		100	84 - 114	P
Nickel	mg/Kg	23.9	23.1		97	83 - 113	P
Potassium	mg/Kg	480	464		97	81 - 116	P
Selenium	mg/Kg	95.7	92.5		97	78 - 111	P
Silver	mg/Kg	3.6	3.68		102	82 - 112	P
Sodium	mg/Kg	140	146		104	83 - 118	P
Thallium	mg/Kg	95.7	93.9		98	83 - 111	P
Vanadium	mg/Kg	14.4	13.8		96	82 - 114	P
Zinc	mg/Kg	9.6	9.90		103	82 - 113	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Nobis Group **SDG No.:** P5306
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** P5306 **SAS No.:** P5306

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165728BS Mercury	mg/Kg	0.26	0.22		86	80 - 124	CV