



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Nobis Group</u>				SDG No.:	<u>P5306</u>				
Contract:	<u>NOBI03</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>P5306</u>	SAS No.:	<u>P5306</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB25	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	12/19/2024	12:00	LB134010

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group **SDG No.:** P5306
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** P5306 **SAS No.:** P5306

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB46	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	12/19/2024	12:05	LB134010
CCB47	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	12/19/2024	13:01	LB134010
CCB48	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	12/19/2024	13:38	LB134010
CCB49	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	12/19/2024	14:02	LB134010
CCB50	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	12/19/2024	14:30	LB134010

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>P5306</u>								
Contract: <u>NOBI03</u>		Lab Code: <u>CHEM</u>		Case No.: <u>P5306</u>		SAS No.: <u>P5306</u>				
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Antimony	2.00	+/-2.00	U	0.25	2.00	P	12/23/2024	12:25	LB134070
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	12/23/2024	12:25	LB134070
	Barium	10.0	+/-10.0	U	1.25	10.0	P	12/23/2024	12:25	LB134070
	Beryllium	1.00	+/-1.00	U	0.25	1.00	P	12/23/2024	12:25	LB134070
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	12/23/2024	12:25	LB134070
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	12/23/2024	12:25	LB134070
	Copper	2.00	+/-2.00	U	1.50	2.00	P	12/23/2024	12:25	LB134070
	Lead	1.00	+/-1.00	U	0.75	1.00	P	12/23/2024	12:25	LB134070
	Nickel	1.00	+/-1.00	U	0.25	1.00	P	12/23/2024	12:25	LB134070
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	12/23/2024	12:25	LB134070
	Silver	1.00	+/-1.00	U	0.50	1.00	P	12/23/2024	12:25	LB134070
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	12/23/2024	12:25	LB134070
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	12/23/2024	12:25	LB134070
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	12/23/2024	12:25	LB134070

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>P5306</u>								
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P5306</u>	SAS No.: <u>P5306</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Antimony	2.00	+/-2.00	U	0.25	2.00	P	12/23/2024	12:43	LB134070
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	12/23/2024	12:43	LB134070
	Barium	10.0	+/-10.0	U	1.25	10.0	P	12/23/2024	12:43	LB134070
	Beryllium	1.00	+/-1.00	U	0.25	1.00	P	12/23/2024	12:43	LB134070
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	12/23/2024	12:43	LB134070
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	12/23/2024	12:43	LB134070
	Copper	2.00	+/-2.00	U	1.50	2.00	P	12/23/2024	12:43	LB134070
	Lead	1.00	+/-1.00	U	0.75	1.00	P	12/23/2024	12:43	LB134070
	Nickel	1.00	+/-1.00	U	0.25	1.00	P	12/23/2024	12:43	LB134070
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	12/23/2024	12:43	LB134070
	Silver	1.00	+/-1.00	U	0.50	1.00	P	12/23/2024	12:43	LB134070
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	12/23/2024	12:43	LB134070
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	12/23/2024	12:43	LB134070
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	12/23/2024	12:43	LB134070
CCB02	Antimony	0.16	+/-2.00	J	0.25	2.00	P	12/23/2024	13:23	LB134070
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	12/23/2024	13:23	LB134070
	Barium	10.0	+/-10.0	U	1.25	10.0	P	12/23/2024	13:23	LB134070
	Beryllium	1.00	+/-1.00	U	0.25	1.00	P	12/23/2024	13:23	LB134070
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	12/23/2024	13:23	LB134070
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	12/23/2024	13:23	LB134070
	Copper	2.00	+/-2.00	U	1.50	2.00	P	12/23/2024	13:23	LB134070
	Lead	0.27	+/-1.00	J	0.75	1.00	P	12/23/2024	13:23	LB134070
	Nickel	1.00	+/-1.00	U	0.25	1.00	P	12/23/2024	13:23	LB134070
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	12/23/2024	13:23	LB134070
	Silver	1.00	+/-1.00	U	0.50	1.00	P	12/23/2024	13:23	LB134070
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	12/23/2024	13:23	LB134070
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	12/23/2024	13:23	LB134070
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	12/23/2024	13:23	LB134070
CCB03	Antimony	0.22	+/-2.00	J	0.25	2.00	P	12/23/2024	13:58	LB134070
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	12/23/2024	13:58	LB134070
	Barium	10.0	+/-10.0	U	1.25	10.0	P	12/23/2024	13:58	LB134070
	Beryllium	1.00	+/-1.00	U	0.25	1.00	P	12/23/2024	13:58	LB134070
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	12/23/2024	13:58	LB134070
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	12/23/2024	13:58	LB134070
	Copper	2.00	+/-2.00	U	1.50	2.00	P	12/23/2024	13:58	LB134070
	Lead	0.17	+/-1.00	J	0.75	1.00	P	12/23/2024	13:58	LB134070
	Nickel	1.00	+/-1.00	U	0.25	1.00	P	12/23/2024	13:58	LB134070
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	12/23/2024	13:58	LB134070
	Silver	1.00	+/-1.00	U	0.50	1.00	P	12/23/2024	13:58	LB134070

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Nobis Group</u>				SDG No.:	<u>P5306</u>				
Contract:	<u>NOBI03</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>P5306</u>	SAS No.:	<u>P5306</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB03	Thallium	1.00	+/-1.00	U	0.50	1.00	P	12/23/2024	13:58	LB134070
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	12/23/2024	13:58	LB134070
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	12/23/2024	13:58	LB134070

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>P5306</u>								
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P5306</u>	SAS No.: <u>P5306</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Antimony	2.00	+/-2.00	U	0.25	2.00	P	12/27/2024	15:38	LB134133
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	12/27/2024	15:38	LB134133
	Barium	10.0	+/-10.0	U	1.25	10.0	P	12/27/2024	15:38	LB134133
	Beryllium	1.00	+/-1.00	U	0.25	1.00	P	12/27/2024	15:38	LB134133
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	12/27/2024	15:38	LB134133
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	12/27/2024	15:38	LB134133
	Copper	2.00	+/-2.00	U	1.50	2.00	P	12/27/2024	15:38	LB134133
	Lead	1.00	+/-1.00	U	0.75	1.00	P	12/27/2024	15:38	LB134133
	Nickel	1.00	+/-1.00	U	0.25	1.00	P	12/27/2024	15:38	LB134133
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	12/27/2024	15:38	LB134133
	Silver	1.00	+/-1.00	U	0.50	1.00	P	12/27/2024	15:38	LB134133
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	12/27/2024	15:38	LB134133
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	12/27/2024	15:38	LB134133
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	12/27/2024	15:38	LB134133

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>P5306</u>								
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P5306</u>		SAS No.: <u>P5306</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Antimony	0.23	+/-2.00	J	0.25	2.00	P	12/27/2024	15:51	LB134133
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	12/27/2024	15:51	LB134133
	Barium	10.0	+/-10.0	U	1.25	10.0	P	12/27/2024	15:51	LB134133
	Beryllium	1.00	+/-1.00	U	0.25	1.00	P	12/27/2024	15:51	LB134133
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	12/27/2024	15:51	LB134133
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	12/27/2024	15:51	LB134133
	Copper	2.00	+/-2.00	U	1.50	2.00	P	12/27/2024	15:51	LB134133
	Lead	0.21	+/-1.00	J	0.75	1.00	P	12/27/2024	15:51	LB134133
	Nickel	1.00	+/-1.00	U	0.25	1.00	P	12/27/2024	15:51	LB134133
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	12/27/2024	15:51	LB134133
	Silver	1.00	+/-1.00	U	0.50	1.00	P	12/27/2024	15:51	LB134133
	Thallium	0.10	+/-1.00	J	0.50	1.00	P	12/27/2024	15:51	LB134133
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	12/27/2024	15:51	LB134133
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	12/27/2024	15:51	LB134133
CCB02	Antimony	0.23	+/-2.00	J	0.25	2.00	P	12/27/2024	16:25	LB134133
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	12/27/2024	16:25	LB134133
	Barium	10.0	+/-10.0	U	1.25	10.0	P	12/27/2024	16:25	LB134133
	Beryllium	1.00	+/-1.00	U	0.25	1.00	P	12/27/2024	16:25	LB134133
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	12/27/2024	16:25	LB134133
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	12/27/2024	16:25	LB134133
	Copper	2.00	+/-2.00	U	1.50	2.00	P	12/27/2024	16:25	LB134133
	Lead	0.26	+/-1.00	J	0.75	1.00	P	12/27/2024	16:25	LB134133
	Nickel	1.00	+/-1.00	U	0.25	1.00	P	12/27/2024	16:25	LB134133
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	12/27/2024	16:25	LB134133
	Silver	0.080	+/-1.00	J	0.50	1.00	P	12/27/2024	16:25	LB134133
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	12/27/2024	16:25	LB134133
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	12/27/2024	16:25	LB134133
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	12/27/2024	16:25	LB134133