

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Nobis Group

SDG No.: P5306

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB165737BL		WATER		Batch Number:		PB165737		Prep Date:	12/19/2024	
	Mercury	0.20	<0.20	U	0.16	0.20	CV	12/19/2024	13:17	LB134010
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB165737TB		WATER		Batch Number:		PB165737		Prep Date:	12/19/2024	
	Mercury	0.20	<0.20	U	0.16	0.20	CV	12/19/2024	14:25	LB134010

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Nobis Group

SDG No.: P5306

Instrument: P7

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB165834BL	WATER			Batch Number:		PB165834		Prep Date:	12/18/2024	
	Antimony	1.00	<1.00	U	0.13	1.00	P	12/23/2024	12:56	LB134070
	Arsenic	0.50	<0.50	U	0.13	0.50	P	12/23/2024	12:56	LB134070
	Barium	5.00	<5.00	U	0.63	5.00	P	12/23/2024	12:56	LB134070
	Beryllium	0.50	<0.50	U	0.13	0.50	P	12/23/2024	12:56	LB134070
	Cadmium	0.50	<0.50	U	0.25	0.50	P	12/23/2024	12:56	LB134070
	Chromium	1.00	<1.00	U	0.38	1.00	P	12/23/2024	12:56	LB134070
	Copper	1.00	<1.00	U	0.75	1.00	P	12/23/2024	12:56	LB134070
	Lead	0.50	<0.50	U	0.38	0.50	P	12/23/2024	12:56	LB134070
	Nickel	0.50	<0.50	U	0.13	0.50	P	12/23/2024	12:56	LB134070
	Selenium	2.50	<2.50	U	2.25	2.50	P	12/23/2024	12:56	LB134070
	Silver	0.50	<0.50	U	0.25	0.50	P	12/23/2024	12:56	LB134070
	Thallium	0.50	<0.50	U	0.25	0.50	P	12/23/2024	12:56	LB134070
	Vanadium	2.50	<2.50	U	0.13	2.50	P	12/23/2024	12:56	LB134070
	Zinc	2.50	<2.50	U	0.75	2.50	P	12/23/2024	12:56	LB134070
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB165834TB	WATER			Batch Number:		PB165834		Prep Date:	12/18/2024	
	Antimony	0.88	<1.00	J	0.13	1.00	P	12/23/2024	13:52	LB134070
	Arsenic	0.50	<0.50	U	0.13	0.50	P	12/23/2024	13:52	LB134070
	Barium	5.00	<5.00	U	0.63	5.00	P	12/23/2024	13:52	LB134070
	Beryllium	0.50	<0.50	U	0.13	0.50	P	12/23/2024	13:52	LB134070
	Cadmium	0.50	<0.50	U	0.25	0.50	P	12/23/2024	13:52	LB134070
	Chromium	1.00	<1.00	U	0.38	1.00	P	12/23/2024	13:52	LB134070
	Copper	1.00	<1.00	U	0.75	1.00	P	12/23/2024	13:52	LB134070
	Lead	0.070	<0.50	J	0.38	0.50	P	12/23/2024	13:52	LB134070
	Nickel	0.50	<0.50	U	0.13	0.50	P	12/23/2024	13:52	LB134070
	Selenium	2.50	<2.50	U	2.25	2.50	P	12/23/2024	13:52	LB134070
	Silver	0.50	<0.50	U	0.25	0.50	P	12/23/2024	13:52	LB134070
	Thallium	0.50	<0.50	U	0.25	0.50	P	12/23/2024	13:52	LB134070
	Vanadium	2.50	<2.50	U	0.13	2.50	P	12/23/2024	13:52	LB134070
	Zinc	0.37	<2.50	J	0.75	2.50	P	12/23/2024	13:52	LB134070