

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Nobis Group **SDG No.:** P5306
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** P5306 **SAS No.:** P5306
ICS Source: EPA **Instrument ID:** P7

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Antimony	1.11	1.5	74	-2.5	5.5	12/23/2024	12:29	LB134070
	Arsenic	0.35	0.1	350	-1.9	2.1	12/23/2024	12:29	LB134070
	Barium	1.42	1.2	118	-18.8	21.2	12/23/2024	12:29	LB134070
	Beryllium	0.30			-2	2	12/23/2024	12:29	LB134070
	Cadmium	0.22	0.7	31	-1.3	2.7	12/23/2024	12:29	LB134070
	Chromium	20.8	21.0	99	17	25	12/23/2024	12:29	LB134070
	Copper	7.22	8.0	90	4	12	12/23/2024	12:29	LB134070
	Lead	4.56	4.0	114	2	6	12/23/2024	12:29	LB134070
	Nickel	5.46	6.0	91	4	8	12/23/2024	12:29	LB134070
	Selenium	0	0.3		-9.7	10	12/23/2024	12:29	LB134070
	Silver	0.040			-2	2	12/23/2024	12:29	LB134070
	Thallium	0.10			-2	2	12/23/2024	12:29	LB134070
	Vanadium	0.17	0.5	34	-9.5	10.5	12/23/2024	12:29	LB134070
	Zinc	11.4	11.0	104	7	15	12/23/2024	12:29	LB134070
ICSAB01	Antimony	20.8	22.0	94	18	26	12/23/2024	12:32	LB134070
	Arsenic	21.4	19.0	113	16.2	21.9	12/23/2024	12:32	LB134070
	Barium	21.1	22.0	96	2	42	12/23/2024	12:32	LB134070
	Beryllium	21.0	19.0	110	16.2	21.9	12/23/2024	12:32	LB134070
	Cadmium	19.5	20.0	98	17	23	12/23/2024	12:32	LB134070
	Chromium	41.0	40.0	102	34	46	12/23/2024	12:32	LB134070
	Copper	25.9	25.0	104	21	29	12/23/2024	12:32	LB134070
	Lead	23.2	25.0	93	21.3	28.8	12/23/2024	12:32	LB134070
	Nickel	25.2	24.0	105	20.4	27.6	12/23/2024	12:32	LB134070
	Selenium	20.3	19.0	107	9	29	12/23/2024	12:32	LB134070
	Silver	18.8	18.0	104	15.3	20.7	12/23/2024	12:32	LB134070
	Thallium	19.2	21.0	91	17.9	24.2	12/23/2024	12:32	LB134070
	Vanadium	20.4	19.0	107	9	29	12/23/2024	12:32	LB134070
	Zinc	31.5	29.0	109	25	33	12/23/2024	12:32	LB134070
ICSA01	Antimony	1.13	1.5	75	-2.5	5.5	12/27/2024	15:42	LB134133
	Arsenic	0.32	0.1	320	-1.9	2.1	12/27/2024	15:42	LB134133
	Barium	1.36	1.2	113	-18.8	21.2	12/27/2024	15:42	LB134133
	Beryllium	0.29			-2	2	12/27/2024	15:42	LB134133
	Cadmium	0.31	0.7	44	-1.3	2.7	12/27/2024	15:42	LB134133
	Chromium	19.8	21.0	94	17	25	12/27/2024	15:42	LB134133
	Copper	7.11	8.0	89	4	12	12/27/2024	15:42	LB134133
	Lead	4.48	4.0	112	2	6	12/27/2024	15:42	LB134133
	Nickel	5.24	6.0	87	4	8	12/27/2024	15:42	LB134133
	Selenium	0.26	0.3	87	-9.7	10	12/27/2024	15:42	LB134133
	Silver	0.040			-2	2	12/27/2024	15:42	LB134133
	Thallium	0.10			-2	2	12/27/2024	15:42	LB134133
	Vanadium	0.20	0.5	40	-9.5	10.5	12/27/2024	15:42	LB134133
	Zinc	10.9	11.0	99	7	15	12/27/2024	15:42	LB134133

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Nobis Group **SDG No.:** P5306
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** P5306 **SAS No.:** P5306
ICS Source: EPA **Instrument ID:** P7

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Antimony	21.5	22.0	98	18	26	12/27/2024	15:45	LB134133
	Arsenic	21.8	19.0	115	16.2	21.9	12/27/2024	15:45	LB134133
	Barium	21.9	22.0	100	2	42	12/27/2024	15:45	LB134133
	Beryllium	21.2	19.0	112	16.2	21.9	12/27/2024	15:45	LB134133
	Cadmium	19.5	20.0	98	17	23	12/27/2024	15:45	LB134133
	Chromium	40.4	40.0	101	34	46	12/27/2024	15:45	LB134133
	Copper	26.4	25.0	106	21	29	12/27/2024	15:45	LB134133
	Lead	24.4	25.0	98	21.3	28.8	12/27/2024	15:45	LB134133
	Nickel	25.6	24.0	107	20.4	27.6	12/27/2024	15:45	LB134133
	Selenium	20.6	19.0	108	9	29	12/27/2024	15:45	LB134133
	Silver	19.6	18.0	109	15.3	20.7	12/27/2024	15:45	LB134133
	Thallium	20.7	21.0	99	17.9	24.2	12/27/2024	15:45	LB134133
	Vanadium	20.3	19.0	107	9	29	12/27/2024	15:45	LB134133
	Zinc	31.6	29.0	109	25	33	12/27/2024	15:45	LB134133