

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Nobis Group **Level:** LOW **SDG No.:** P5306
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** P5306 **SAS No.:** P5306
Matrix: Water **Sample ID:** P5306-02 **Client ID:** OU4-VSL-07-121224DUP
Percent Solids for Sample: NA **Duplicate ID** P5306-02DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Antimony	ug/L	20	0.60	JD	5.00	UD	200		P
Arsenic	ug/L	20	3.57	D	3.65	D	2		P
Barium	ug/L	20	106	D	106	D	0		P
Beryllium	ug/L	20	1.85	JD	1.70	JD	8		P
Cadmium	ug/L	20	2.50	UD	2.50	UD			P
Chromium	ug/L	20	20.1	D	19.4	D	4		P
Copper	ug/L	20	39.6	D	38.5	D	3		P
Lead	ug/L	20	2.78	D	2.10	JD	28	*	P
Nickel	ug/L	20	68.5	D	68.0	D	1		P
Mercury	ug/L	20	0.20	U	0.20	U			CV
Selenium	ug/L	20	25.0	UD	25.0	UD			P
Silver	ug/L	20	0.30	JD	2.50	UD	200		P
Thallium	ug/L	20	0.25	JD	2.50	UD	200		P
Vanadium	ug/L	20	624	D	622	D	0		P
Zinc	ug/L	20	620	D	617	D	0		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Nobis Group **Level:** LOW **SDG No.:** P5306
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** P5306 **SAS No.:** P5306
Matrix: Water **Sample ID:** P5306-02MS **Client ID:** OU4-VSL-07-121224MSD
Percent Solids for Sample: NA **Duplicate ID** P5306-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Antimony	ug/L	20	285	D	283	D	1		P
Arsenic	ug/L	20	248	D	244	D	2		P
Barium	ug/L	20	1490	D	1480	D	1		P
Beryllium	ug/L	20	304	D	295	D	3		P
Cadmium	ug/L	20	308	D	303	D	2		P
Chromium	ug/L	20	309	D	302	D	2		P
Copper	ug/L	20	2550	D	2530	D	1		P
Lead	ug/L	20	1200	D	1190	D	1		P
Nickel	ug/L	20	352	D	343	D	3		P
Mercury	ug/L	20	3.86		3.92		2		CV
Selenium	ug/L	20	283	D	293	D	3		P
Silver	ug/L	20	0.53	JD	0.53	JD	0		P
Thallium	ug/L	20	296	D	290	D	2		P
Vanadium	ug/L	20	895	D	881	D	2		P
Zinc	ug/L	20	2930	D	2900	D	1		P