

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Nobis Group **SDG No.:** P5306
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** P5306 **SAS No.:** P5306

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165737BS Mercury	ug/L	4.0	4.16		104	82 - 119	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Nobis Group **SDG No.:** P5306
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** P5306 **SAS No.:** P5306

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165834BS							
Antimony	ug/L	250	257		103	85 - 117	P
Arsenic	ug/L	250	254		102	84 - 116	P
Barium	ug/L	1300	1280		98	86 - 114	P
Beryllium	ug/L	250	244		98	83 - 121	P
Cadmium	ug/L	250	251		100	87 - 115	P
Chromium	ug/L	250	250		100	85 - 116	P
Copper	ug/L	2500	2510		100	85 - 118	P
Lead	ug/L	1300	1220		94	88 - 115	P
Nickel	ug/L	250	241		96	85 - 117	P
Selenium	ug/L	250	251		100	80 - 120	P
Silver	ug/L	250	262		105	85 - 116	P
Thallium	ug/L	250	242		97	82 - 116	P
Vanadium	ug/L	250	249		100	86 - 115	P
Zinc	ug/L	2500	2530		101	83 - 119	P