

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

OU4-VSL-14-121224MS

Contract: NOBI03

Lab Code: CHEM Case No.: P5306

SAS No.: P5306 SDG NO.: P5306

Lab Sample ID: P5306-15MS

Date(s) Analyzed: 12/26/2024 12/26/2024

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP ID: 0.32 (mm)

GC Column: (2): RTX-CLP2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
2,4,5-T	1	9.49	9.44	9.54	110	12.2
	2	10.25	10.20	10.30	97.4	
2,4,5-TP(Silvex)	1	9.20	9.15	9.25	91.3	10.9
	2	9.84	9.79	9.89	81.9	
Dalapon	1	2.62	2.57	2.67	114	13.1
	2	2.67	2.62	2.72	130	
2,4-D	1	8.33	8.28	8.38	159	13.4
	2	8.93	8.88	8.98	139	
2,4-DB	1	10.06	10.01	10.11	64.6	18.4
	2	10.82	10.77	10.87	77.7	
DICHLORPROP	1	8.10	8.05	8.15	112	13.2
	2	8.61	8.56	8.66	98.1	
Dinoseb	1	11.27	11.22	11.32	25.5	37.8
	2	11.20	11.15	11.25	17.4	
DICAMBA	1	7.39	7.34	7.44	143	11.1
	2	7.89	7.84	7.94	128	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

OU4-VSL-14-121224MSD

Contract: NOBI03

Lab Code: CHEM Case No.: P5306

SAS No.: P5306 SDG NO.: P5306

Lab Sample ID: P5306-15MSD

Date(s) Analyzed: 12/26/2024 12/26/2024

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP ID: 0.32 (mm)

GC Column: (2): RTX-CLP2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Dalapon	1	2.62	2.57	2.67	118	14.9
	2	2.68	2.63	2.73	137	
DICHLORPROP	1	8.10	8.05	8.15	112	13.5
	2	8.61	8.56	8.66	97.8	
2,4-D	1	8.33	8.28	8.38	163	14.5
	2	8.93	8.88	8.98	141	
2,4,5-TP(Silvex)	1	9.20	9.15	9.25	90.9	10.4
	2	9.84	9.79	9.89	81.9	
2,4,5-T	1	9.49	9.44	9.54	114	13.9
	2	10.25	10.20	10.30	99.2	
2,4-DB	1	10.06	10.01	10.11	64.6	18.5
	2	10.82	10.77	10.87	77.8	
Dinoseb	1	11.27	11.22	11.32	26.0	39.1
	2	11.20	11.15	11.25	17.5	
DICAMBA	1	7.39	7.34	7.44	142	10.4
	2	7.89	7.84	7.94	128	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB165776BS

Contract: NOBI03

Lab Code: CHEM Case No.: P5306

SAS No.: P5306

SDG NO.: P5306

Lab Sample ID: PB165776BS

Date(s) Analyzed: 12/23/2024 12/23/2024

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP ID: 0.32 (mm)

GC Column: (2): RTX-CLP2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Dalapon	1	2.62	2.57	2.67	161	3.8
	2	2.68	2.63	2.73	155	
DICHLORPROP	1	8.10	8.05	8.15	171	5.4
	2	8.61	8.56	8.66	162	
2,4-D	1	8.33	8.28	8.38	171	6
	2	8.94	8.89	8.99	161	
2,4,5-TP(Silvex)	1	9.20	9.15	9.25	174	6.5
	2	9.84	9.79	9.89	163	
2,4,5-T	1	9.50	9.45	9.55	174	6.5
	2	10.26	10.21	10.31	163	
2,4-DB	1	10.07	10.02	10.12	165	1.8
	2	10.83	10.78	10.88	162	
Dinoseb	1	11.27	11.22	11.32	171	7.3
	2	11.20	11.15	11.25	159	
DICAMBA	1	7.40	7.35	7.45	171	6.6
	2	7.90	7.85	7.95	160	