

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: P5306 SAS No.: P5306 SDG NO.: P5306

Continuing Calib Date: 12/26/2024 Initial Calibration Date(s): 12/23/2024 12/23/2024

Continuing Calib Time: 18:40 Initial Calibration Time(s): 11:23 12:59

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.39	7.40	7.30	7.50	0.01
2,4-DCAA	7.21	7.21	7.11	7.31	0.00
Dalapon	2.62	2.62	2.52	2.72	0.00
DICHLORPROP	8.10	8.10	8.00	8.20	0.00
2,4-D	8.33	8.33	8.23	8.43	0.00
2,4,5-TP(Silvex)	9.20	9.21	9.11	9.31	0.01
2,4,5-T	9.49	9.50	9.40	9.60	0.01
2,4-DB	10.06	10.07	9.97	10.17	0.01
Dinoseb	11.27	11.28	11.18	11.38	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: P5306 SAS No.: P5306 SDG NO.: P5306

Continuing Calib Date: 12/26/2024 Initial Calibration Date(s): 12/23/2024 12/23/2024

Continuing Calib Time: 18:40 Initial Calibration Time(s): 11:23 12:59

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.89	7.90	7.80	8.00	0.01
2,4-DCAA	7.70	7.70	7.60	7.80	0.00
Dalapon	2.67	2.67	2.57	2.77	0.00
DICHLORPROP	8.61	8.61	8.51	8.71	0.00
2,4-D	8.93	8.94	8.84	9.04	0.01
2,4,5-TP(Silvex)	9.84	9.84	9.74	9.94	0.00
2,4,5-T	10.25	10.26	10.16	10.36	0.01
2,4-DB	10.82	10.83	10.73	10.93	0.01
Dinoseb	11.20	11.20	11.10	11.30	0.00