

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: P5306 SAS No.: P5306 SDG NO.: P5306

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL01 Date Analyzed: 12/23/2024

Lab Sample No.: HSTDCCC750 Data File : PS028794.D Time Analyzed: 14:11

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.496	9.397	9.597	734.150	712.500	3.0
2,4,5-TP(Silvex)	9.204	9.106	9.306	736.040	712.500	3.3
2,4-D	8.330	8.232	8.432	719.010	705.000	2.0
2,4-DB	10.067	9.969	10.169	711.030	712.500	-0.2
2,4-DCAA	7.209	7.110	7.310	766.590	750.000	2.2
Dalapon	2.619	2.520	2.720	708.540	682.500	3.8
DICAMBA	7.395	7.297	7.497	730.800	705.000	3.7
DICHLORPROP	8.101	8.002	8.202	719.010	705.000	2.0
Dinoseb	11.274	11.176	11.376	721.470	705.000	2.3

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: P5306 SAS No.: P5306 SDG NO.: P5306

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL01 Date Analyzed: 12/23/2024

Lab Sample No.: HSTDCCC750 Data File : PS028794.D Time Analyzed: 14:11

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.260	10.160	10.360	713.330	712.500	0.1
2,4,5-TP(Silvex)	9.841	9.742	9.942	716.070	712.500	0.5
2,4-D	8.939	8.840	9.040	705.580	705.000	0.1
2,4-DB	10.825	10.726	10.926	720.650	712.500	1.1
2,4-DCAA	7.699	7.600	7.800	742.000	750.000	-1.1
Dalapon	2.675	2.574	2.774	682.490	682.500	0.0
DICAMBA	7.897	7.798	7.998	712.440	705.000	1.1
DICHLORPROP	8.611	8.512	8.712	701.520	705.000	-0.5
Dinoseb	11.203	11.104	11.304	691.380	705.000	-1.9