

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: P5306 SAS No.: P5306 SDG NO.: P5306

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL02 Date Analyzed: 12/23/2024

Lab Sample No.: HSTDCCC750 Data File : PS028800.D Time Analyzed: 16:34

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.495	9.397	9.597	746.360	712.500	4.8
2,4,5-TP(Silvex)	9.204	9.106	9.306	749.460	712.500	5.2
2,4-D	8.329	8.232	8.432	732.400	705.000	3.9
2,4-DB	10.065	9.969	10.169	724.540	712.500	1.7
2,4-DCAA	7.208	7.110	7.310	784.350	750.000	4.6
Dalapon	2.620	2.520	2.720	717.130	682.500	5.1
DICAMBA	7.395	7.297	7.497	744.950	705.000	5.7
DICHLORPROP	8.100	8.002	8.202	732.170	705.000	3.9
Dinoseb	11.273	11.176	11.376	736.450	705.000	4.5

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: P5306 SAS No.: P5306 SDG NO.: P5306

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL02 Date Analyzed: 12/23/2024

Lab Sample No.: HSTDCCC750 Data File : PS028800.D Time Analyzed: 16:34

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.259	10.160	10.360	716.190	712.500	0.5
2,4,5-TP(Silvex)	9.841	9.742	9.942	719.750	712.500	1.0
2,4-D	8.939	8.840	9.040	709.650	705.000	0.7
2,4-DB	10.824	10.726	10.926	725.030	712.500	1.8
2,4-DCAA	7.699	7.600	7.800	749.440	750.000	-0.1
Dalapon	2.676	2.574	2.774	681.460	682.500	-0.2
DICAMBA	7.896	7.798	7.998	719.320	705.000	2.0
DICHLORPROP	8.610	8.512	8.712	712.290	705.000	1.0
Dinoseb	11.203	11.104	11.304	701.160	705.000	-0.5