

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: P5306 SAS No.: P5306 SDG NO.: P5306

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL03 Date Analyzed: 12/23/2024

Lab Sample No.: HSTDCCC750 Data File : PS028811.D Time Analyzed: 20:57

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.492	9.397	9.597	764.390	712.500	7.3
2,4,5-TP(Silvex)	9.201	9.106	9.306	767.440	712.500	7.7
2,4-D	8.326	8.232	8.432	733.770	705.000	4.1
2,4-DB	10.063	9.969	10.169	741.330	712.500	4.0
2,4-DCAA	7.206	7.110	7.310	783.300	750.000	4.4
Dalapon	2.621	2.520	2.720	733.750	682.500	7.5
DICAMBA	7.392	7.297	7.497	761.470	705.000	8.0
DICHLORPROP	8.097	8.002	8.202	749.570	705.000	6.3
Dinoseb	11.270	11.176	11.376	757.100	705.000	7.4

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: P5306 SAS No.: P5306 SDG NO.: P5306

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL03 Date Analyzed: 12/23/2024

Lab Sample No.: HSTDCCC750 Data File : PS028811.D Time Analyzed: 20:57

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.255	10.160	10.360	728.020	712.500	2.2
2,4,5-TP(Silvex)	9.837	9.742	9.942	731.260	712.500	2.6
2,4-D	8.936	8.840	9.040	718.240	705.000	1.9
2,4-DB	10.820	10.726	10.926	737.840	712.500	3.6
2,4-DCAA	7.696	7.600	7.800	759.590	750.000	1.3
Dalapon	2.675	2.574	2.774	692.920	682.500	1.5
DICAMBA	7.894	7.798	7.998	730.170	705.000	3.6
DICHLORPROP	8.607	8.512	8.712	722.900	705.000	2.5
Dinoseb	11.199	11.104	11.304	712.830	705.000	1.1