

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: P5306 SAS No.: P5306 SDG NO.: P5306

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL04 Date Analyzed: 12/26/2024

Lab Sample No.: HSTDCCC750 Data File : PS028820.D Time Analyzed: 11:07

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.492	9.397	9.597	794.470	712.500	11.5
2,4,5-TP(Silvex)	9.200	9.106	9.306	803.120	712.500	12.7
2,4-D	8.326	8.232	8.432	766.560	705.000	8.7
2,4-DB	10.062	9.969	10.169	769.040	712.500	7.9
2,4-DCAA	7.207	7.110	7.310	816.550	750.000	8.9
Dalapon	2.622	2.520	2.720	774.990	682.500	13.6
DICAMBA	7.393	7.297	7.497	792.890	705.000	12.5
DICHLORPROP	8.098	8.002	8.202	775.840	705.000	10.0
Dinoseb	11.269	11.176	11.376	788.080	705.000	11.8

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: P5306 SAS No.: P5306 SDG NO.: P5306

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL04 Date Analyzed: 12/26/2024

Lab Sample No.: HSTDCCC750 Data File : PS028820.D Time Analyzed: 11:07

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.253	10.160	10.360	723.860	712.500	1.6
2,4,5-TP(Silvex)	9.836	9.742	9.942	730.730	712.500	2.6
2,4-D	8.934	8.840	9.040	707.750	705.000	0.4
2,4-DB	10.819	10.726	10.926	730.050	712.500	2.5
2,4-DCAA	7.695	7.600	7.800	751.940	750.000	0.3
Dalapon	2.676	2.574	2.774	701.470	682.500	2.8
DICAMBA	7.893	7.798	7.998	723.730	705.000	2.7
DICHLORPROP	8.606	8.512	8.712	714.020	705.000	1.3
Dinoseb	11.197	11.104	11.304	713.420	705.000	1.2